May 3, 2018

Medtronic Core Valve LLC
% Mark Job
Regulatory Technology Services, LLC
1394 25th Street, NW
Buffalo, Minnesota 55313

Re: K181001
Trade/Device Name: Medtronic Confida™ Brecker Guidewire
Regulation Number: 21 CFR 870.1330
Regulation Name: Catheter Guide Wire
Regulatory Class: Class II
Product Code: DQX
Dated: April 12, 2018
Received: April 16, 2018

Dear Mr. Job:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good

manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (https://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/) and CDRH Learn (http://www.fda.gov/Training/CDRHLearn). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (http://www.fda.gov/DICE) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Nicole G. Ibrahim -S

for

Bram D. Zuckerman, M.D.
Director
Division of Cardiovascular Devices
Office of Device Evaluation
Center for Devices and Radiological Health

Enclosure

DEPARTMENT OF HEALTH AND HUMAN SERVICES
Food and Drug Administration

Indications for Use

Form Approved: OMB No. 0910-0120
Expiration Date: 06/30/2020
See PRA Statement below.

510(k) Number *(if known)*

K181001

Device Name
Medtronic Confida™ Brecker Guidewire

Indications for Use *(Describe)*
The Medtronic Confida™ Brecker Guidewire is intended for use to introduce and position catheters during diagnostic and interventional procedures within the chambers of the heart, including transcatheter aortic valve implantation (TAVI).

Type of Use *(Select one or both, as applicable)*

☒ Prescription Use (Part 21 CFR 801 Subpart D) ☐ Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRAStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

5.0 510(k) Summary

As required by the Safe Medical Devices Act of 1990, coded under Section 513, part (1)(3)(A) of the Food, Drug and Cosmetic Act, a 510(k) summary upon which substantial equivalence determination is based is as follows:

Date Prepared:	March 30, 2018
Applicant:	Medtronic CoreValve LLC 1851 E Deere Ave Santa Ana, CA 92705 USA Establishment Registration NO. 2025587
Contact Person:	Gerardine Drummond Senior Regulatory Affairs Specialist Medtronic Ireland Parkmore Business Park West Galway, Ireland Phone: (+353-91)-708914 Fax: (+353-91)-708672 Email: Gerardine.drummond@medtronic.com
Subject Device Name:	Medtronic Confida™ Brecker Guidewire
Model Number:	GWBC30
Common Name:	Guidewire
Classification Name:	Wire, Guide, Catheter
Product Code:	DQX
Predicate Device(s):	Medtronic Confida™ Brecker Guidewire (K150465 cleared August 03, 2015)

Device Description:	The Medtronic Confida™ Brecker Guidewire was developed for use in diagnostic and interventional procedures including transcatheter aortic valve implantation (TAVI) procedures for the treatment of aortic valve disease. The Medtronic Confida™ Brecker Guidewire (model number GWBC30) is available in one size and is comprised of a stainless steel wire, 0.035" in diameter, and 260cm in length. The distal end of the Medtronic Confida™ Brecker Guidewire is comprised of a preformed 360° curved tip. The Medtronic Confida™ Brecker Guidewire is an external communicating device having limited contact (<24 hours) with circulating blood. The Medtronic Confida™ Brecker Guidewire has a polytetrafluoroethylene (PTFE) coating applied to the entire length of the device in order to aid in lubricity. The Medtronic Confida™ Brecker Guidewire is a patient contacting, single use only device and not intended for re-use or re-sterilization. The device is intended to be used only by physicians trained in percutaneous, intravascular techniques and procedures. The Medtronic Confida™ Brecker Guidewire is sterilized via Ethylene Oxide to a Sterility Assurance Level (SAL) of 10^{-6}. The purpose of this 510(k) is to notify FDA of a change to the PTFE coating applied to the core wire section of the Confida™ Brecker Guidewire. The change is to replace the existing low Chromium VI PTFE coating with Chromium VI free PTFE coating.
Statement of Intended Use:	The Medtronic Confida™ Brecker Guidewire is intended to facilitate the placement of devices during diagnostic and interventional procedures.
Statement of Indications for Use:	The Medtronic Confida™ Brecker Guidewire is intended for use to introduce and position catheters during diagnostic and interventional procedures within the chambers of the heart, including transcatheter aortic valve implantation (TAVI).
Contraindications:	The Medtronic Confida™ Brecker Curve™ Guidewire is contraindicated for patients presenting with an intolerance to anticoagulation therapy and unheparinized patients. The guidewire is contraindicated for use in the coronary arteries and in the cerebrovasculature.

Comparison to Predicate device:

The Medtronic Confida™ Brecker Guidewire is identical to the predicate device for the following characteristics:

- Intended use
- Indications for use
- Contraindications
- Target population
- Fundamental scientific technology, including design
- Operating principle
- Packaging materials
- Shelf life
- Sterility assurance level and method of sterilization

Summary of Non-Clinical Data

In order to demonstrate substantial equivalence of the subject device, the Medtronic Confida™ Brecker Guidewire, to the predicate device, the following non-clinical evaluations were performed:

- PTFE Coating Adhesion Test (Coating Integrity)
- ISO Cytotoxicity Study
- ISO Maximization Sensitization Study
- ISO Intracutaneous Study
- ISO/USP Pyrogen Study Material Mediated
- ISO Acute Systematic Toxicity Study
- *In Vivo* Thromboresistance
- *In Vitro* Hemolysis Study (Modified ASTM – Extraction Method)
- SC5b-9 Compliment Activation

All testing met the specified acceptance criteria

Conclusion:

Based on the information above, the subject device, the Medtronic Confida™ Brecker Guidewire, is substantially equivalent in intended use, performance, and fundamental scientific technology to the predicate device, the Medtronic Confida™ Brecker Guidewire (K150465).