

* Removed 2 stds + 5 vials 27 & made aliquots
x 2 std remain in kit box.



UmanDiagnostics

Rapid detection of brain disease

2 Kits lot.

Recv: - 12/5/2021 (C#2)
22/04/2021 (C#1)

Certificate Of Analysis

Product	Article no	Lot no	Manufacturing date	Expiration date
NF-light® ELISA CE	10-7001	70804	2020-03	2021-09

Components	Lot no
Standard Reconstitution Volume: 1700 µl	70805
Tracer 50x	70806
Conjugate 65x	70807
Conjugate diluent	70808
Sample diluent	70809
TMB substrate	70810
Wash solution 10x	70811
Stop Solution	70812
Plate	70813

Additional material
Instruction for use (Version 2019-02)
Plate cover 2 pcs
Tubes for conjugate solution 2 pcs

Storage
Store at +2- +8 °C

Analysis	Acceptance Criteria	Results
CSF Quality control sample 1 (pg/ml)	1758 – 2829 pg/ml	2275 pg/ml
CSF Quality control sample 2 (pg/ml)	3152 - 5370 pg/ml	3944 pg/ml
Background (A ₄₅₀ - A ₆₂₀) (AU)	< 0.1 AU	0.03 AU

Date of approval	Signature
2020-05-20	Malin Stigbrand Malin Stigbrand, Quality Assurance

UmanDiagnostics AB
www.umandiagnostics.se

Visiting address:
Tvistevägen 48C
907 36 Umeå, Sweden

Organization Number:
556698-2418

Contact:
info@umandiagnostics.se
Tel +46 90 777 880

VAT Number:
SE556698241801