



Control(s) Certificate of Analysis

Date: 4/23/2022

Species: Human, Mouse, Non Human Primate, Rat

The following items/volumes have been certified by ProteinSimple:

Assay	Control Vial Part Number	Control Vial Lot Number	Reconstitution Volume (uL)	Nominal Concentration (pg/mL)	Low QC (pg/mL)	High QC (pg/mL)
NFL	898349	P303035	900	128625	66.8 - 111	2871 - 4785

Expected Values

The ranges in the above tables represent the average measured concentration +/- 25% for the target assay. Due to possible variations in techniques, it is recommended that each laboratory determine its own target range.

Disclaimer

Please verify the lot number located on the control vial.

Certificate(s) of Analysis for additional lot numbers can be downloaded at our website:

www.proteinsimple.com.

Certified By:

Jennifer Pendleton

Quality Representative

2022-April-23

Date

Signature