



ChroPlate - ProteinG [↗](#)

Alexandra Ehl¹, David Frommholz¹, Nadine Stefanczyk¹

¹DALEX Biotech

[dx.doi.org/10.17504/protocols.io.uyjexun](https://doi.org/10.17504/protocols.io.uyjexun)



David Frommholz
DALEX Biotech



ABSTRACT

Purification Guide for the Isolation of Antibodies with ChroPlate Filtration Plates by DALEX Biotech.

Easy and quick high throughput antibody purification from various sources and species.

Each well of the ChroPlate has a binding capacity of > 1 mg (tested with human polyclonal Ig, binding varies between species and clones).

The proprietary resin does not shrink or swell in aqueous buffers.

High pressure stability.

pH stability short term 2 - 8, long term 3 - 8.

Excellent thermal stability up to 15 minutes at 80 °C in aqueous buffers at neutral pH.

Can be dried for long term storage (80 °C for > 2 h).

EXTERNAL LINK

<https://dalex-biotech.com/>

PROTOCOL STATUS

Working

Official product protocol by DALEX Biotech

GUIDELINES

For optimal binding and purity, the pH of the sample should be 7.5-8.5 and should contain 150-300 mM NaCl. An easy way to achieve this is by adding 1/11 volume 0.5 M Tris, 2 M NaCl (pH 8.0) to your sample. For screening of binding condition for e.g. a monoclonal antibody this parameters might be varied.

Purification works best with an antibody concentration of 1-2 mg/ml in your sample.

It is advisable that all fractions are collected (Sample, flow through, wash, and eluate) in separate plates for analysis, e.g. SDS-PAGE

MATERIALS TEXT

Materials provided in the kit:

ChroPlate

Dummy plate

Wash buffer

Elution buffer

Neutralization buffer

Materials not provided in the kit:

Tween-20

10x PBS

SAFETY WARNINGS

The buffers in the kit include sodium azide (CAS No. 26628-22-8) as a preservative.

For safety information on this chemical(s) check <http://www.dguv.de/ifa/gestis-database>

BEFORE STARTING

Make sure your sample is free of particulate matter. You can remove particles by centrifugation or filtration (0.45 µm).

How do you want to purify?

- 1 Do you want to purify by centrifugation or by vacuum filtration? Please choose below.

step case

Centrifugation

no description provided

Equilibration

- 2 Add 500 µl wash buffer to each well, place the ChroPlate on top of a deep-well plate, and centrifuge 5 minutes at 1000 g in a swing-out rotor. For counterbalance of the centrifuge a dummy filter plate is included in the kit.

Load and Wash

- 3 Place the ChroPlate on a clean deep-well plate. Add up to 1.5 ml sample to every well. Centrifuge 5 minutes at 1000 g in a swing-out rotor.

NOTE

It is advisable to keep the flow-through and wash fractions for later analysis, e.g. SDS-PAGE.

The centrifugation time depends on the sample's volume and viscosity. Volumes larger than 1 ml and viscous samples e.g. serum usually require more than 5 minutes centrifugation time.

For optimal binding and purity, the pH of the sample should be 7 - 8 and should contain 150 - 300 mM NaCl. An easy way to achieve this is by adding 1/11 volume 10x PBS to your sample.

- 4 Empty the deep-well plate or place the ChroPlate on a clean one. Add 500 µl wash buffer to each well and centrifuge 5 minutes at 1000 g in a swing-out rotor.

NOTE

In case of unspecific hydrophobic and/or ionic interactions include up to 1 % Tween-20 and/or up to 0.5 M NaCl in the wash buffer.

- 5 Repeat the previous step.

NOTE

For increased purity, repeat the washing step a third time.

Elution

- 6 Place the ChroPlate on a clean deep-well plate. Add 100 µl elution buffer to each well and centrifuge 5 minutes at 1000 g in a swing-out rotor. Repeat two more times.

- 7 Add one drop of neutralization solution to every well of the deep-well plate and mix gently.

Equilibration

step case

Vacuum filtration

no description provided

- 2 Assemble your vacuum device according to the manufacturer's instructions.
Add 500 µl wash buffer to each well, place the ChroPlate on a waste receptacle and apply a vacuum of 100 mbar for 2 minutes.

Load and Wash

- 3 Place the ChroPlate on a clean deep-well plate. Add up to 1.5 ml sample to every well. Apply a vacuum of 100 mbar for 2 - 3 minutes.

NOTE

It is advisable to keep the flow-through and wash fractions for later analysis, e.g. SDS-PAGE.

Strength and time of the vacuum depends on the sample's volume and viscosity. Volumes larger than 1 ml and viscous samples e.g. serum usually require more than 2 minutes.

For optimal binding and purity, the pH of the sample should be 7 - 8 and should contain 150 - 300 mM NaCl. An easy way to achieve this is by adding 1/11 volume 10x PBS to your sample.



This is an open access protocol distributed under the terms of the [Creative Commons Attribution License](https://creativecommons.org/licenses/by/4.0/), which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited