

CB-X™ Protein Assay for simple aqueous samples

G-Biosciences

Abstract

Citation: G-Biosciences CB-X™ Protein Assay for simple aqueous samples. **protocols.io**

dx.doi.org/10.17504/protocols.io.e33bgqn

Published: 21 Jun 2016

Guidelines

INTRODUCTION

Protein assays are routinely used in many research fields to estimate proteins in a vast array of buffers and conditions. A major problem for researchers is to select a protein assay from the vast selection on the market that is compatible with their protein sample. CB-X™ Protein Assay eliminates this problem as it is designed to be compatible with all commonly used buffers and conditions in protein isolation, storage and assays.

For protein samples in simple, uncomplicated aqueous buffers CB-X™ is a highly sensitive, single reagent assay that can be performed in 5 minutes. CB-X™ Protein Assay uses a protein dye that is an improvement on the Bradford Coomassie dye reagent.

For complicated protein samples CB-X™ Protein Assay is supplied with reagents to clean up the samples and remove all reagents and chemicals that interfere with accurate protein estimation. These reagents include detergents, chaotropes, reducing agents, alkylating agents, sugars, high salt concentrations, buffering agents and chelating agents (Table 1). The clean up stage and subsequent protein assay is performed in a single tube to ensure no protein loss and to maintain the accuracy of the assay.

DETERGENTS		REDUCING AGENTS	
Brij* 35	2%	2-mercaptoethanol	1M
CHAPS	2%	DTT	1M
CHAPSO	2%	CHAOTROPES	

Nonidet* P-40	2%	Guanidine.HCl	6M
SDS	2%	Urea	6M
Triton* X-100	2%	SALTS	
Tween* 20	2%	Ammonium sulfate	1M
Deoxycholate	0.1%	MISCELLANEOUS	
SUGARS		EDTA	0.1M
Glucose	1M	HEPES	0.1M
Sucrose	25%	MES	0.1M

Table 1: CB-X™ Protein Assay is compatible with many interfering agents

Figure 1 demonstrates the efficiency of the CB-X™ Protein Assay. If interfering agents are present, such as detergents, or if an artifact results are produced then the protein samples are treated with the clean up reagents and the protein is then assayed generating a linear response.




Figure 1: CB-X™ Protein Assay. Inhibitory effects of detergents on protein assays are abolished with CB-X™. Protein solutions containing 1% Triton® X-100 or 1% SDS were assayed using a standard Coomassie dye protein assay. The same protein samples with 1% Triton® X-100 or 1% SDS were assayed using CB-X™ protein assay. A linear response to increasing protein concentration was visualized, indicating no interference by the detergents.

CB-X™ Protein Assay is supplied with a lot specific CB-X™ Tables. These allow researchers to perform

single protein clean ups, subsequent assays and then look up their absorbance in the CB-X™ Table to find the protein concentration. The CB-X™ Table eliminates the need for multiple protein standards and saves considerable time and effort. The CB-X™ Table is prepared with a complex protein mixture that compares well with proteins from mammalian, plant, bacteria and yeast sources. An optional set of bovine serum albumin standards or non-animal protein standards are supplied with Cat. # 786-12X or 786-894 respectively for generating curves when using CB-X™ Assay Dye alone or for researcher's who prefer to generate their own standard curve or to generate their own CB-X™ Table for their specific conditions.

The CB-X™ Protein Assay is reliable over the range of 0.5-50µg per assay. The regular size kit contains enough CB-X™ Assay Dye for 500 protein assays and enough clean up reagents for 250 clean ups. (*Patents Pending*)

ITEM(S) SUPPLIED

Description	Cat. # 786-11X	Cat. # 786-12X	Cat. # 786-894	Cat. # 786-12XT
CB-X™	2 x 125ml	2 x 125ml	2 x 125ml	10ml
CB-X™ Assay Dye	2 x 250ml	2 x 250ml	2 x 250ml	10ml
CB-X™ Solubilization Buffer-I	15ml	15ml	15ml	1ml
CB-X™ Solubilization Buffer-II	15ml	15ml	15ml	1ml
CB-X™ Protein Standard [2mg/ml]	-	5ml		-
Non-Animal Protein Standard [2mg/ml]	-	-	5ml	-
CB-X™ Table: Lot Specific	1	1	1	1

STORAGE CONDITION

The kit is shipped at ambient temperature. Upon arrival, store CB-X™ Assay Dye and protein standard at 4°C and other kit components at room temperature. The CB-X™ reagent must be chilled to -20°C for optimal efficiency, for additional convenience the CB-X™ reagent can be stored at -20°C. When stored and used properly this kit is good for 12 months.

ITEMS NEEDED AND NOT SUPPLIED WITH THIS KIT

- Centrifuge
- Assay tubes (G-Biosciences Cat # 786-008)
- Disposable polystyrène cuvettes (G-Biosciences Cat # 786-009)

Materials

CB-X™ Protein Assay [786-12XT](#) by [G-Biosciences](#)

Assay Tubes (2ml) [786-008](#) by [G-Biosciences](#)

Disposable polystyrène cuvettes (1ml) [786-009](#) by [G-Biosciences](#)

Protocol

PREPARATIONS BEFORE USE

Step 1.

Chill CB-X™ at -20°C.

ⓘ NOTES

Colin Heath 07 Jun 2016

NOTE: CB-X™ can be stored at -20°C for prolonged storage and additional convenience.

PREPARATIONS BEFORE USE

Step 2.

Prior to use, mix CB-X™ Assay Dye by gently inverting the bottle several times. DO NOT SHAKE THE BOTTLE.

PROTOCOL 1: FOR SIMPLE, UNCOMPLICATED AQUEOUS SAMPLES

Step 3.

Dilute the supplied protein standards in the same buffer as your protein samples. We recommend duplicates of 5-6 dilutions within the range of 0.1-1.0mg/ml.

PROTOCOL 1: FOR SIMPLE, UNCOMPLICATED AQUEOUS SAMPLES

Step 4.

Transfer 50µl diluted protein standards and test samples to the assay tubes. Set up a blank containing 50µl of the buffer used to generate the standards.

PROTOCOL 1: FOR SIMPLE, UNCOMPLICATED AQUEOUS SAMPLES

Step 5.

Add 1ml CB-X™ Assay Dye into each tube and mix well. Incubate for 5 minutes at room temperature.

DURATION

00:05:00

PROTOCOL 1: FOR SIMPLE, UNCOMPLICATED AQUEOUS SAMPLES

Step 6.

Read the absorbance at 595nm.

PROTOCOL 1: FOR SIMPLE, UNCOMPLICATED AQUEOUS SAMPLES

Step 7.

Prepare a standard calibration plot for the determination of protein concentration of the unknown samples.

NOTES

Colin Heath 07 Jun 2016

NOTE: This assay can be performed in titer plates. Use 10µl protein standards or samples and 200µl CB-X™ Assay Dye and follow the above protocol.