TOPICAL REVIEW

Active upper limb prostheses: a review on current state and upcoming breakthroughs

To cite this article: Andrea Marinelli *et al* 2023 *Prog. Biomed. Eng.* **5** 012001

View the article online for updates and enhancements.

You may also like

Progress in Biomedical Engineering

TOPICAL REVIEW

Active upper limb prostheses: a review on current state and upcoming breakthroughs

Andrea Marinelli[1,2,6,*] ![ORCID], Nicolò Boccardo[1,5,6], Federico Tessari[1], Dario Di Domenico[1,3] ![ORCID], Giulia Caserta[1], Michele Canepa[1,5], Giuseppina Gini[4], Giacinto Barresi[1], Matteo Laffranchi[1], Lorenzo De Michieli[1,6] and Marianna Semprini[1,6] ![ORCID]

1   RehabTechnology Lab, Italian Institute of Technology, Via Morego, 30, 16163 Genova, GE, Italy
2   Bioengineering Lab, University of Genova, DIBRIS, Genova, Italy
3   Department of Electronics and Telecommunications, Politecnico di Torino, Turin 10124, Italy
4   Department of Electronics, Computer Science and Bio–Engineering, Politecnico di Milano, Piazza L. da Vinci, 31, I–20133 Milano, MI, Italy
5   The Open University Affiliated Research Centre at Istituto Italiano di Tecnologia (ARC@IIT), via Morego 30, 16163 Genova, Italy
6   Equal contribution.
*   Author to whom any correspondence should be addressed.

E-mail: andrea.marinelli@iit.it

Keywords: bionics, biosignals, closed-loop, embodiment, feedback, prosthetic hand, user experience

Abstract

The journey of a prosthetic user is characterized by the opportunities and the limitations of a device that should enable activities of daily living (ADL). In particular, experiencing a bionic hand as a functional (and, advantageously, embodied) limb constitutes the premise for promoting the practice in using the device, mitigating the risk of its abandonment. In order to achieve such a result, different aspects need to be considered for making the artificial limb an effective solution to accomplish ADL. According to such a perspective, this review aims at presenting the current issues and at envisioning the upcoming breakthroughs in upper limb prosthetic devices. We first define the sources of input and feedback involved in the system control (at user-level and device-level), alongside the related algorithms used in signal analysis. Moreover, the paper focuses on the user-centered design challenges and strategies that guide the implementation of novel solutions in this area in terms of technology acceptance, embodiment, and, in general, human-machine integration based on co-adaptive processes. We here provide the readers (belonging to the target communities of researchers, designers, developers, clinicians, industrial stakeholders, and end-users) with an overview of the state-of-the-art and the potential innovations in bionic hands features, hopefully promoting interdisciplinary efforts for solving current issues of upper limb prostheses. The integration of different perspectives should be the premise to a transdisciplinary intertwining leading to a truly holistic comprehension and improvement of the bionic hands design. Overall, this paper aims to move the boundaries in prosthetic innovation beyond the development of a tool and toward the engineering of human-centered artificial limbs.

1. Introduction

Over the past 20 years, poly-articulated upper limb prostheses (ULPs) have undertaken several technological and scientific developments to satisfy the different needs of the upper limb amputee community. Nonetheless, in a recent study, Salminger *et al* (2020) observed overall abandonment rates of ULPs of about 44% in a population of mainly (92%) myoelectric prostheses users. They also highlighted how the past decade of developments still presents technological limiting factors that did not permit the restoration of the full functionalities of a missing limb, hence leading to a substantial increased rate of prosthesis abandonment. The main cause of such ineffectiveness mainly resides in a non-sufficiently patient-tailored design process (Salminger *et al* 2020).

According to the American Orthotic and Prosthetic Association (Aopa 2016), partial amputations, i.e. finger amputations, represent the majority of upper-limb losses (75.6%), while trans-radial and trans-humeral amputations constitute a percentage oscillating between 5% and 6%. Despite this, the level of impairment caused by trans-radial and trans-humeral amputations is greater than for partial amputations.

Without tracing back all the evolution of ULPs—the reader might find useful the reviews of Trent *et al* (2019) and Ribeiro *et al* (2019). Trent *et al* (2019) work focuses on a classification of the upper-limb prostheses architectures based on the type of adopted actuation, e.g. passive, body-powered or active. On the other hand, Ribeiro *et al* (2019)'s research investigates the most relevant control signals used for the man-machine interface.

This work focuses on trans-radial and trans-humeral devices, excluding partial amputations, and it details the latest and most technologically advanced solutions, namely poly-articulated myoelectric prostheses. Moreover, this review aims at presenting and analyzing the key elements of state-of-the-art ULPs in a user-centered and human-in-the-loop fashion and to provide guidelines for the development of such prostheses and the relative control algorithms, to possibly achieve solutions capable of promoting the systems use and overcoming the elevated abandonment rates observed so far. Overall, the reader could take advantage of this review as an analytical collection of solutions constituting a premise to provide the user with a seamless control experience.

2. Upper limb prosthetics classification: a twofold perspective

An ULP system can be observed from two main points of views: its mechatronics, namely the combination of the mechanical and electronic components necessary for its operation, and the control strategies and algorithms implemented to orchestrate its functions. Research groups have therefore attempted to solve the prostheses abandonment problem by addressing different technological and scientific challenges, either focusing on mechatronic design, or on control strategies aimed at increasing the human-machine interaction and, in some cases, introducing feedback sources, as detailed in the next sections.

ULP control can be divided into two synergistically interacting sub-systems: the user-level and the device-level, as depicted in figure 1. The user-level includes the patients and the most proximal device component interacting with the user (i.e. the socket), while the device-level extends from the socket to the ULP device. These two sub-systems overlap at the socket level, which is involved in a bidirectional flow of information. On one hand, it receives inputs from the user (i.e. movement intentions) and translates them into movement commands for the device; on the other hand, it receives information (both from the device and the environment) and communicates it to the user through sensory feedback (figure 1). Importantly, the socket itself severely limits the user comfort, and together with the prosthetic weight highly contributes to the prosthetic abandonment.

Even if the state-of-the-art in prosthetic research encompasses studies based on psychological processes too, commercial ULP systems have focused on restoring functional capabilities by capitalizing on the device-level only, therefore on mechatronic, and several solutions can be found on the market for trans-radial level of amputations. Commercially available systems merge basic functionalities and aesthetic requirements, targeting the clinical needs given by a certain kind of amputation, rather than focusing on each patient's specific needs.

Commercial solutions range from tri-digital hands, e.g. VaryPlus Speed, SensorHand Speed by Ottobock (Ottobock 2020c) and Motion Control Hand by Fillauer (Fillauer 2021); through polyarticulated hand under-actuated, e.g. Michelangelo by Ottobock (Ottobock 2020b); to fully actuated polyarticulated hand, e.g. BeBionic by Ottobock (Ottobock 2020a), i-Limb by Ossur (Ossur 2020b), Vincent Hand by Vincent Systems (Systems 2020), TASKA hand by Taska Prosthetics (Taska 2022), BrainRobotics Hand by BrainRobotics (Brainrobotics 2022) and Ability Hand by Psyonic (Psyonic 2022).

In the last decades, many research groups have focused on the mechatronic development of ULP devices, entrusting the intelligence of the device to the embedded mechanics in a very thorough design, structuring the development of the concept of under-actuation, such as the Vanderbilt Multigrasp Hand (Bennett *et al* 2014), the MIA Hand (Controzzi *et al* 2016), the SoftHand Pro (Godfrey *et al* 2018), the KIT Hand (Weiner *et al* 2018), and the Hannes Hand (Laffranchi *et al* 2020).

On the other hand, there is a family of very dexterous devices, not yet market-ready, that mimic the complexity of the human hand, implementing a fully-actuated multi-degrees of freedom mechatronics, e.g. the University of Bologna Hand (Meattini *et al* 2019) or the Shadow Hand (Company 2020).

High level of amputations, as the trans-humeral ones, require prosthetic elbows, such as the Dynamic Arm (Ottobock 2022a), the Dynamic Arm Plus (Ottobock 2022b), and the ErgoArm (Ottobock 2022c) from Ottobock; the Espire Elbow (Classic, Classic Plus, Pro and Hybrid,) from Steeper Inc. (Steeper 2022); and the

Figure 1. Graphical representation of a ULP system and its elements. The user level (left panel) includes: input data sent from subject to the prosthesis (input signals), artificial sensory feedback information delivered from the prosthesis to the user (sensory feedback), and external sources of interaction (external factors), such as actuation coming from the unimpaired limb or environmental/accidental sources of feedback such as vision and sound. The device-level (right panel) includes the control commands used to drive the prosthesis and the feedback information collected by the end-effector. The user-device interface is characterized by a bidirectional exchange of information (overlap of the two panels).

Fillauer Motion E2 Elbow (Fillauer 2022a) and the Utah Arm 3 (Fillauer 2022b) from Fillauer. In the research context, full robotic arms include the DLR hand system (Grebenstein *et al* 2011), the APL modular prosthetic limb (Johannes *et al* 2011), the LUKE Arm (Bionics 2022), the Rehabilitation Institute of Chicago arm (Lenzi *et al* 2016), and Edinburgh Modular Arm System (Gow *et al* 2001).

However, this great variety of products does not match with the elevated abandonment rates, demonstrating the lack of satisfaction of the patients' needs from a mechatronic perspective. In particular, structural and supporting part lack of adjustability of user size, allow limited kinematic and motion possibilities and more advanced systems present limited operational time (Harte *et al* 2017). This leads to limited satisfaction and feeling of security. Moreover, these systems generally present poor personal and social acceptance because of limited anthropomorphism, high weight and presence of acoustic disturbances during use (Harte *et al* 2017), This suggests that ULP development should not only focus on the device level, but improvements at the user level could play a key role for truly meeting the user requirements and consequently obtain device acceptance. Motivated by this, in this review, we analyze all the possible approaches that could potentially address the user needs in terms of device controllability, robustness and hence embodiment and user experience. To this end, it is fundamental not only to focus on the functionality restoration but also on the sensory information recovery, which are fundamental to effectively control the device. All the described approaches range from improvements in decoding user intentions, hence analyzing all possible input sources and their related control strategies, to inclusion of additional sources of feedback capable to restore the sensory information. These approaches tackle the issues related to poor device control because of lack of intuitiveness and sensory feedback.

Therefore, in this review we present current and emerging methods in ULP development, detailing various sources of input and feedback signals, as well as control strategies. We also highlight current challenges and open issues in the field, specifically focusing on the importance of user experience and involvement in the design and development process. This is fundamental to promote patient-tailored approaches leading to the development of truly personalized devices, which are currently lacking. We finally provide an overview of the most promising approaches that if followed, may one day provide upper limb amputees with a true substitute of their missing arm.

3. Input and feedback signals for prosthetic control

Prosthetic control is regulated by a flow of signals, as depicted in figure 2. Input signal runs from the user to the device and they are often of biological or electrophysiological nature, in which case are called biosignals. Signals flowing in the opposite direction convey information from the device to the user and are therefore defined as sensory feedback signals. Moreover, some external factors convey to the user additional source of feedback (i.e. incidental feedback), such as visual or auditory information that can be used to estimate the prosthesis state (Wilke *et al* 2019, Sensinger and Dosen 2020, Gonzalez *et al* 2021).

Figure 2. Graphical representation of information flow of a possible ULP architecture. Input flow (top panel): from user (input signals i.e. from EMG sensors) to prosthesis (control commands i.e. through power train). Feedback flow (bottom panel): from prosthesis (end-effector feedback i.e. from tactile force sensors) to user (sensory feedback i.e. through vibrotactile motors).

Figure 3. Input sources for ULP.

Input signals include all the sources of information that can be taken from the amputee and translated into motor commands for driving the prosthesis (e.g. electromyography—EMG), see section 3.1. Instead, sensory feedback information encompasses different prosthetic sensing solutions acquired either from the prosthetic device or from the environment, see section 3.2 that can be translated into sensory stimuli for the amputee (e.g. vibrotactile stimulation, see section 3.3). All types of signals can be classified according to their level of invasiveness, with consequent advantages and drawbacks.

3.1. Input signals

In recent years, many research activities have focused on the extraction of useful information from the biological signals in order to suitably control ULPs. Traditionally, the surface EMG (sEMG) is the most widespread signal for prosthesis control but its use still faces many drawbacks (Kyranou *et al* 2018). In the following, we describe various methods to employ EMG as input signal for ULP control and we also explain how other input sources can be exploited to obtain more dexterous prosthetic behavior, overcoming the limitations of current ULP systems.

Figure 3 collects input signals for ULP control that will be described in the following subsections, ranging from those used by commercial systems, up to those currently under investigation.

3.1.1. Biosignals

The term biosignal indicates every possible signal that can be detected and measured from biological beings, humans—in our case. Usually, the term is used for signals of electric nature (i.e. EMG), but actually every signal collected from the activity of different tissues or organs belonging to the human body, can be considered as a biosignal.

We here adopt this latter definition to group input sources that are described next. Given its large use both in research and commercial ULP devices, EMG deserves a dedicated subsection, while other biosignals are grouped together. We also dedicate a whole subsection to brain-derived signals, which are especially used in brain-machine and brain-computer interfaces (BMIs, BCIs), but that are also showing potential use for ULP applications. Table 1 summarizes biosignals for ULP control that will be described in the following subsections.

3.1.1.1. EMG

While cosmetics, electronic components and computational efforts have undergone a significant improvement, the control strategies currently used in prosthetic applications have not changed since their first appearance in the 1960s (Schmidl 1965). The EMG has been one of the major sources to control ULPs (Merletti and Farina 2016). These signals carry information about neuromuscular activity, and they are used to retrieve human intention. EMG is indeed a technique for studying the activation of the skeletal muscles through the recording of electrical potentials produced by muscle contraction (Hudgins *et al* 1993). The theory behind the sEMG electrodes is that they form a chemical equilibrium between the detecting surface of the electrode and the skin of the body through electrolytic conduction, so that the current can flow into the electrode.

Multiple methods have been used to obtain the intended gesture from the processed EMG signals, all of which exploit the fact that the amputees can still generate different and repeatable muscular patterns related to each forearm movement with residual muscles of the stump. Low-density EMG is commonly used in prosthetic application, both in research and commercial context. Noteworthy, EMG signals can also be collected with invasive methods. The sEMG can be thus classified according to the level of resolution and density of the sensors. In the following, we provide an overview of the different types of EMG-based biosignals.

3.1.1.1.1. sEMG

The sEMG can be classified according to the number of electrodes used (figure 4). Low-density EMG generally refers to the use of a small (<10) number of EMG bipolar sensors, that can be either wet, i.e. contain an electrolytic substance that serves as interface between skin and electrodes, or dry (Jamal 2012). Conversely, high-density EMG is typically composed by wet monopolar sensors spread on a planar patch, around 1 cm apart, and with the ground reference generally placed on the wrist or on the elbow (Drost *et al* 2006). Importantly, sEMG electrodes also differ in their electronic configuration, as they can be either preamplified or not (Zheng *et al* 2021). Merletti and Muceli (2019) provided a guide with the best practice to acquire and manipulate EMG data according with the different aims, from signal analysis to motion prediction.

Prosthetic control with low-density EMG is generally obtained by using two bipolar electrodes placed on antagonist muscles. This configuration allows the control of the prosthetic system in a robust and simple way (Hudgins *et al* 1993). However, the detection of complex and simultaneous movements of the phantom limb can be improved by using an array of EMG electrodes placed on the superficial skin of the residual forearm (Coapt 2017, Dellacasa Bellingegni *et al* 2017, Ottobock 2019, Marinelli *et al* 2020), The use of sEMG in prosthetic applications has become the most widespread source of information about voluntary movement (Schmidl 1965) because of the direct correlation between EMG activity and subjects' intentions.

Differently from the low-density, the high-density sEMG (HD-sEMG) is based on a higher number of electrodes placed on a small portion of the body. Recently, a growing number of researchers has focused on the use of these electrodes aiming to increase the amount of collected data, although at the cost of a greater computational burden. HD-sEMG sensors have been used to discriminate muscular patterns related to different gestures. Their signals can be handled in various ways to retrieve unique and repeatable information, as described in section 4. These sensors have to be positioned according to the distribution of the underlining muscle fibers and this configuration provides a low resolution map of the synergistic activation of the muscles during movement production (Winters 1990, Sartori *et al* 2018). For example, from contraction of the muscles under the acquisition grids, it is possible to extract bi-dimensional images, in which the EMG amplitude is mapped to a color scale. These maps can be thus handled by complex algorithms, as the ones used for objects detection in robotic navigation (Chen *et al* 2020). The main limitation of the HD-sEMG, which currently bounds its application to a laboratory scenario, is the

Table 1. Biosignals used as input sources in prosthetic applications.

		Measured Property	Sensors' placement	PROs	CONs	Sensor Fusion	Examples
Electromyography (EMG)	Surface EMG	Muscle electric potentials	On the skin over targeted muscles 2–32, up to 192 sensors	Non-invasive, long-term use, a large number of people	Sweating, electrodes shift, muscle fatigue, electromagnetic noise	NIRS, IMU, FMG, SMG, MMG	(Merletti *et al* 2010) up to 27 gestures
	Invasive EMG		Underneath the skin, on or inside targeted muscles 4–8 sensors	High signal/noise ratio, directly on the nerve, no shift with respect to the source	Invasive, infections		(Cipriani *et al* 2014, Ortiz-catalan *et al* 2020)
Force-myography (FMG)		Change of muscle morphology measured on the skin surface	Over targeted muscle, over related tendons 8, up to 126 sensors	Physiologic, small size, high signal/noise ratio, flexible	Muscle fatigue, sensors shift, pre-load force, small spatial resolution, crosstalk	EMG	(Xiao and Menon 2019) up to 8 gestures
Mechano-myography (MMG)		Muscle fiber oscillations using microphone or accelerometers	Over targeted muscle 6–20 sensors	Low cost, no pre-amplification, no precise positioning, no skin impedance or sweat influence	Ambient acoustic noise, Adjacent muscle crosstalk, Sensor displacement	EMG, IMU	(Wilson and Vaidyanathan 2017, Guo *et al* 2017a, Castillo *et al* 2020) up to 5 gestures
Sono-myography (SMG)		Change of muscle morphology	Over targeted muscle, over related tendons transducers of different shapes	Deep and superficial muscles, some models are cheap and energy-efficient	Probe shift, tissue impedance, no wireless, some models expensive and bulky	EMG	(Dhawan *et al* 2019) up to 15 gestures
Near-Infrared Spectroscopy (NIRS)		Tissue oxygenation through the amount of scattered light	Over targeted muscle 2–4 sensors	Deep and superficial muscles, high spatial resolution, no electronic interference	Ambient light, Muscle fatigue, tissues heating	EMG, IMU	(Paleari *et al* 2017) up to 9 gestures
Electrical Impedance Tomography (EIT)		Tissue impedance	Over targeted muscle, over related tendons 8, up to 64 sensors	No need precise positioning	Low time resolution, sweating, electromagnetic noise, high consumption	—	(Zhang *et al* 2016, Wu *et al* 2018) up to 8 gestures
Capacitance sensing		Tissue capacitance	Over targeted muscle, over related tendons 3 receiver sensors	Non-invasive, low cost, deep and superficial muscles	Sweating, Electromagnetic noise, displacement, ambient temperature	—	(Cheng *et al* 2013, Truong *et al* 2018) up to 2 gestures
Magneto-myography		Magnetic fields generated by muscle	Over/inside targeted muscle 7 sensors	Not sensitive to sensor's shift and sweat	Magnetic interference, can be invasive, movement artifacts	—	(Zuo *et al* 2020) concept
Peripheral neural interfaces (PNIs)		Electrical activity of the nerves	Microelectrode arrays placed on different fascicles within the median and ulnar nerves	Intuitive, direct maps of complex movements, high accuracy, robust	Invasive, difficult to separate EMG and PNI components, recording channels really closed each other	—	(Nguyen *et al* 2020) up to 15 DoFs

(Continued.)

Table 1. (Continued.)

	Measured Property	Sensors' placement	PROs	CONs	Sensor Fusion	Examples
Intracortical neural signals	Intracortical neural signals from the brain, action potentials of individual neuron	16–192 high-density channels electrodes inserted into the motor cortex tissue	Accurate and capable of collecting the most information-rich data, high spatial resolution	Very invasive, influenced by tissue reactions	—	(Hochberg *et al* 2006, Hochberg *et al* 2012, Collinger *et al* 2013, Wodlinger *et al* 2014) 7–10 DoFs
Electrocorticography (ECoG)	Electrical activity of brain's surface	32–128 high-density channels on sensorimotor regions	Less attenuated than EEG, good spatial resolution and wide frequency content	Surgical procedure and lack to measure single cell activity	—	(Wang *et al* 2013, Fifer *et al* 2013, Bleichner *et al* 2016, Hotson *et al* 2016) 4 gesture recognition and wrist movements
Electroencephalography (EEG)	Electrical activity of the brain	6–32 channels headsets	Not invasive, low cost, portable, stable, and very easy to use	Signal attenuated by the dura, the skull, and the scalp, loss of important information	—	(Mcfarland *et al* 2010, Yang *et al* 2012, Edelman *et al* 2019, Fuentes-Gonzalez *et al*, 2021) single Degree of Freedom (DoF)
Functional near-infrared spectroscopy (fNIRS)	Activity-related brain oxygenation, near-infrared led, and a photodetector measure the amount of IR light absorbed by the hemoglobin in the brain	10–200 channels of optodes	Non-invasive, simultaneous detection information under the skin, low cost	Few centimeters penetration of cortical tissue, not great accuracy, and system too cumbersome	—	(Syed *et al* 2020) 3 DoFs trans-humeral amputees

skin-electrode contact since it requires conductive gel to reduce the interface impedance. The wet area is mainly needed to reduce artifacts in the EMG signals since it is generally acquired in monopolar configuration. Another disadvantage of this technique consists in the fact that computation is time-consuming.

Overall, the main drawback of sEMG-based approaches is constituted by the influence that skin impedance, sweat, and electrode shift have on the stability of the input signals (De Luca 1997). Additionally, muscle crosstalk and the difficulty to reach deep muscles further limit the quality of the collected signal. In the context of ULP, the use of sEMG can be further complicated by the fact that the amputation strongly affects muscles strength and organization and therefore signal quality, as discussed in section 6.4.

3.1.1.1.2. Invasive EMG and surgical procedures
The invasive approach has been exploited to explore the activity related to the production of movement for many years (Adrian and Bronk 1929) and it is still investigated by many groups. However, the main drawback of this approach is constituted by the surgery and by the technological barriers still faced by the available equipment. On the other hand, invasive electromyography (iEMG) allows to measure single motor unit action potentials, enabling a higher selectivity and a better accuracy of the input signal, overcoming the limitations imposed by sEMG. There are several examples of iEMG, which vary in the type of electrodes and level of invasiveness, as detailed hereafter.

EMG can be invasively detected by inserting electrodes into the internal surface of muscles (Merletti and Farina 2009). This invasive technique exploits two different percutaneous electrodes: needles and fine wires

Figure 4. sEMG electrodes. (A) Bipolar dry sensors, Ottobock and IIT/INAIL (Marinelli *et al* 2021) respectively. (B) High-density wet sensors (OT Bioelettronica).

(Jamal 2012, Rubin 2019). The most used are needle electrodes. These electrodes are concentric, and their bare hollow needles contain an insulated fine wire into their cannula, which is exposed on the beveled tip, which is the active recording site. Wire electrodes are typically made of non-oxidizing and stiff materials with insulation, they can be implanted more easily and are usually less painful than needle electrodes.

Since both these sensors are percutaneous, i.e. passing through unbroken skin and leaving an open passage between the internal structures of the body and the external world, the risk of infection is quite probable. For this reason, and because of their intrinsic discomfort due to the percutaneous wire that can easily break, their usage is limited to laboratory research (Hargrove *et al* 2007, Cloutier and Yang 2013a). A detailed description of invasive electrodes both to record biological signals and to deliver electrical stimulation can be found in Raspopovic *et al* (2021a).

In the last decades, growing attention has been paid to the development of intramuscular electrodes that could be implanted under the skin of the subject to achieve the advantages of invasive sensors and simultaneously avoid the risks and inconvenience of percutaneous instruments. For example, Weir *et al* (2008) developed an implantable myoelectric sensor (IMES), a system able to receive and process up to 32 implanted sensors with wireless telemetry. A transcutaneous magnetic link between the implanted electrodes and the external coil allows reverse telemetry, which transfer data from the sensors to the controller, commanding the control of the prosthesis, and forward telemetry to supply power and configuration settings to the electrodes. These sensors are designed for permanent long-term implantation without any kind of servicing requirement and have been tested on animals. Four months after the implantation of IMESs in the legs of three cats, the sensors were still functioning (Weir *et al* 2008). Intramuscular electrodes have been used in prosthetic application to decode 12 different hand gestures from 4 healthy subjects (Cipriani *et al* 2014). Moreover, it has been shown that the application of this invasive approach enhances the simultaneous control of multi-DoFs system (Smith *et al* 2014).

Recently, the group of Ortiz-Catalan showed an invasive procedure for ULP control. They positioned EMG electrodes under the skin of amputated subjects and sutured them directly on the external surface of the muscles (Ortiz-Catalan *et al* 2020). More precisely, sensors were sewn onto the epimysium of the two heads of the biceps' muscles and the long and lateral heads of the triceps muscles. These invasive electrodes were used in combination with an osseointegrated prosthesis, i.e. a system obtained following a very invasive surgical procedure, which allows to anchor the prosthesis to the remaining limb's bone (Ortiz-Catalan *et al* 2020). In the context of ULP, osseointegration is offered for trans-humeral amputees, and the prosthesis is anchored to the humerus with two mechanical elements: the fixture, a screw made of titanium placed inside a hole made in the bone that becomes osseointegrated, and the abutment, placed within the fixture and extending outside of the body in a percutaneous way, onto which the prosthesis is connected. This technique was tested on four osseointegrated patients.

This latter example indicates that also surgical approaches can be taken to improve the quality of the collected EMG. A promising surgical technique that is performed in case of high-level amputation is targeted muscle reinnervation (TMR). This method was developed by the group of Kuiken in the early 2000s and consists in transferring residual arm nerves to alternative muscle sites. Following reinnervation, these target muscles are able to produce EMG that can be collected and used to control prosthetic arms (Kuiken *et al* 2009). This strategy works at the condition that each reinnervated muscle produces an EMG signal in response to only one transferred nerve, with the consequence that native nerves innervating the target muscle has to be cut during the surgical procedure to avoid unwanted EMG signals (Kuiken *et al* 2017). In the last 15 years, TMR has allowed intuitive control of ULP to several subjects with high-level amputation for

whom standard ULP devices allowed a poor restoration of motor functions (Kuiken *et al* 2017). Importantly, given that it is performed on complex amputations, this technique is strongly tailored to each patient's physical and clinical status (Cheesborough *et al* 2015, Mereu *et al* 2021).

Recently, a new surgical method for improving EMG-based control has emerged: the regenerative peripheral nerve interface (RNPI) (Vu *et al* 2020a). Just as TMR, its goal is to turn a muscle into a biological amplifier of the motor command, in order to improve the quality of the EMG signal recorded, processed and used to drive the prosthesis. To this end, RNPI exploits the regeneration capabilities of nerves and muscles, to implant a transected nerve into a free muscle graft. Following regeneration, revascularization and reinnervation by the transected nerve, the muscle graft effectively becomes a stable peripheral nerve bioamplifier, able to produce high-amplitude EMG signals (Urbanchek *et al* 2012). The potential of this novel interface has been tested by Vu *et al* (2020b): they used EMG signals collected by intramuscular bipolar electrodes implanted into RNPIs obtained in amputated individuals, who could successfully perform real-time control of an artificial hand. Surprisingly, subjects were able to control the device with a high level of accuracy even 300 d post-implantation, without recalibration of the control algorithm.

Another surgical technique, not directly related to EMG signals but worth mentioning, is cineplasty, an old method revived in the last years with a new and more modern approach. This method was introduced for the first time by Vanghetti in 1899 and then replicated by Sauerbruch ten years later (Tropea *et al* 2017). It consisted of the direct mechanical linking of residual muscles and/or residual tendons of the affected limb to the prosthesis through external cables (i.e. Bowden cables). In 2001, Heckathorne and Childress (2001) implemented an evolution of this surgical solution for the control of one DoF ULP by exploiting exteriorized tendons directly linked to a force sensor.

3.1.1.2. Other biosignals

The limitations imposed by the use of EMG (either invasive or non-invasive), have led researchers to study new approaches, aiming at increasing algorithms robustness and accuracy. Some may be soon used in commercial prosthetic systems, while others represent promising research scenarios, but still far from real-life applications. We here describe some of these peripheral signals, both non-invasive and invasive.

For example, forcemyography (FMG) has been widely investigated in the past 20 years (Xiao and Menon 2019) (table 1). This approach is based on force sensors able to record muscle stiffness around the forearm during different movements. The muscle deformation of the stump can be measured with various types of sensors, such as: force sensing resistors (Prakash *et al* 2020), optical fiber transducers (Fujiwara *et al* 2018), capacitance-based deformation sensors (Truong *et al* 2018), Hall-effect based deformation sensors (Kenney *et al* 1999), barometric sensors (Shull *et al* 2019), thin arrays of adhesive stretchable deformation sensors (Jiang *et al* 2019), or high density myo-pneumatic sensors for topographic maps of pressures and residual kinetic images of the stump (Phillips and Craelius 2005, Radmand *et al* 2016). The accuracy of the sensors may limit the robustness of FMG-based control. Therefore, FMG is often fused with other input sources, such as inertial measurement unit (IMU) (Ferigo *et al* 2017) or EMG (Nowak *et al* 2020). FMG is indeed complementary to EMG due to its capability to get information about extrinsic hand muscles placed in several layers underneath the skin, and therefore difficult to be detected with the EMG sensors. Moreover, with respect to EMG-based control strategies, FMG is not influenced by electrode shifting.

Another technique is mechanomyography (MMG), which measures the lateral oscillations, detected as low-frequency vibrations (in the range of 1–100 Hz), generated by deformation in muscle fibers actively involved in the contraction (table 1). This approach can be considered as the mechanical counterpart of EMG and it is also known as acousticmyography, phonomyography or vibromyography, depending on the type of sensor used. It can actually be based on different types of sensors, such as: low mass accelerometers (Farina *et al* 2008, Youn and Kim 2010), microphones (Castillo *et al* 2020, Meagher *et al* 2020), piezoelectric contact (Orizio *et al* 2008, Tanaka *et al* 2011), force sensing resistors (Esposito *et al* 2018), and laser distance sensors (Scalise *et al* 2013). With respect to EMG, this technique shows some advantages: it is low cost, it does not require pre-amplification or precise positioning, and signals are not influenced by skin impedance or sweat. However, it is very susceptible to environmental noise and motion. Artifact removal can be implemented with the integration of an IMU, as proposed by (Wilson and Vaidyanathan 2017) and (Woodward *et al* 2017). MMG has also been used in combination with EMG signals (Guo *et al* 2017a), achieving better control performance and robustness.

The sonomyography (SMG) measures muscle volume changes and thickness using reflected ultrasound waves (table 1). Wave amplitude depends on the acoustic impedance of the tissue, and it can be detected using ultrasound transducers. Currently, no portable prosthetic systems based on SMG have been developed, but the results obtained using this technique are very promising. For example, Dhawan *et al* (2019) were able to detect 11 different movements in real-time placing the sensor on the stump of a trans-radial amputee, obtaining better results than using EMG signals alone. This non-invasive approach allows a faster user

training and the detection of both superficial and deep muscles, but even a small shift of the sensor can change the cross-section view and bring to the failure of the control algorithm. SMG signals have been used in combination with EMG signals, leading, to improved performances with respect to EMG alone (Xia *et al* 2019, Engdahl *et al* 2020a).

Near-infrared spectroscopy (NIRS) is a non-invasive technique measuring the level of oxygenation of active muscles under contraction (table 1). The detection unit consists of a near-infrared led emitter and a photodetector, placed on the skin surface. The emitted infrared (IR) light is partly absorbed by the tissue, mostly by hemoglobin, and partly scattered back to the skin surface and detected by the photodetector. NIRS thus detects changes in the amount of IR light scattered back due to muscle contraction (Schneider *et al* 2003). This technique has a high spatial resolution and is immune to electronic interference. However, tissue heating may take place after prolonged use. Recently, Paleari *et al* (2017) developed a wireless NIRS unit for hand gesture recognition, indicating the potentiality of this technique for ULP control. NIRS has indeed been used in this context in conjunction with EMG (Guo *et al* 2017b) and IMU (Zhao *et al* 2019).

The electrical impedance tomography (EIT) measures the internal electrical impedance of the tissues in the cross-section plane covered by specific surface electrodes (table 1), which may range from 8 to 64 (Padilha Leitzke and Zangl 2020). The measurement is executed by exciting a sine wave of electrical current (amplitudes ranging from 10 μA to 10 mA and frequencies from 10 kHz to 1 MHz (Grushko *et al* 2020)) and by recording the voltages collected by surface electrodes. The detected changes in phase and amplitude represent the distribution changes of internal conductivity within the affected area, identifying patterns of movement. Wearable systems for ULP control have been developed, such as the ones proposed by Zhang *et al* (2016) capable to recognize hand gestures, and by Wu *et al* (2018), who also tested an EIT-based hand prosthesis control system on healthy people, achieving an accuracy of 98.5% with a grouping of three gestures and an accuracy of 94.4% with two sets of five gestures. This non-invasive method does not require a precise positioning of the electrodes, it only needs changes in impedance to be large enough. On the other hand, the current available systems have slow measurement and long processing time, leading to a high-power consumption. Moreover, the technique is affected by surface electrodes issues, namely skin contact conditions, electromagnetic interference, etc.

Capacitance sensing measures capacitance variations between two or more conductors (table 1). A capacitance exists when the two conductors are separated by a given distance *d*. In ULP context, electrodes may be placed on the prosthetic fingers, which work as capacitor plates. When a user performs a gesture, the skin deformation will cause a change in distance (*d*) between the conductors. This technique was used for hand gesture prediction in (Cheng *et al* 2013) and in (Truong *et al* 2018), using wearable systems. This technique is low cost, non-invasive, and it is capable to detect deep and complex signals, but it owns the standard disadvantages affecting surface electrodes, and it is susceptible to ambient temperature changes.

Magnetomyography is a promising approach aimed at measuring the magnetic fields produced by electrical currents propagating through muscles during contraction (table 1). This technique foresees the placement of magnetometers on the muscle, either non-invasively or beneath the skin, following a surgical procedure. The magnetometers convert the magnetic fields into measurable quantities, such as currents or voltages that can be used for the control of the prosthesis. Small implantable magnetometers have been proposed in Zuo *et al* (2020), but they still need to be clinically tested. This technique is less sensitive to sensors' shift or sweat but may be strongly influenced by the environmental magnetic noise and the magnetic field of the Earth.

Peripheral neural interfaces measure the electrical activity of the motor peripheral with an invasive approach (table 1). There are three types of electrodes: extraneural, like CUFF or flat interface nerve electrodes (FINEs), which embrace the nerve; intraneural, which run longitudinally (longitudinal intrafascicular electrodes (LIFEs)) or transversally (transverse intrafascicular multichannel electrodes (TIMEs) or Utah slanted electrode array (USEA)) through the nerve; and regenerative, such as SIEVE or Microchannel, attached between the two extremities of a severed nerve (del Valle and Navarro 2013, Raspopovic *et al* 2021b). Nguyen *et al* (2020) enabled an amputee to control a 15 DoFs prosthesis, by using 4 implanted LIFE arrays, 2 in the medial nerve and 2 in the ulnar one. The negative aspects of this method reside in its profound invasiveness and in the acquisition of noisy signals.

3.1.1.3. Brain signals

The first neuroprosthetic application on humans was reported by the group of Donoghue, who demonstrated that tetraplegic individuals implanted with arrays of microelectrodes over the motor cortex were able to remotely control the movement of a cursor on a screen (Hochberg *et al* 2006). This clinical trial was soon followed by another from the same group reporting the control of reaching and grasping actions of a robotic arm (Hochberg *et al* 2012). The group of Schwartz also showed similar results of an individual with tetraplegia successfully controlling a seven DoF robotic arm (Collinger *et al* 2013). In all these examples,

intracortical brain signals were used, i.e. action potentials of individual neurons were detected with an array of electrodes inserted into the brain, usually in the motor cortex (table 1).

Less invasive measurements of cortical currents using electrocorticography (ECoG) have been widely used for neuroprosthetic control in the lab. ECoG detects the electrical activity of the brain with strips of electrodes laid on the brain's surface, usually in the motor cortex area. ECoG signals have been used for hand gesture recognition (Bleichner *et al* 2016), for the control of a virtual prosthesis (Wang *et al* 2013) and of a robotic limb (Fifer *et al* 2013), and also with a detached prosthesis with active digits (Hotson *et al* 2016).

ECoG provides an ideal trade-off between the invasiveness of intracortical recordings and the poor spatial resolution of electroencephalography (EEG) (Thakor *et al* 2014). However, whether non-invasively collected signals convey enough motor information to control a neuroprosthetic hand is still debated (Fukuma *et al* 2016).

EEG measures the electrical activity of the brain with an external helmet made of electrodes (table 1). In a ULP application, a motor imagery task is typically used, and the subject only needs to think about the movement. EEG signals corresponding to the intention of the movement are therefore used to drive the end-effector. Recently, McDermott *et al* 2021 were able to extract from EEG recordings relevant brain states in real-time and indicated such states as prospective therapeutic targets for motor neurorehabilitation (McDermott *et al* 2021). Similarly, the group of Wolpaw showed that paralyzed patients could use EEG signals to control a cursor in three-dimensional space (Mcfarland *et al* 2010), suggesting that noninvasive EEG-based BCIs can be exploited for control of robotic devices or neuroprostheses. EEG-based neuroimaging is indeed emerging as a useful tool for robotic device control, as demonstrated by Edelman *et al* (2019).

Another promising technique is the functional near-infrared spectroscopy (fNIRS), which detects activity-related brain oxygenation. The instrumentation is the same used for NIRS, i.e. a near-infrared led, and a photodetector are used to measure the amount of scattered back light and, therefore, the amount of IR light absorbed by the hemoglobin in the brain, which increases during brain activity. In 2020, Syed *et al* (2020) used these hemodynamic brain responses to control a ULP for trans-humeral amputees with three DoFs, gaining eight out of ten classified movements in real-time.

These examples demonstrate that groundwork for brain control of motor prosthetics has been laid. However, it has been limited to the lab and mostly addresses paralyzed patients. Nevertheless, there is a growing interest in brain-derived measures for prosthetic applications and different recording techniques have been investigated for ULP control.

3.1.2. Other techniques under investigation

Besides the detection of physiological changes in residual muscles of the stump or in the brain during movements, described above, there are many other input sources and techniques capable or with the potentiality to control an upper limb prosthesis. Some of these are mainly used in the research field and since they lack usability, they do not find a real application in everyday life of amputees, or they are conceived for patients without the possibility to exploit other more convenient and intuitive sources (i.e. tetraplegic people). Some of them, instead, have been still only proposed as proof-of-concept.

The most studied approach is based on the use of IMUs. IMU sensors are cheap, small and can therefore by easily embedded in the prosthesis. They can increase the amount of data useful to successfully discriminate between different gestures of ULP during distinct phases of the reaching movement. These devices exploit accelerometers, gyroscopes and magnetometers to understand which is the actual altitude, position and orientation of the prosthesis. These sensors deliver information through quaternions and they are often used together with EMG to improve the classifier robustness (Georgi *et al* 2015). Zhang *et al* (2011) depicted the possibility to manipulate objects and perform complex tasks using both IMU and EMG sensors. As a matter of fact, the accelerometers can capture information that sEMG sensors cannot easily detect, such as hand withdrawal or rotation (Chen *et al* 2007). It has been shown that the use of IMU sensor coupled to EMG is more advantageous than increasing the number of EMG sensors (Fougner *et al* 2011). Similar results have been achieved by Krasoulis *et al* (2017), who have combined EMG and IMU to feed pattern recognition systems (see section 4.3.1). They demonstrated that this combination could significantly improve the real-time completion rates compared to the traditional methods, exclusively based on sEMG signals. Moreover, the data coming from IMU can be used alone to control a single module, usually the wrist or the elbow (Merad *et al* 2018) or to realize other types of control for rehabilitation purposes, such as shadow control, in which the control policy consists in replicating the movement captured by the IMU sensors (Rapetti *et al* 2020). These devices were also placed on feet to directly control an ULP by implementing precise foot movements (Resnik *et al* 2014). The adoption of IMU sensors is specifically promising in sensor fusion approaches, as discussed in section 3.1.3. Besides EMG signals, IMU data have been also combined with NIRS (Zhao *et al* 2019) and MMG (Wilson and Vaidyanathan 2017, Woodward *et al* 2017).

Table 2. Alternative input sources investigated for the control multi-DoF prosthesis devices.

Input source	Measured property	Sensors' placement	PROs	CONs	Sensor fusion	Examples
IMU	Specific force, angular rate, orientation of the body	Up to 8 IMU sensors located on feet	Non-invasive, simple, low cost	Problems during walking, not intuitive, unnatural	EMG, NIRS, MMG	(Resnik *et al* 2014) DEKA Arm control
Myokinetic control	Change of muscle morphology trough magnetic fields	Permanent magnet markers implanted over targeted muscles and external three-axis magnetic field sensors placed in the socket	Intuitive control, force and position feedback	Magnetic interferences, misalignments between socket and initial position, invasive	—	(Tarantino *et al* 2017, Clemente *et al* 2019)
Voice	Throat vibration Voice commands	Piezoelectric sensor on the throat Microphone near mouth	Ease of use sequence of movements	External noise, input sound level, unintuitive control	EMG IMU	(Mainardi and Davalli 2007) (Alkhafaf *et al* 2020)
Tongue	Pressures made by the tongue	Board of coils on the palate and activation unit on the tip of the tongue	Mobile, wireless, invisible	Unintuitive control, uncomfortable	EMG	(Johansen *et al* 2016, Johansen *et al* 2021)
Feet	Pressures made by the feet	Insole made of force sensing resistors	Simple low cost	Unintuitive control, problem during walking, need of accurate calibration	IMU EMG	(Carrozza *et al* 2007)
Optical myography (OMG)	Skin surface deformations caused by underlying muscle contraction	Single low-resolution camera and marker-based tracking methods	Simple low cost	No space for camera in the socket, low robustness	—	(Nissler *et al* 2016, Wu *et al* 2019)

Table 2 summarizes the use of IMU and other input sources investigated for the control of ULP, many of which are described in (Grushko *et al* 2020).

3.1.3. Integrative sources
An integrative input source is not used as the main responsible for the actual command of the prosthesis, but it is used to help and to facilitate its control, which usually depends on EMG signals. The integrative input sources work in parallel and together with the main ones, integrating their information and implementing the so-called data-fusion or sensor-fusion methods, see also section 4.3.3. Table 3 summarizes integrative sources found in the literature that have been used for ULP control.

3.2. Prosthetic sensing
Natural movements occur with a bidirectional flow of neural information, i.e. motor commands on one direction and sensory feedback on the other. In prosthetic applications, while many efforts have been spent to provide signals carrying motor intentions, a less explored path is the integration of sense of touch into the prosthesis (Clemente *et al* 2015). This lack is highly responsible for the missing perception of the prosthesis as part of one's own body and is also precluding a closed-loop control of the prosthesis.

More recently, the scientific community has started exploring different methods to equip prosthetic devices with perception of tactile and pressure information (Schmitz *et al* 2008, Tee *et al* 2012, Hammock *et al* 2013, Lucarotti *et al* 2013, Taunyazov *et al* 2021), although often resulting in very complex, unreliable, or unpractically cumbersome solutions. The few solutions tested on real prosthetic setups impacted on their anthropomorphism and dexterity.

To integrate touch sensors into robotic and prosthetic devices (figure 2, end-effector feedback) (Lucarotti *et al* 2013, Iskarous and Thakor 2019, Dimante *et al* 2020), different technologies have been investigated and employed (Ciancio *et al* 2016), namely capacitive (Jamali *et al* 2015, Maiolino *et al* 2013), resistive (Beccai *et al* 2005, Tee *et al* 2012, Zainuddin *et al* 2015), piezoelectric (screen printed piezoelectric polymer, PVDF)

Table 3. Integrative sources of information used to improve prosthesis control.

Integrative input source	Instruments and measured information	Application	PROs	CONs	Fusion	Examples
Computer vision	Two cameras used to collect images and estimate depth Depth estimated by the color intensity of the pixel collected by the camera	Estimation of size, distance and grasp type for a semi-autonomous control of the prosthesis	Ease of use fixing of errors without looking at the prosthesis automatic help in controlling the prosthesis	Expensive, cumbersome and uncomfortable	EMG IMU	(Markovic *et al* 2014) Stereovision (depth?) (Mouchoux *et al* 2021) Depth and color camera RGB
Eye movements	4 Superficial electrodes for the measuring of the corneo-retinal standing potentials between the front and the back of the human eye Camera mounted on a pair of glasses measuring the reflection of infra-red (IR) light from the eyeball	Estimation of the position/length/ width/ orientation of a final target and preparation of the preshape and direction of the hand	Ease of use automatic help in controlling the prosthesis	Distinction with random eye movements, cumbersome and uncomfortable	EMG	(Hao *et al* 2013) Electro-oculography (Krausz *et al* 2020) Eye tracking glasses
Optical sensor	Led-based optical sensor mounted on fingertips Miniature reflective optic sensor that combines an Infrared LED and a phototransistor in the same package.	Slip detection and eventual automatic suppression	Accurate, robust simple, low cost and low power consumption	Poor detection with transparent surfaces	EMG	(Sani and Meek 2011) LED motion detection sensor (Nakagawa-Silva *et al* 2018) Reflective optic sensor
IMU	Accelerometers, Gyroscopes, Magnetometers	Decreased #sensors, better controllability, artifact detection	Non-invasive, simple, low cost, motion artifact deletion	Prone to error cumulate over time	EMG, NIRS, MMG	(Krasoulis *et al* 2017, Krasoulis *et al* 2019b) up to 6 gestures

(Alameh *et al* 2018), and magnetic sensors (Ahmadi *et al* 2011). Other examples include technologies based on electrical impedance (Zainuddin *et al* 2015, Wu *et al* 2018), pressure and electrical impedance (Lin *et al* 2009), optical fibers (Bragg fiber (Massari *et al* 2019)), micro-electro-mechanical systems (texture sensing (Mazzoni *et al* 2020)) combined with spiking based on Izhikevich neuron model (Gunasekaran *et al* 2019) and optoelectronic (Alfadhel and Kosel 2015).

Examples of the application of these sensors into prosthetic devices include the E-dermis (piezoelectric sensors integrated on the Bebionic's fingertips) (Osborn *et al* 2018), E-skin (integrating different types of sensors) (Iskarous and Thakor 2019), and BioTac (impedance sensor integrated on the Shadow Hand (Robot 2022)) (Fishel and Loeb 2012).

Among commercial devices, the SensorHand Speed (Ottobock 2021) made by Ottobock is the only one including tactile sensors based on resistive technology (Ottobock 2021).

Therefore, tactile sensation is the first step toward novel and more efficient control strategies that do make use of feedback information (Raspopovic *et al* 2014). To this end, artificial intelligence can be exploited to detect the grasp of different objects from sensor data (Alameh *et al* 2020).

3.3. Sensory feedback

Sensory feedback patterns are designed to enrich the perceived responsiveness of the device and the subjective experience of its use as a limb (Antfolk *et al* 2013b, Svensson *et al* 2017, Raspopovic *et al* 2021a). Such a result derives from the elicitation of physiological and psychological reactions that promote embodiment processes (described in section 5.1). Furthermore, such stimulations (haptic feedback in many

cutting-edge devices) are designed as a fundamental component of bidirectional human-machine interfaces empowering prosthetic control (Navaraj *et al* 2019). Establishing such a closed-loop can trigger learning processes even for artificial sensations (Cuberovic *et al* 2019), pointing at somatosensory plasticity processes. These phenomena provide the user with an engaging guidance within a natural interaction, facilitating the execution of prosthetic maneuvers during calibration, training, and daily use. Importantly, such an enhanced practice will ease the production of consistent biosignals that will progressively become easier to interpret as user commands.

However, current commercial prostheses generally do not incorporate an explicit haptic feedback but the incidental feedback, like visual and the sound cues, could be exploited by the user to estimate the prosthesis state (Wilke *et al* 2019). For example, the acoustic feedback provides a guidance on how to reach target during the rehabilitation session, in this way the rehabilitation step can be more interactive and engaging if appropriately designed (never obnoxious, possibly plausible). Overall, the next sub-sections will discuss the design of sensory feedback in prosthetics, distinguishing invasive and non-invasive stimulation modalities.

3.3.1. Non-invasive methods

Non-invasive feedback restoration for upper limb amputees is a hot topic in the research community, and yet it has not achieved broad clinical application (Sensinger and Dosen 2020). Many solutions have been proposed, but the main problem lays in their poor robustness. (Ribeiro *et al* 2019) highlighted the most widespread types of non-invasive feedback, described in table 4.

The most investigated feedback relies on the sense of touch and therefore consists of cutaneous stimulation. This can be performed with different modalities namely, vibrational, mechanotactile or electrical stimulation.

The vibrational feedback is generally implemented with the addition of eccentric rotating motors placed in contact with the skin surface of the stump (Ribeiro *et al* 2019). This method is generally employed to augment the robustness of the control system by providing the user with additional information regarding the position of the prosthetic device but it lacks intuitiveness, as the association between perceived sensation and the corresponding information has to be learned by the user. For example, in Bark *et al* (2014), the motors were placed in four distinct areas of the stump to guide the user through the desired trajectory while grasping object and the results showed a significant decrease in the root mean square angle error of their limb during the learning process. More recently, Markovic *et al* (2019) proposed a joint-oriented feedback criterion consisting of three vibromotors placed on the arm to provide the information on which joint is currently activated by the user, thus restoring proprioceptive sensation. The experiment was performed by 12 able-body subjects and 2 amputees controlling 3 DoF prosthesis, and it was found that the myoelectric multi-amplitude control outperformed the pattern recognition method when the feedback was applied.

Differently from the vibrational, the mechanotactile feedback is based on the application of linear actuators on the skin and provides pressure sensation. Antfolk *et al* (2013a) exploited this technique and proposed a multisite mechanotactile system to investigate the localization and discrimination threshold of pressure stimuli on the residual limbs of trans-radial amputees. They demonstrated that subjects were able to discriminate between different location of sensation and to differentiate between three different levels of pressure. This study demonstrated that it is possible to transfer tactile input from an artificial hand to the forearm skin after a brief training period. Recently, Svensson *et al* (2017) used it to translate the interaction between a virtual reality (VR) environment and a virtual hand into user sensation. The authors showed that by placing the tactile actuators in correspondence with the areas of the skin involved in object manipulation, subjects were able to feel a real touch sensation that increased their sense of body ownership. For example, pressure applied to the prosthetic fingers was perceived as a tactile sensation on the skin (Svensson *et al* 2017).

The electrical feedback is based on transcutaneous stimulation. The elicited sensations range from perception of pressure (Jorgovanovic *et al* 2014) to slip sensations (Xu *et al* 2015), depending on the electrical parameters (i.e. current amplitude, pulse frequency, pulse width). One advantage of this approach with respect to the vibrotactile and mechanotactile ones is the lack of moving components avoiding problems of electrode displacement and, thus, improving the sensors-skin contact. Nevertheless, it is important to take into account that the noise introduced by the electric stimulation can corrupt the acquisition of muscular activity, causing errors if the ULP is myoelectrically controlled. Moreover, the perceptions are not strictly confined to the zone under the stimulating device but they can spread in a wider region if the area above a nerve is considered.

Another sensory modality exploited for feedback delivery is the acoustic one. Gigli *et al* (2020) recently tested a novel acquisition protocol with additional acoustic feedback in 18 able-body participants to improve myoelectric control. The protocol consisted in dynamically acquiring EMG data in multiple arm positions while returning an acoustic signal to urge the participants to hover with the arm in specific regions of their peri-personal space. The results showed that the interaction between user and prosthesis during the data

Table 4. Non-invasive methods for sensory feedback in ULP.

Feedback sense		Instruments and feedback information	Application	PROs	CONs	Examples
Touch (cutaneous stimulation)	Vibrational	Eccentric rotating motors, proprioception, force	Array over the forearm or over the arm	Non-invasive, robustness control, brief training period, intuitive, cheap, small	(Bark *et al* 2014, Markovic *et al* 2019) up to 3 DoFs or different force levels	Non-physiological, need calibration, coupled intensity and rotation frequency, position displacement
	Mechanotactile	Linear actuator, pressure sensation, spatial touch sensation	Detected areas to reproduce real touch sensation, array over the arm	Non-invasive, intuitive, brief training period, decoupled intensity and frequency	(Antfolk *et al* 2013a, Svensson *et al* 2017, Tchimino *et al* 2021) different pression level, touch sensation	Need spatial and intensity calibration, bulky, position displacement
	Electrical	Transcutaneous stimulation using bipolar electrodes, pressure, slip, proprioception	Array over the forearm or arm	No electrode displacement, low power consuming, high sensor skin contact, intensity or frequency modulation	(Jorgovanovic *et al* 2014, Xu *et al* 2015, Garenfeld *et al* 2020) touch location, pression, proprioception	Noise during acquisition, long calibration, not localized sensation
Sound (acoustic)		Acoustic speaker, proprioceptive movements	Laptop speaker to guide the training acquisition and improve the pattern recognition strategy	Low cost, no calibration, intuitive	—	(Gigli *et al* 2020) multiple arm positions
Vision (visual)		Camera on board, external camera	Head-mounted displays, laptop displays, virtual reality, augmented reality	Increase perceptual experience, engagement, intuitive, promote training	Bulky, not portable, uncomfortable	(Clemente *et al* 2016, Markovic *et al* 2017, Sharma *et al* 2018, Hazubski *et al* 2020, Sun *et al* 2021b) trajectory, force

acquisition step was able to significantly improve myoelectric control. Auditory feedback has also been employed to convey artificial proprioceptive and exteroceptive information. Lundborg *et al* (1999) and Gonzalez *et al* (2012) employed auditory feedback by encoding the movement of different fingers into different sounds. The method demonstrated that the inclusion of auditory feedback reduces the mental effort and increase the human-machine interaction; furthermore, better temporal performance and better grasping performance were obtained.

In the last years, there have been some examples exploiting vision to deliver sensory feedback. Indeed, visual stimulation can be provided as explicit feedback through screens during game-like exercises, helping the prosthetic user to learn how to control the device (e.g. adjusting trajectory or grasping force) (Markovic *et al* 2018). However, adding sensory information to the prosthetic user's perceptual experience in real contexts requires solutions like augmented reality (AR, occurring when computer-generated items overlay a real setting) or mixed reality (MR, a term that represented different combinations of real and virtual items) (Milgram and Kishino 1994, Speicher *et al* 2019). AR and MR environments, implemented through wearable solutions like head-mounted displays, can support the actual control of a prosthetic device through visual feedback that does not occlude the real context (Clemente *et al* 2016, Markovic *et al* 2017, Hazubski *et al*

2020). However, they can also be used for prosthetic use training (Anderson and Bischof 2014, Sharma *et al* 2018)—in such a case, VR (a fully computer-generated setting) can offer visual feedback too (Lamounier *et al* 2010, Sun *et al* 2021b), especially within game-based frameworks (Nissler *et al* 2019) for engaging the users and motivating their activity.

3.3.2. Invasive methods

There are different technologies that can be employed to provide a sensation directly to the nerve (Cutrone and Micera 2019, Raspopovic *et al* 2021a). The most used employ intrafascicular electrodes, such as TIME and wire and thin-film LIFE, which can both record muscle activity (e.g. iEMG) and stimulate nerves. Other solutions are characterized by the fact that the electrodes are placed around the nerves, such as cuff electrodes and FINEs.

The first example of ULP with sensory stimulation dates back to 1979 and it was based on the remapping between pressure signals acquired by prosthesis sensors to an amplitude-frequency modulation. This consisted of a series of pulses delivered with a pulse rate proportional to the increment of the pinch force and provided through dry electrodes placed over the skin in correspondence of the median nerve, as described in Shannon (1979). Later, the group of Micera employed thin-film intrafascicular electrodes longitudinally implanted in peripheral nerves (tf-LIFE4) to deliver electrical stimulation. With this method, they were able to elicit sensation of missing hand in the fascicular projection territories of the corresponding nerves and to modulate the sensation by varying the pulse width and pulse frequency (Benvenuto *et al* 2010). Importantly, this method avoids muscle crosstalk, fundamental for guaranteeing myoelectric control. More recently, new bioinspired paradigms have been suggested to better induce natural sensations (Raspopovic *et al* 2021a). In particular, the study of Oddo *et al* (2016) showed that it is possible to restore textural features recorded by an artificial fingertip. This device embedded a neuromorphic real-time mechano-neuro-transductor, which emulated the firing dynamics of SA1 cutaneous afferents. The emulated firing rate was converted into temporal pattern of electrical spikes that were delivered to the human median nerve via percutaneous microstimulation in one trans-radial amputee.

Valle *et al* (2018) suggested a 'hybrid' encoding strategy based on simultaneous biomimetic frequency and amplitude modulation. This kind of stimulation was perceived more natural with respect to classical stimulation protocol, enabling better performance in tasks requiring fine identification of the applied force. This paradigm was tested and validated during a virtual egg test (Valle *et al* 2018), where the subject needed to modulate the force applied to move sensorized blocks. This encoding strategy not only improves gross manual dexterity in functional task but also improved the prosthesis embodiment, reducing abnormal phantom limb perceptions.

Similarly, Osborn *et al* (2018) implemented a neuromorphic feedback paradigm based on Izikevich neuron model to generate the current spike train to inject directly in the median and ulnar nerves, using beryllium copper (BeCu) probes. Their prosthesis proposes a neuromorphic multilayered artificial skin to perceive touch and pain. Their transcutaneous electrical nerve stimulation allows to elicit innocuous and noxious tactile perceptions in the phantom hand. The multilayered electronic dermis (e-dermis) produces receptor-like spiking neural activity that allows to discriminate object curvature, including sharpness in a more natural sensation spanning a range of tactile stimuli for prosthetic hands. The authors were able not only to restore finger touch discrimination and objects recognition, but also to provide a pain sensation when the prosthesis touched sharp objects. In particular, they found that pain sensation is generated by a stimulation of 15–20 Hz.

Tan *et al* (2014) suggested that simple electronic cuff placed around nerves in the upper arm can directly activate the neural pathways responsible for hand sensations. This neural interface enabled the restoration of different sensations at many locations on the neuroprosthetic hand. Different stimulation patterns could transform the typical 'tingling sensation' of electrical stimulation into multiple different natural sensations, enabling the amputees to perform fine motor tasks and improving the embodiment.

In George *et al* (2019) a biomimetic method was described to restore both force and haptic sensation. The sensory feedback was implemented to restore the force sensation and promote objects recognition: USEA electrodes were used to deliver stimulation proportional to the variation of contact force exchanged between the prosthesis and the object during manipulation. Instead, the haptic sensation was based on the distribution of stimulation delivered during contact with the object with a fixed frequency and amplitude. The characteristic of this encoding scheme is based on electrical biphasic, charge—balanced of 200 or 320 μs phase durations. The biomimetic model describes the instantaneous firing rate of the afferent population using the contact stimulus position, velocity, and acceleration simulating all tactile fibers to any spatiotemporal deformation of the skin and hand. This strategy allows the amputee to augment the active exploration experience and to discriminate object size and stiffness.

Liu *et al* (2021b) have shown that primary afferents encode different stimulus features in distinct yet overlapping ways: scanning speed and contact force are encoded primarily in firing rates, whereas texture is encoded in the spatial distribution of the activated fibers, and in precisely timed spiking sequences. When multiple aspects of tactile stimuli vary at the same time, these different neural codes allow for information to be multiplexed in the responses of single neuron and populations of neurons. Exploiting this sensory architecture with invasive methods may lead to the development of prosthetic devices able to truly evoke natural sensations.

Another promising approach is targeted sensory reinnervation (TSR), i.e. the sensory version of TMR, which consists in coupling a pressure sensor placed on the prosthetic device to surgically redirected cutaneous sensory nerves (Marasco *et al* 2011). This technique strongly helps discrimination of objects size and stiffness during active exploration, especially if the tactile feedback is biomimetic (George *et al* 2019). Recently, Marasco *et al* (2021) have developed a prosthetic system based on both targeted sensory and motor reinnervation. TSR was used to deliver both touch and kinesthetic feedback. The authors showed that the system was able to significantly improve device control and promote embodiment.

These results indicate that, in order to close the loop on user and provide useful sensation (regardless the specific feedback modality), an optimal feedback control policy is necessary (Sensinger and Dosen 2020), as discussed in section 4.4.

4. Prosthetic control strategies and algorithms

Although the focus of this section is on the active prosthesis, it is worth mentioning that an important portion of the amputees still uses body-powered prosthesis (Carey *et al* 2015). These are cable-operated devices usually equipped with split hook or hand as terminal part (Millstein *et al* 1986).

Ranging from standard control approaches (e.g. dual-site control (Scott and Parker 1988)) to simultaneous control of multiple degrees of freedom (e.g. pattern recognition (Hahne *et al* 2018)), the literature offers disparate solutions for ULP control depending on the type of input signal and the sensors density.

In general, prosthetic control is performed at different levels. The low level refers to motor actuation (figure 5(D)) and, more in general, to the control of the active degrees of freedom of the device; the medium level consists of the translation of movement intentions into joint references and gestures (figure 5(C)); the high-level control translates input signals collected from the user (figure 5(A)) into movement intentions (figure 5(C), yellow panel—layer 1). In the next sections, we describe these different levels of control and provide examples of the different strategies that can be used.

4.1. Low-level control: from control commands to motor actuation

The low-level control combines the well-known strategies implemented in the automation industry to operate autonomous machines, e.g. industrial robots. We will not detail the structure and mathematical formality of these control architectures. However, if the readers are curious, a more complete and detailed analysis of robot lower-level control is provided by the comprehensive work of Siciliano *et al* (2010).

In brief, at the base of these controls, there is always an active and controllable actuator, that for upper limb prosthetic solutions coincides—most of the times—with an electrical motor (either brushed or brushless) often coupled to a dedicated transmission system (e.g. a planetary gear) to reach the desired torque-speed characteristic. It is possible to present the low-level control of ULPs as the combination of three possible nested controllers: the current, the speed and the position control loops (figure 5(D)).

The current control loop takes care of reliably tracking desired current trajectories. To be implemented, it requires the presence of reliable and precise current measurement sensors. The current control also provides a relatively good force/torque control of the system, being the current absorbed by the actuator directly proportional to the generated output torque. On top of the current controller, it is usually found a speed control loop to regulate the rotational speed of the motor and, thus, the speed of the actuated system. The combination of an external speed controller with an internal current control guarantees the possibility of safely operate the actuating unit in terms of desired speeds and torques. Sometimes, on top or in substitution to the speed controller, systems also implement a position control loop. The position controller guarantees the tracking of desired angular trajectories. It is therefore preferable to use the speed controller if the goal is to precisely track given trajectories in specific time intervals. The implementation and application of speed and position controllers can be performed either before (fast shaft) or after (slow shaft) of the transmission system. The decision depends on the availability of sensing devices (e.g. angular sensors such as encoders or resolvers) to measure the required physical quantities.

Figure 5. Architecture of ULP control: actuation and feedback. Input signals collected from the user (A) are processed into the embedded processing unit (C) to generate control commands for the single joint control unit (D). Feedback information coming from the prosthesis or its interaction with the environment (E) are also processed in the embedded processing unit (C) to deliver sensory feedback (B). The embedded processing unit (C) can be set up by different layers: layer 1 (intention detection, yellow panel) is the software turning the input signals (A) sampled by master board into detected movement intentions, by means of specific control algorithms (e.g. machine learning or deep learning (DL) algorithms); layer 2 (human-robot interaction, green panel) is the software responsible of processing prosthesis position (joint and Cartesian space control) and external information (tactile identification, (E)); layer 3 (encoding haptic feedback, blue panel) is the software responsible for encoding the information processed in layer 2 into sensory feedback. The output of the embedded processing unit are control commands (mediated by actuator drivers) both to move the device and to provide sensory feedback. This has a direct impact on the user experience (F) in terms of learning how to use the device (training) and of user-prosthesis integration (embodiment).

All these controllers are implemented in a negative feedback architecture and typically controlled by means of PID controllers, whose proportional (P), integrative (I) and derivative (D) parameters are tuned to reach the desired system response in terms of control reactivity (rise time and settling time), precision (steady-state error and overshoot) and stability. It is worth mentioning that a negative feedback architecture is typically only bounded to the low-level control of the prosthesis, while higher level controllers and especially high-level control (see section 4.3) are often treated in an open-loop fashion, where the user directly generates the reference control signal without any feedback verification. The generated reference commands will then be directly sent to the low-level controller.

4.2. Mid-level control: from movement intention to control commands

The mid-level techniques (figure 5(C), yellow panel—layer 1) aim to synthetize the control commands to suitably activate the electric motors of the multiple DoFs ULP (actuation drivers in figure 5(C)). These signals are the input of the aforementioned low-level control.

A major classification of the mid-level control strategies for multi-DoFs robots divides them in two categories: joint-space and task-space (Cartesian) controllers (Siciliano *et al* 2008, Corke and Khatib 2011).

Joint-space control strategies directly feed the commands to each of the actuated joints, namely DoFs, of the upper-limb robotic device. It is a direct approach that does not require any particular mathematical manipulation. In such a scenario, the mid-level control receives information from the high-level (see section 4.3), then it assigns specific commands to each low-level controller (see section 4.1). The logic used to assign the control commands is strongly based on the kind of information coming from the high-level side. Nonetheless, it will most likely reduce to a set of independent commands for each of the actuated joints.

On the other side of the spectrum, we have task-space based control strategies. In this case, the control commands for each of the joints are the results of a mathematical manipulation that involves the transformation from the Cartesian space to the joint space. If the aim is to regulate the Cartesian trajectory, the controller will need to translate the Cartesian trajectories into joint angles, by means of a process known as inverse kinematics. If instead the aim is to regulate the Cartesian force, the controller will transform the Cartesian forces into joint forces (or torques) utilizing the process of inverse dynamics.

Both these approaches are well known to robotic applications and will not be treated in detail in this review. Nonetheless, the authors suggest the comprehensive works of Corke and Khatib (2011) and Siciliano *et al* (2008) to get the fundamentals of the aforementioned approaches.

In general, Cartesian based controls are more intuitive for the external user, namely any subject interacting with the robot as an external tool. In fact, the robot behavior can be more naturally interpreted being the forces or the trajectories referred to the three-dimensional space we are used to deal with. However, from a computational and complexity point of view, task-space controllers require a bigger effort and introduce limitations to their application, e.g. singularities, redundancies. On the other hand, joint space control behavior is less intuitive to predict but it is easier and less complex to implement.

Which approach is better for upper-limb prosthetic devices is still unclear. However, it is important to notice that, even if Cartesian controls are more intuitive from an external perspective, they might appear more complex from an internal perspective, such as the one of a prosthesis user, where the motion of the arm is more likely imagined in terms of joint motions and not Cartesian ones.

4.3. High-level control: from input signals to movement intentions

This section summarizes the most assessed techniques for ULP control. Considering the prostheses available on the market but also the research activities, the main input source exploited to control such devices is the EMG. On the basis of the EMG type multiple control strategies can be employed, and the last decades of studies on active prostheses mainly focused on the control strategy design and development.

The most common control strategy is based on dual site control which consist in two electrodes placed in two antagonist muscles (Scott and Parker 1988). This solution allows the control of the motor in two directions according to the muscle amplitude of the selected electrodes. The synthetized reference usually is proportional to the amplitude of the muscle signal in term of speed or force. With the introduction of multiple DoFs, a co-contraction strategy has been implemented to switch between controlled joints (Resnik *et al* 2018). This allows the control of a single DoF at a time using two electrodes as in dual-site control. When both muscles are simultaneously contracted the control signal switches the joint to be controlled. This is a simple solution yet unnatural and lacking intuitiveness.

Another diffused strategy to control prosthesis with multiple active DoF is the finite state machine (Moon *et al* 2005). Commercially available ULPs implement this strategy to switch the position of the thumb to reproduce different types of grasp (Ottobock 2020a, 2020b, Ossur 2020b). For example, the Michelangelo hand allows to switch the thumb position when a signal of opening is triggered with the hand in a fully opened configuration (Ottobock 2020b).

With the aim of increasing the number of controlled DoFs, many different methods were proposed, such as muscle synergies, feature extraction (FE), multi-amplitude threshold control and machine learning methods. Muscle synergies capture muscle activation invariance during motor production and can be exploited as control variables for ULP, with aim of obtaining a biomimetic human-like behavior (d'Avella and Bizzi 2005). The main idea is to extract motion primitives from muscle synergies and combine them to generate complex arm movements (Jiang *et al* 2013, Liu *et al* 2021a). Furui *et al* (2019) propose a biomimetic control based on muscle synergies to extract motion primitives and combine them to generate complex movements. FE methods foresee the computation of some EMG-based metrics that reflect movement intentions (Guo *et al* 2015). Multi-amplitude threshold methods work as dual-site control, but they associate different amplitudes of the input signal to different DoFs (Markovic *et al* 2019). Although robust, these techniques are poorly used because they lack intuitiveness (Markovic *et al* 2019). Machine learning methods will be described in the following paragraph.

#### 4.3.1. Machine learning algorithms

Figure 6 illustrates the main machine learning methods employed for ULP control. These methods generally solve a pattern recognition problem in which, given the input signal, an output movement have to be identified.

The first Pattern Recognition (PR)-based control schemes arose around the second half of 1960s (Scheme and Englehart 2011). In this configuration, the acquired EMG signals are elaborated by the controller to determine the action to be performed by the prosthesis. The five pillars of this computation process are: pre-processing, data segmentation, FE, classification, and post-processing. Each step is briefly described in table 5.

EMG-based pattern recognition controllers are now investigated by many groups and are even available in commercial prostheses (Coapt 2017, Ottobock 2019, i-biomed 2021).

The PR-based controllers apply linear and non-linear methods to classify the EMG signal into a possible large number of movements. The two main families of classification methods used in this context are regression (Hahne *et al* 2014) and classification techniques (Hudgins *et al* 1993). While the former is usually

Figure 6. Division of machine learning approaches for ULP control.

Table 5. Pattern recognition steps.

Pre-processing	During this phase, the incoming signals are firstly filtered to delete the interferences, such as acquisition noise and artifacts.
Data segmentation	This process divides the signals into time-windows, overlapping or adjacent (Parajuli *et al* 2019).
Features extraction	It reduces the signal information into a set of representative features in time domain (e.g. variance, zero crossing, etc.), frequency domain (e.g. mean frequency, spectral properties, etc) or time-frequency domain (e.g. the wavelength transform, an alternative to the traditional Fourier Transform useful for noise-removal and data compression (Hartwell *et al* 2018)), as described in Boostani and Moradi (2003). Importantly, this part can greatly affect the computational costs.
Classification	This is the crucial step for the classifier, where the controllers recognize and classify the signals input information and generate an output for the actuators.
Post-processing	It has the main goal to reduce as much as possible the misclassification. An example is the majority vote strategy, in which the current output is calculated on the previously most recognized class. The majority vote scheme is used for eliminating spurious misclassifications caused by too short windows on which the most recurrent class is selected; it employs the previous classification results and evaluates the current output on the basis of the previously most recognized class (Englehart *et al* 2003).

simple to implement and train, the latter are generally more difficult to employ. The embedding of neural networks (NN) in an ULP strictly depends on the structure of the algorithm (number of layers and neurons), since complex architecture requires high computational effort (Hagan *et al* 1997).

Statistical regression models usually produce good results in terms of high accuracy percentages. However, the out-of-laboratory results are particularly poor, because these techniques are extremely sensitive to changes of the input signals (Parajuli *et al* 2019). Motivated by this issue, in the last decade, many groups focused on classification-based techniques to implement more reliable decoders. Importantly, training classifiers requires longer than training linear models, however, the formers can achieve better results during real-time execution. Different classifiers have been exploited in ULP control such as support vector machine, regularized least squares, whereas the gold-standard is the linear discriminant analysis (Scheme and Englehart 2011, Cloutier and Yang 2013b, Di Domenico *et al* 2021). Among NN, the most common architecture is the multi-layer perceptron (MLP) (Amrani *et al* 2017, Shahzaib and Shakil 2018). The MLP is a supervised machine learning (ML) technique, which exploits labeled data to train the algorithm. It is characterized by three types of layers: input, hidden and output layer. The first one contains the same number of neurons as the input signals (for example, features extracted from EMG signals), the second stage can have one or more layers where there are all the trainable neurons, while the last layer comprises all the output nodes representing the results (for example, classification likelihood of each class of movement). Neurons of a certain layer are fully connected to the neurons in the next layer via nonlinear activation functions. However, as for the regression algorithms, the performance results obtained in the lab are not easily replicated in the real-life scenario. Moreover, the complexity of the controlled prosthesis (e.g. the number of DoFs) corresponds to a higher number of neurons in the NN, with important consequence not only on the computational burden, but also on the memory consumption.

When considering an increase in the number of controllable DoFs, current pattern recognition approaches demonstrated poor performance (Piazza *et al* 2020). As a matter of fact, to enhance the

classification rate (i.e. number of correctly recognized movement) a greater content of information should be handled. The higher the amount of input data, the more complex would the ML algorithm be.

Therefore, HD-sEMG can be exploited to increase the amount of muscular information but this comes at the cost of higher computational burden. It has been proven that the use of this type of data can be helpful in increasing the robustness against electrode shift (Pan *et al* 2015), allowing an improvement of the classification by exploiting spatial images of the muscular contractions (Geng *et al* 2016), and for retrieving measures of motor unit potentials, which can be difficult to assess without invasive techniques (Merletti *et al* 2008).

Different techniques can be exploited to extract motor units' activity from the HD-sEMGs. The main used decomposition algorithm is the blind source separation (with the Convolution Kernel Compensation described by Holobar and Zazula (2007)) which seems to be the most suitable since it does not make any prior assumptions on the action potential shapes. The main problems related to this technique is the lack of a useful output for the prosthesis control, since the decomposition provides an extraction of principal features of the EMG signals. On the one hand, the algorithm returns reliable information about neural activity, but, on the other hand, it increases the computational burden required to the system. Indeed, the Holobar algorithm has been used together with ML algorithms to control robotic arm in real-time (Barsakcioglu and farina 2018).

Another approach includes the exploitation of ML algorithms where the input EMG signals are considered as numeric values and the definition of the output is based on a Black Box technique. Therefore, the mathematical tools contained in the Black Box do not take into account the biomechanics of the amputated limb and they are not specific for prosthetic applications.

It is relevant to feed the ML algorithm via a set of EMG signals (muscular patterns) specific for different prosthesis movements in such a way that the classifier does not misclassify. However, it is not always feasible to acquire the same signals for each movement due to different sources of errors (i.e. muscle fatigue, sweating, electrode misalignment). Indeed, more complex classifiers belonging to the DL field are exploited to make the control more robust. A possible application can be the use of convolutional neural network (CNN), which exploits dimensionality reduction to extract complex features from the activation maps of the HD-sEMG without dramatically increasing the computation time (Olsson *et al* 2019). This type of algorithm is also ideal for increasing the number of DoFs (and therefore the number of classes to be recognized) while keeping a quite high accuracy rate (Hartwell *et al* 2018). Moreover, Zhai *et al* (2017) has proved that the exploitation of CNN can help in removing issues of daily life noise, updating its feature map to include this new information, avoiding the need of periodical readjustment.

Adaptive technique based on reinforcement learning (Vasan and Pilarski 2017, Wu *et al* 2022) has been recently investigated, with the aim of facilitating the learning process of prosthetic use. This approach is promising as it points towards the development of a 'human–prosthesis symbiosis in which human motor control and intelligent prosthesis control function as one system', as defined by the group of Huang *et al* (2021).

Other DL algorithms take into account time series with feedback loops with prior hidden layers (Sun *et al* 2021a). This architecture allows storing the history of the input signals by considering the information of previous time instants, also resulting in performance improvements with respect to simpler DL architectures (Amado Laezza 2018).

Recently, novel DL strategies have also been proposed for ULP: recurrent NNs process temporal or sequential information; temporal convolutional networks take advantage of a one-dimensional convolution layer running along the time dimension to learn the time dependence of a given input signal (Li *et al* 2021); transformers are attention-based architectures applied to HD-sEMG data (Burrello *et al* 2022, Montazerin *et al* 2022).

Overall, the main problem related to ML applied to the bionic field is the evident gap between the results observed in a closed safe environment, such as a laboratory, and in real daily life (Resnik 2011).

4.3.2. Model-based approaches

To overcome the limitations of ML algorithms for ULP control, some groups investigated the model-based approach, which consists of an accurate description of the muscles and bones involved in the movements starting from the Hill model of muscle fiber (Winters 1990). For example, the neuromusculoskeletal model extracts from the residual EMGs the activation dynamics of the limb (Pan *et al* 2018, Zhao *et al* 2022). The activation dynamics combined with the kinematics of the limb produces the contraction dynamics. This consists of the modification of fiber length involved in the motion along the specific DoFs. In particular, Sartori *et al* (2018) implemented a control strategy based on the physiology and kinematics of a real hand and tested it with an amputated subject performing some complex grasping tasks. This approach needs a calibration step to scale the model to the subject specific activation EMGs. Results showed great stability over

the noise introduced by sensors or movements artifacts. Moreover, the amputee was able to reproduce simultaneous multi-DoF gestures. The limitation of this approach is its susceptibility to electrode shift and fatigue condition that affects the EMG acquisition. The real-life scenario is yet to be tested, but preliminary results appear very promising (Sartori *et al* 2018).

4.3.3. Sensor/data-fusion and other techniques

For ULP control using different input sources together with or without EMG signals, other methods can be adopted. In case of force myography, the same algorithms used for EMG input can be applied. For example, machine learning techniques can be used to analyze and synthetize output starting from FMG input (Cho *et al* 2016). The adoption of other input signals different from EMG clearly requires the implementation of ad-hoc methods for their processing. For example, voice control introduces audio analysis method to detect and translate command into prosthetic movements (Mainardi and Davalli 2007, Alkhafaf *et al* 2020). Further, tongue control allows the motion of the prosthesis using a wireless controller resembling a dental retainer and providing the functionality of a wireless joystick or keyboard (Johansen *et al* 2016).

The high complexity of ULP control has led to the development of sensor fusion approaches, in which input signals of different nature are simultaneously collected and then processed to estimate the intended movement more reliably and accurately.

On low-density sEMG, we can find robust and semi-autonomous control solutions based on custom multi-amplitude algorithms, as those implemented on the Michelangelo hand with closed-loop multi amplitude control (CMAC) (Markovic *et al* 2019, Mouchoux *et al* 2021). The adoption of IMU sensors may lead to further improvements, such as the automatic adaptation to unexpected external factors, including sweat, muscle fatigue, mental stress, electrode re-positioning and weather conditions. The state-of-the-art algorithms have to cope with these challenging issues. Therefore, the combination of EMG and IMU as input to a classifier could provide useful localization information of the hand position, which could delete possible false positives, actively improving the obtained accuracy (Krasoulis *et al* 2017, 2019b). Moreover, it has been observed that integrating EMG, IMU and artificial vision sensors could benefit both the classifier accuracy and the increment of available DoFs (Mouchoux *et al* 2021). Other promising research advancements demonstrated that mixing EMG with FMG could lead to an improved multi-DoFs control as proposed by (Nowak *et al* 2020). Similarly, Jiang *et al* (2020) proposed a sensor fusion approach among EMG and FMG. Moreover, by fusing FMG and IMU, other interesting results were presented by (Ferigo *et al* 2017). In addition, other research activities treated NIRS fused with EMG (Guo *et al* 2017b) and IMU respectively (Zhao *et al* 2019).

In conclusion, a data fusion aims at compensating some of the main limiting factors of single input approaches (such as EMG-based or others) as these latter suffer from artifacts, electrodes shift, etc.

4.4. Control strategies for the sensory feedback and closed-loop approaches

Recent developments in the prosthetic field have focused attention on sensory feedback restoration. In particular, many groups began studying how to provide the user with information about the interaction between the prosthetic system and the physical world. This information needs to be collected (figure 5(E)), processed (figure 5(C), green panel—layer 2) and encoded into control signals (figure 5(C), blue panel—layer 3) for the feedback system (figure 5(B), e.g. vibromotors, electrostimulation, etc).

The control strategy implemented to encode this information depends on the type of sensation to restore as, for instance, tactile feedback (pressure, temperature, pain) or proprioception feedback (gestures, joint movements). To this aim, different solutions have been developed.

Mamidanna *et al* (2021) focused their research activity on the force feedback that the prosthesis applies to the grasped objects by using vibromotors attached to the forearm skin. To do that, an encoding scheme of the current absorbed by the prosthetic motor was translated into vibromotors amplitude. Other sensorized solutions have been developed to directly translate the prosthesis interaction to user sensation like artificial skin able to translate the distribution of pressure and intensity to tactile and pain sensations on users with invasive interfaces (Jiang *et al* 2019). Similarly, Markovic *et al* (2019) implemented a proprioceptive feedback translating prosthesis movements into vibration orientation and shape to be intuitively interpreted by users.

In addition to prosthetic feedback, some groups are working on user feedback in terms of providing information about how the prosthesis is controlled by means of closed-loop approaches. For example, Schweisfurth *et al* (2016) have tested on amputees a ULP system in which EMG input used to drive the prosthesis was translated into intensity of vibromotors activation. In this configuration, the amount of EMG activity detected is directly proportional to prosthesis grasping strength and to intensity of vibration amplitude. In another work, the control commands generated by the user and translated into joint angles were encoded as proprioceptive information delivered through electrical stimulation (Garenfeld *et al* 2020). This allowed user to understand if the intended control command was correctly detected by the algorithms.

Similarly, Tecnalia developed a ULP system with sensory feedback by merging into a unique device EMG acquisition and electrical stimulation (Štrbac *et al* 2016). Although this solution significantly reduced the problem of encumbrance, it still faces some issues mainly related to the artifacts that the stimulation produces on the EMG signal and that cannot be removed using standard signal processing algorithms (Li *et al* 2019).

As for decoding of movement intention from input signals, the interpretation of feedback information needs a calibration procedure aimed at familiarizing the user with the ULP device. In this context, it is fundamental to guide the user to: (a) produce the correct input signal to perform the desired movement, and (b) to intuitively convert the feedback signal into useful information for motor planning. A user-centered approach can maximize and speed-up the learning process, as detailed in next section.

5. User-centered solutions in upper limb prosthetics

As observed in previous sections, multiple efforts in research and development offer heterogeneous technological solutions to enable a proficient control of an ULP device. However, selecting a sub-set of these solutions is compulsory for implementing and validating them. Accordingly, this section will discuss user-centered solutions based on the technologies described in the previous paragraphs, highlighting the opportunity of overcoming a separation (often, an opposition) between (for instance) user-centered features and technical ones, or between ADL-related performance and biomimetic one. Nevertheless, we decided to proceed in the selection of each solution by pragmatically moving from one perspective toward the other.

Fundamentally important criteria for performing such a selection should come from the analysis of the prosthetic user experience (figure 5(F)). Indeed, special attention should be paid to the user needs in order to promote the daily use of prosthetic devices, a prerequisite for checking the validity of any technological solution presented in previous sections.

In particular, the prosthetic technology acceptance constitutes a dramatic issue in this domain. Overall, ULP technology acceptance (Longfellow 2014) is tied to several interdependent functional factors related to ease of use (sensory feedback, control), dexterity (motion complexity, force output, actuation speed, manipulation), body integration (anthropomorphism, autonomy, weight), technology transfer (cost, reliability). Further factors embrace several domains, namely clinical (age, level of amputation, fitting timespan), cultural (education, social conditions, living environment, country development), and personal (psychological attitudes, subjective expectations, occupation, activity, environment).

Low acceptance can contribute to the abandonment of a prosthetic hand, erasing any chance of improvement in the control skills of the users (Castellini 2020). Thus, it is important to promote an intrinsically motivated and continuous ULP practice, which must be experienced by the users as immediate and rewarding in order to achieve high degrees of technology acceptance (rodgers *et al* 2019) and integration (Shaw *et al* 2018). The users must also feel engaged enough to surpass the impact of feeling social stigma or the doubts on the functional impact of the system on daily life. Intuitive patterns of system control play a critical role in this context to facilitate a spontaneous use of the system and to improve the user experience (Krasoulis *et al* 2019a).

Obviously, the absence of appropriate acceptance, usability, and user engagement creates a barrier for the introduction (and the further development) of any technological improvement in prosthetics.

5.1. Toward user-centered upper limb prosthetics
In order to improve the ULP acceptance, different approaches can be adopted, especially in terms of user research (Figliolia *et al* 2019). A review of Cordella *et al* (2016) provided a rich set of guidelines for enhancing the prosthetic hand technology acceptance through the analysis of the user requirements, considering literature and case studies like Luchetti *et al* (2015). Among these requirements: the capability to accomplish basic grasping actions during activities of daily living with minimal visuo-attentional focus, high dexterity, appropriate strength control; biomimetic features of sensory feedback and anthropomorphism; duration and reliability of the device and its component; technical features with impact on comfort like heat dissipation and motor noise reduction.

All features must be designed according to individual preferences. These preferences can depend on demographic factors, type and level of amputation, pain symptoms, and type of prosthesis (e.g. body-powered or myoelectric) (Biddiss and Chau 2007, Biddiss *et al* 2007, Davis and Onge 2017, Uellendahl 2017, Smail *et al* 2020, Kerver *et al* 2020). The amputees' preferences must also be investigated to design virtual and augmented environments for prosthetic use training (Garske *et al* 2021b). If appropriately devised, game-like engaging exercises can motivate the user to train, feeding the prosthetic with consistent biosignals that efficiently represent different types of grasps, an advantageous condition for ML based control (Tabor *et al* 2017). This can possibly happen with a successful generalization if the training is adequately designed with solutions like task switching (Heerschop *et al* 2021). Overall, the training designers

should focus not only on playfully engaging the user to train the muscles, but also on accurately representing prosthetic use tasks to enable the related skills transfer (Garske *et al* 2021a). Furthermore, the parameters of meaningful and, possibly, ecological interactive settings can be experimentally controlled by the clinician or the researcher (Resnik *et al* 2011, Bouwsema *et al* 2014, Markovic *et al* 2017, Paljic 2017, Nissler *et al* 2019, Boschmann *et al* 2021, Phelan *et al* 2021). In addition, interactive settings can be adjusted to the individual needs and reactions. To understand the individual needs in prosthetic use (training and daily activities), the improvement of user research methodologies themselves becomes a priority to promote effective co-creation frameworks. Recent works (Jones *et al* 2021) described surveys and workshops to investigate the point of view of amputees and all stakeholders (clinicians, academics, experts and managers in industry and charity), observing a gap between laboratories and the real life of prosthetic users (whose issues are typically misrepresented by media too). Interestingly, initiatives like the Cybathlon competitions for assistive and prosthetic technology users are also devised for overcoming such a gap (Riener 2016).

The users' involvement in iterative activities of design and evaluation of products and of product services is highly important (O'Sullivan *et al* 2017). Such activities must be planned for checking and improving the usability of prostheses as medical devices according to the international standards (Pelayo *et al* 2021) and for estimating the impact of user experience on the technology acceptance (Longo 2018, Lah *et al* 2020). Obviously, user-centered evaluation methodologies and metrics must be adjusted to the specific case of ULPs (Resnik 2011, Zahabi *et al* 2019), especially considering how their user interface is not based just on buttons, plugs, and leds and their behavior and feedback are eminently biomimetic in hand-like manipulation tasks.

The functional resemblance of the ULP design to a real hand is a wise strategy for promoting a positive interaction between user and prosthesis. Such an approach (implicitly and explicitly) aims at building artificial limbs that are spontaneously used by the amputees as their own. Such a 'prosthetic ownership' experience is deeply investigated within the domain of the embodiment research, crossing disciplines like cognitive psychology and robotics according to the roadmap in Beckerle *et al* (2018).

The embodiment phenomenon can be constituted across its components, i.e. self-location, ownership, and agency, by the sensation that an artifact is integrated in one's body scheme (Kilteni *et al* 2012, Maimon Mor and Makin 2020, Toet *et al* 2020). Overall, the technology embodiment promotes intuitive control with improved user experience and acceptance (Makin *et al* 2017, Nelson *et al* 2020, Toet *et al* 2020). About ULPs, the embodiment improves: (a) movement control (Grechuta *et al* 2017), (b) object discrimination and manipulation (Tan *et al* 2014), (c) manual accuracy and sensitivity. Furthermore these processes contribute to: (d) the reduction of the phantom-limb pain (Page *et al* 2018) and (e) the mitigation of the risk of prosthesis abandonment (Mcdonnell *et al* 1989, Beckerle *et al* 2019).

Obviously, we must ponder how to measure and to stimulate the prosthetic embodiment. Overall, the embodiment evaluation is typically entrusted to methods (questionnaires, biosignal analysis, proprioceptive drift) based on the rubber hand illusion (RHI) studies (Botvinick and Cohen 1998, Tsakiris and Haggard 2005, Ehrsson *et al* 2008, Romano *et al* 2021). RHI can be also implemented on its different versions—e.g. virtual hand illusion (Pyasik *et al* 2020, Beckerle 2021) and robotic hand illusion (Romano *et al* 2015, Huynh *et al* 2019). However, these methodologies are still debated in cognitive studies (Gallagher *et al* 2021) which show the complexity of the processes underlying the embodiment itself.

Understanding such processes is required for designing appropriate strategies to enhance the embodiment of an artificial limb. First of all, it must be said that daily prosthetic practice, individual characteristics (like the cause of limb absence), and multisensory feedback congruency play fundamental roles in this process, which does not necessarily require cosmetic improvements or specific control patterns (body-powered or myoelectric) (Dornfeld *et al* 2016, Engdahl *et al* 2020b, Moore *et al* 2021, Zbinden *et al* 2021). Embodiment training strategies can also be explored in virtual and augmented settings (Barresi *et al* 2021), even if the generalization of their effects to actual prostheses must be investigated. Importantly, establishing optimal techniques to promote the embodiment of an artificial limb is a way to fully engage the user in exploring the prosthetic device and its potential, further improving its embodiment too within a virtuous circle.

However, it is necessary to consider what technological challenges must be faced for achieving a truly 'biomimetic' experience as the prosthetic embodiment to improve the use and the acceptance of an artificial limb.

5.2. Promoting prosthetic use and acceptance through improved mechatronics and control

To improve ULPs, two main classes of approaches can be taken: one focusing on mechatronic development and the other on control implementation. From the mechatronic side we suggest optimized actuation, anthropomorphism, human-like grasping behavior, and biomimetic performance as key factors to take into account for promoting ULP use and acceptance, while from the control perspective we identify robust control strategies and use of smart prostheses. Moreover, we believe the inclusion of multimodal sensory feedback a

fundamental prerequisite of next generation prostheses. Each approach contributes to approximate the prosthetic user experience and the user-prosthesis system performance to, respectively, the sensations provided by a natural limb and its spontaneous and effective usage. However, all approaches face issues that must be solved in order to obtain a robustly controlled prosthetic system easily accepted by the user. These approaches thus constitute research challenges, summarized in table 6 with their actual potential solutions.

These technological approaches constitute the premise for many kinds of breakthrough in prosthetics. Obviously, we need to consider that tradeoff calculations must be made for selecting the most rational set of features that can be combined for providing a satisfying (without creating excessive expectations in users and any stakeholder) and (also economically) sustainable design of the devices. However, the features we described, and their user-centered synergies, can make us foresee the perspectives on bionic hands innovation that will be discussed in next section.

6. Perspectives on tomorrow's upper limb prosthetics

By analyzing the state-of-the-art techniques for ULP input (section 3.1) and feedback (section 3.2) signals, it emerges that there exist two parallel directions for future development, namely the non-invasive and invasive approaches. This is due to different reasons: first of all, because non-invasive solutions may provide the amputees with a plug and play device ready to be used for ADLs, while invasive solutions still need to overcome technological barriers before they can be routinely adopted by the majority of amputee population. Moreover, the specific choice of non-invasive vs invasive strategy is highly dependent on the level of amputation. For example, for trans-humeral amputees, TMR represents the most promising opportunity for restoration of lost functionality, which could not be achieved with non-invasive approaches. In the following, we describe possible direction for future prosthetics. In particular, for non-invasive solution we suggest the use of multiple input sources and of sensory feedback. For invasive solution, we recommend strategies promoting a direct translation of user intentions into prosthetic movements. These directions are also outlined in table 7.

6.1. Short-term non-invasive solutions
The most widespread technique for non-invasive control is based on sEMG. This technique has lots of advantages, such as low cost, direct correlation between muscle activation and movements, and intuitive control, as described in sections 3.1.1.1 sEMG and section 4.3. Nevertheless, EMG-based systems lack of robustness, due to EMG susceptibility to artifacts of biological nature (sweating, hair, muscle fatigue), instrumental source (electromagnetic disturbances), intrinsic to the device (movement artifacts, electrode shift, variations in contraction depending on the orientation of the arm), or intrinsic to the control algorithm (optimization of the classifier).

For this reason, there is ample room for improvement in non-invasive approaches for ULP control. One research direction points toward the use of sensors fusion techniques, in which multiple input data is taken into account to estimate the movement intentions, as described in section 4.3.3. However, the use of great number of sensors may be impractical for creating an embedded system for everyday applications. Future ULP should be equipped with multi-modal control based on minimal number of sensors (Jiang *et al* 2012, Di Domenico *et al* 2021). Moreover, recent studies are investigating the miniature technology to limit the encumbrance within the socket (Marinelli *et al* 2021).

With the aim of improving ULP control, we pose that a fundamental element is the feedback. This will allow a closed-loop interaction between user and prosthesis, a essential prerequisite for promoting the integration of the prosthesis into the body scheme and for facilitating the controllability of the entire system. In this direction, many studies have pointed out the usefulness of vibrotactile stimulation for providing sensory feedback information (Sensinger and Dosen 2020). This technique is cheap, easily integrated into a socket, and its modulation in frequency and intensity allows to provide various information.

In the last years, wearable technology has largely expanded in many fields, influencing also prosthetics. For example, the CTRL Labs have realized a wearable wristband, which reads EMG signals and translates them into finger movements (Melcer *et al* 2018), later improved by the Facebook Reality Labs by means of advanced ML algorithms (Basu 2021). This technology could be exploited for prosthetic applications with a strong impact for ADLs usage.

Another promising approach consists in providing sensory information by stretching the skin of the stump. This can be done with a wearable haptic device producing rotational skin stretch according to the movement of the controlled device (Kayhan *et al* 2018, Battaglia *et al* 2019).

These examples show that innovative ULP solutions can be adopted for restoring lost functionality in the short-term using non-invasive approaches.

Table 6. Current issues affecting ULP mechatronics and control.

	Challenge description	Current and possible solutions
MECHATRONICS	Optimized actuation The actuation architecture (i.e. the number of actuators employed) influences the overall performance of the device in terms of: • amount of possible gestures/configurations (therefore controllability and dexterity); • contribution to ULP weight; • grasping strength; • acoustic noise during actuation (i.e. reduction stage).	There are two possible strategies to optimize the actuation of ULPs. From a qualitative point of view, a fully actuated system might lead to independent control of each single joint thus allowing to replicate the full amount of gestures of a real hand. However, this solution is typically characterized by a heavier and bulkier or poorly performing system. Moreover, this configuration prevents the use of power motors to generate human-like grasping strength. On the other hand, an underactuated device might guarantee compactness, light weightiness and the possibility to achieve more efficient actuation and therefore higher performance, at the cost of passive uncontrolled degrees of freedom. Both actuation solutions are still affected by acoustic noise during prosthetic movements and this constitutes room for improvement for future development.
	Anthropomorphism Anthropomorphism also represents a key design feature for an ULP. In fact, users are more prone to adopt and utilize anthropomorphic devices that anatomically and functionally resemble their missing limb as much as possible (Varol *et al* 2014).	In two useful reviews on mechanical and anthropomorphic aspects of prosthetic hands, Belter and colleagues proposed a list of guidelines to achieve, by mechanical and mechatronic means, the desired hand anthropomorphism in terms of size, weight, shape and kinematic capabilities (Belter and Dollar 2011, Belter *et al* 2013). The group of Metta proposed a systematic approach to benchmark different robotic and prosthetic hands in terms of shape, feature and performance, observing a continuous need for weight, payload and generic grasps improvement while maintaining an anthropomorphic appearance (Vazhapilli Sureshbabu *et al* 2019). However, it should be mentioned that a minority of ULP users do not recognize anthropomorphism as priority, focusing their needs on functionality. In some cases, the ULP is deliberately unconventional and worn as a fashion gadget or stylized wearable art pieces (De Oliveira Barata 2021).
	Human-like grasping behavior Human-like grasping behavior represents the aesthetic capability of the ULP to synergistically operate and adapt its configuration and to robustly perform different sets of grasping tasks.	Underactuated solutions greatly simplify the accomplishment of this due to their intrinsic capability of conforming to the object to be manipulated during grasp (Catalano *et al* 2014, Weiner *et al* 2018, Laffranchi *et al* 2020). On the other hand, in systems with such architecture, human-like behavior is limited exclusively to the grasping function.
	Biomimetic performance The biomimetic performance is intended as the capacity of reaching the desired biomechanical force and speed requirements in the different activities of daily living.	Achieving the force and speed of the biological hand in an anthropomorphic prosthesis is an extremely challenging task. High integration density can be facilitated by adopting mechatronic architectures with high efficiency. This can be achieved by using once again underactuation, which centralizes the power source (motor) and minimizes the number of transmission components which dissipate mechanical power (Laffranchi *et al* 2020).

(Continued.)

Table 6. (Continued.)

	Challenge description	Current and possible solutions
CONTROL	Robust control strategies	
	The control strategy is a fundamental element to provide the ability of simultaneous actuation of multiple joints and to improve the activities of daily living (ADLs) of the amputees. It is strongly linked to the onboard mechatronics, and in particular the number of active joints plays an important role. In the case of fully actuated systems, the relationship between inputs and outputs is complex and unintuitive, due to the limited amount of input information, for example carried from the superficial EMG signal (i.e. extrinsic muscles, residual muscular activity) (Farina *et al* 2004). Typically, control strategies on commercial devices are based on buttons (Ottobock 2020a) or smartphone apps (Ossur 2020a) to respectively configure the hand grasps and gestures. Considering an underactuated system, the relation between inputs and outputs for trans-radial case is typically based on the real residual muscle of the forearm. In this case, the research is focused on finding a connection between EMG and movements (Marinelli *et al* 2020, Nguyen *et al* 2021). Considering a trans-humeral or interscapular/shoulder disarticulation, the loss of muscles related to the actuation of hand and wrists increases the complexity of relation between inputs and outputs, resulting in a less intuitive control. Smart prostheses	Innovative solutions based on ML algorithms are routinely used for trans-radial amputation and have been proposed not only for underactuated but also for fully actuated ULPs to translate the user intentions into single finger movements (Nguyen *et al* 2021). These methods are intuitive and functional, but they are still not widespread in the market due to their limited robustness over time, requiring frequent recalibration of the entire system (Marinelli *et al* 2020). Possible solutions in this regard are constituted by maximizing the user experience during training (Del Vecchio *et al* 2021), or by incremental learning strategies for device control that allows continual adaptation to the changes in the input signals (Gijsberts *et al* 2014). For trans-humeral amputation, the target muscle reinnervation (TMR) is the most promising solution for simultaneous control of a multi-DoF prosthetic system (Mereu *et al* 2021). Also in this case, ML algorithms represent an interesting approach to relate the muscular activity of the reinnervated limb into more intuitive and physiological movements.
	Current prostheses lack of the possibility to automatically process an incoming stream of information, differently from the human hand that is equipped whit automatic behavior in response to certain stimuli. For example, our hand immediately reacts when touching a burning heat source, even before we consciously perceive the thermal sensation. This kind of features can be a precious improvement of the current solutions in shared-control (Cipriani *et al* 2008, Yang and Liu 2021).	It would be desirable to have a prosthesis able to take decisions in those situations that require an immediate response. This would free the user from the need of constantly monitoring prosthesis status, limiting the damaging events and allowing the user to operate the device only for voluntary motor production, consequently reducing the mental effort. Equipping the prosthesis with specific sensors and related processing can enable shared control solutions aimed at completing the action-perception coupling with the missing contribute of sensory information. We define this solution as feedback-to-prosthesis. This can be obtained by sensors embedded on the prosthesis, which can measure the interactions between the device and the external world. This can be done exploiting different measurements ranging from the motor current (Ajoudani *et al* 2013, Laffranchi *et al* 2020, Deng *et al* 2020), to tactile/pressure sensors (Tomo *et al* 2018), Inertial measurement units (IMU) to understand the actual pose of the prosthesis (Krasoulis *et al* 2019b), or artificial vision (Mouchoux *et al* 2021) to understand which is the shape and orientation of the nearby objects.

(Continued.)

Table 6. (Continued.)

	Challenge description	Current and possible solutions
Feedback	Multimodal sensory feedback	
	The sensory feedback, namely the possibility to restore the feel of interaction with the external world, represents the last key element of ULPs. Current ULP systems rely solely on vision as feedback information. However, we cannot consider just visual (unimodal) feedback to catch the complexity of the human-machine-environment interactions. Multimodal sensory feedback is necessary to empower the prosthetic control training and to trigger the embodiment processes.	Many groups are now exploring novel solution, both invasive and non-invasive, to provide sensory information about prosthetic movement. Regardless the specific methodology used, it is fundamental to achieve an intuitive or easily learnable strategy to associate the perceived feeling with a specific posture of the controlled device. The feedback-to-user can also be adjusted to the user through intelligent solutions, for instance through EMG biofeedback strategies based on the individual monitoring of the physiological input (Dosen *et al* 2015).

Table 7. Perspectives on tomorrow's upper limb prosthetics.

Future perspectives	Short-term non-invasive solutions	Long-term invasive solutions
Increase of input sources	Multiple input data to estimate the movement intentions; data fusion.	Increase the number of myoelectric input sites; neuroprosthetics.
Restoration of sensory feedback	Build artificial feedback, i.e. proprioception and tactile sensations by use of vibrotactile feedback or skin stretching.	Restore natural sensation, i.e. proprioception (by kinesthetic illusion), spatial sensation and phantom limb cortical representation (by refer touch strategies).
Closed-loop	Implementation of a bidirectional communication between user and prosthesis to restore a link between motor and sensory counterparts.	
User-prosthesis co-adaptation	Promote learning with engaging/immersive training and rehabilitation protocols (from user to prosthesis). Adaptive control, advanced PR algorithms (from prosthesis to user). Co-adaptive feedback (feedback-to-user and feedback-to-prosthesis).	
Miniaturization	Miniature technology to limit the encumbrance within the socket.	
Modular architecture	Modular prosthetic system enabling the progressive replacement of the non-invasive input and feedback sources with implanted ones.	
Standardization of amputation and surgery procedures	Standardization of level of amputation to help designing sockets that are simultaneously comfortable, anthropomorphic, and spacious (to integrate the circuitry and the power system).	
	—	Chronic and reliable implants.

6.2. Long-term invasive solutions

The great advantage of invasive approaches is a direct bidirectional contact with the nervous system. This comes at the cost of several issue related to the surgical procedure. Nevertheless, there are some promising approaches whose invasiveness drawbacks are counterbalanced by considerable improvements in device functionality, usability, and embodiment.

Moreover, although still far from clinical usage due to technological barriers, brain-based approaches seem a promising solution for prostheses of the future.

6.2.1. Peripheral bionics implants

Among the invasive approaches, surgical procedures aimed at augmenting the signal containing the motor commands have gained popularity in the last decades. Indeed, TMR is now routinely adopted in case of trans-humeral amputation, and it allows to increase the number of myoelectric input sites, which can be exploited for multi-DoFs control. Differently from invasive procedures for electrodes implantation, TMR is permanent, turning the reinnervated muscles into natural bioamplifiers of motor commands. It is also adopted for phantom limb pain reduction (Mereu *et al* 2021).

Similarly, the more recent RNPI represents another promising technique for bioamplification of motor signals and its successful demonstration in experimental trials encourages their potential adoption in clinical practice (Vu *et al* 2020b).

As for delivery of feedback information, invasive approaches represent the more intuitive and natural solution towards the retrieval of sensations, as demonstrated in (Osborn *et al* 2018, Nguyen *et al* 2021).

However, the main limitation toward the diffusion of these techniques is represented by their invasiveness, i.e. poor compatibility of electrodes, risk of infection due to external cables, scar tissue formation on the nerve, etc.

6.2.2. Neuroprostheses
Ideally, a prosthetic limb should be a perfect replication of the natural limb, both in terms of control and perception, such as Luke Skywalker's arm in the Star Wars saga. In this scenario, control signals should directly derive from the brain and communicate the intended movement to the robotic device, while sensory information should be encoded into stimulation patterns delivered to the brain. The field of neuroprosthetics, among other things, aims at addressing these fascinating goals and in the last 50 years several progresses have been made, indicating that these visionary scenarios might one day become true.

Fetz (1969) demonstrated that monkeys could voluntarily modulate the firing rates of neurons in the primary motor cortex, in the absence of movement. At the same time, Humphrey *et al* (1970) were able to predict arm displacement from the activity recorded from small populations of neurons in the motor cortex. These exciting and pioneering works thus proved the possibility of controlling artificial devices with the mind and eventually led to a rapid flourishing of investigations aimed at interfacing the brain with machines. These studies culminated with the first demonstration by the group of Nicolelis, of a robotic arm controlled with signals produced by an ensemble of neurons recorded from the motor cortex of a rat (Chapin *et al* 1999). At the beginning of this century, BMIs were thus born and scientists were therefore hoping that in few decades, fully functional bionic limbs would have been routinely adopted by amputees and paralyzed individuals (Nicolelis and Chapin 2002). Sadly, this is not at all how the story ended. Indeed, more than 20 years after the first demonstrations of brain-controlled devices, we still do not have the technology nor the computational capabilities to effectively control artificial devices with cortical brain signals.

However, in the last few years, some groups have presented promising examples of paralyzed individuals with neural implants in the motor and premotor cortices controlling artificial limbs for several months/year, while other groups worked on non-invasive applications on neurological populations, as detailed in section 3.1.1.3. Although the target population of these studies is mostly composed by stroke or paralyzed patients, exploitation of results for prosthetics applications clearly emerges, i.e. the possibility to perform device control by reliably and timely accessing to the subject's motor intentions. These examples indeed demonstrate that groundwork for brain control of motor prosthetics has been laid. However, it has been limited to the lab and mostly addressing paralyzed patients, for whom there are currently not viable solutions to enable dexterous device control as for amputees, whose residual motor functions can be successfully leveraged for prosthetic control signals.

In sum, brain control approaches are still far from clinical and personal applications, not only because of the poor controllability that they exert over the prosthetic device, but mainly because of the cumbersome apparatus they need for their collection and processing. However, the dream of brain-controlled devices has spread outside the academic labs and has contaminated also visionary entrepreneurs from venture capitals and tech giants, with the consequent birth of some important companies interested in brain-interfacing technology, such as Neuralink (Musk 2021), Facebook Reality Labs (Zuckerberg 2021), and Google DeepMind (Deepmind 2021). In conclusion, cutting edge research that we are currently witnessing both in academic and non-academic contexts may thus soon push the envelope of neuroprosthetics up to its diffusion in our everyday life, with important consequences also for amputees.

6.3. User-prosthesis co-adaptation
Designing user-centered prosthetic devices and user-centered prosthetic trainings is necessary to guide an appropriate learning of the system, as explained in section 5.1. Indeed, motivating the user to exercise and to get practice in using the bionic hand constitute the main way to improve the prosthetic control. Such approach can facilitate and accelerate the co-adaptation between humans and machines, as described in the following.

Humans implicitly learn how to control devices, even if initially they must adopt explicit strategies. Indeed, it must be underlined that ULP control training is an important step of rehabilitation. In particular, the ability of generating distinct muscle contractions increase with time and exercise. The use of functional tasks, like target achievement test (Simon *et al* 2011) or activities of daily living, allows users to learn how to produce repeatable patterns of contraction to better control the prosthesis. However, the learning process can be long and sometimes stressful, as described in (Zecca *et al* 2002). The development of more engaging training tasks and of a more immersive rehabilitation protocol could promote the learning process by increasing the engagement of the users (Roche *et al* 2019). In this context, user-centered design can truly make a difference in the effectiveness of a training procedure, as better discuss in section 5.

While humans have to learn ULP control, machines need to be trained with growing datasets for classifying the signals in terms of user's commands. For effective ULP control, PR-based algorithms currently represent the most effective solution, as they are able to recognize the human intentions on the basis of training data. An important aspect that can affect the accuracy of the classifiers is thus the way in which these data are collected. Indeed, the prosthesis control might not ensure good performances under different arm positions and several studies have been conducted on the evaluation of the impact of upper limb position during the data acquisition on classifier performance (Geng *et al* 2017). For example, as far as EMG is concerned, muscle activation is not completely the same when performing a given movement under different elbow and shoulder configurations. The signals on which the algorithm is trained are thus different from the ones obtained in a daily living scenario. Indeed, the classification performances strictly depend on the labeled data assigned to the specific movement. Moreover, because the method used for the acquisition strongly affects the classification accuracy, it is important to collect data under the same conditions of ADLs, i.e. by wearing the prosthesis, in order to have the training signals as similar as possible to the online ones. Cipriani *et al* (2011) highlighted that indeed EMG signals do not carry just information about the desired arm movement, but they also contain the muscular contribution to sustain the prosthesis weight. This aspect has to be taken into account in order to avoid misclassification and unwanted prosthetic movements, because as soon as the signals change, the classifier is no longer able to behave properly.

To cope with these problems, a possible solution consists in the adaptation of the control algorithm while using the prosthesis. This is of paramount importance, since the biggest issue of ULP control lies in the variability of the sEMG input signal due to electrodes shift, muscle fatigue, sweat, etc. ML algorithms are highly performing after training, but deterioration of the input signal or subsequent doffing/donning of the prosthesis can lead to misclassifications. For these reasons, amputees must regularly perform the training from scratch of the algorithm, which typically takes long time (Phinyomark and Scheme 2018). To address this issue, the incremental learning focuses on the adjustment of internal weights of the model without the need of re-training (Gijsberts *et al* 2014). In this way the training data is continuously updated, and the control system is capable to cope with possible sources of errors (i.e. unwanted changes in input signals).

A promising approach to innovate the ULP field is the adoption of co-adaptive features through bidirectional human-machine interfaces (De Santis 2021). In this case, the human and the machine reciprocally adjust their activity in order to improve their task-specific joint performance as a human-machine system. Co-adaptive features constitute a convergence of user activity and machine activity (including feedback toward the user and the machine), i.e. an 'agreement' between human and machine on the biosignals produced by the first and on their interpretation performed by the second. Co-adaptation can modulate both (human and machine) training processes within the same framework (Digiovanna *et al* 2008, Yeung *et al* 2019, Zbyszynski *et al* 2019, Igual Bañó 2021). This is performed through the feedback of human and machine on each other, further supporting the opportunity of designing feedback-to-user and feedback-to-prosthesis. Such a research effort requires innovative approaches to training processes as interactions between human and machine. In the next we present two building blocks for a user-prosthesis co-adaptation: closed-loop sensorimotor systems and interactive training.

6.3.1. Toward closed-loop sensorimotor prosthesis

While great progress has been made in recognizing human motor intention and translating it into prosthesis joint movements, sensory feedback restoration is one of the many challenges that many research groups are still addressing (see section 3.2) (Farina *et al* 2021). Obtaining a reliable and efficient way to artificially convey sensory information to prosthetic users would allow to develop smart devices able to truly mimic the behavior of human limbs, establishing the premise for a true co-adaptation between user and machine.

However, such a sensorimotor prosthesis would still need to face technological barriers for its development and use in ADLs. For example, all the circuitry and components needed to operate the device (i.e. acquire signals for motor intention decoding and translate sensory information into stimulation patterns) should fit into the socket space. One possible solution to obtain an embedded device is to rely on a hybrid approach. For example, a non-invasive sensorimotor prosthesis could use EMG signals to extract motor intentions and vibrotactile feedback to deliver sensory information. This configuration requires the embedding of an analog to digital converter (ADC) amplifier to acquire EMG data, and motor drivers to control vibromotors. All these components could be placed on the same board, overcoming the problem of electrical coupling.

However, although this design is very articulated, we believe that this is not the real end point for a fully-integrated prosthetic system capable of maximizing patient acceptance. It simply turns out to be a developmental test bench for what will be the next generation of prostheses on a longer temporal horizon, namely neuroprostheses. By creating a modular architecture, in fact, it will be possible to maintain the physical prosthetic system and to progressively replace the non-invasive input and feedback sources with

implanted ones, namely: (a) the standard sensors for surface EMG replaced by an intraneural implant; (b) the hardware for feedback delivery replaced by a chronic implant; and (c) the hardware on which run the control strategies replaced by a smaller and more performing dedicated hardware able to process neuromorphic algorithms (e.g. Application specific integrated circuit (ASIC)).

The realization of bidirectional ULP systems able to restore both sensory and motor functions will open new research scenario for embodiment process and neuroplastic phenomena. Indeed, about neuroplasticity in amputees using bionic hands as prostheses, Di Pino *et al* (2009) highlighted how the reorganization of the central nervous system after the usage of the device can be the source of indices of prosthetic effectiveness in functional recovery. Furthermore, the effects of the device on the central nervous system can make the prosthesis work as a neurorehabilitative solution mitigating aberrant plasticity phenomena and facilitating positive neural changes. Finally, novel human-machine interfaces should consider neuroplasticity principles for restoring the efferences and afferences of the central nervous system with the lost limb in order to exploit them for connecting a prosthesis.

The processes described above are an excellent expression of the technologies that can lead to a user-prosthesis co-adaptation. They care for providing the human with sensations matching the motor activity and the events occurring on or for the artificial limb, which needs to learn how to offer appropriate feedback to the user. Next paragraphs will describe how this can happen.

6.3.2. Innovation in interactive training

Establishing interactions between user and prosthesis requires a human-centered design of the technology 'behavior'. Since the term co-adaptation implies that two or more entities are adjusting to each other, possibly learning through the interaction itself for reaching a goal that can be the improvement of the human-machine performance. The attention to the interactive aspects of learning and training was recommended in Castellini *et al* (2016). In this context, intriguing opportunities come from theoretical frameworks in psychology like the constructivism to define and improve the paradigm of interactive ML in prosthetic training and control (Nowak *et al* 2018, Bettoni and Castellini 2021). Accordingly, a myocontrol system should learn and forget on demand, under request of each one of the components of the human-machine system. For instance, the users can label the violation of their expectations on the prosthesis interpretation of their commands, starting a novel data collection cycle. Additionally, the machine itself can highlight the need of collecting further data (especially from correct execution of a training exercise) through acoustic feedback to the user. Through this, the myocontrol models are updated.

Moreover, certain biosignal features could guide an automated labeling of the violation of the observer's expectations without any explicit command, as in neuroprosthetic interfaces using Error-related Potentials (Chavarriaga *et al* 2014). Such an advance (still explored in laboratories) could enable self-calibrating intention detection processes, leading to a true (and fruitful) human-machine symbiosis. Such symbiosis is based on the online processing of users' neurocognitive states. The machine interprets such states for implicitly adjusting (without direct and declarative commands of the user) its activity to the individual current capabilities and preferences. Exploiting this process as a further example of the closed-loop previously envisioned, the human-machine system will become a fully functional unit able to enact manual and bimanual biomimetic behaviors (Chavarriaga *et al* 2014).

However, when the practical applications of this kind of implicit learning will move outside the laboratories, its advantages should be evaluated in real-world cases of prosthetic learning. Currently, a human-machine explicit communication component in initial co-training sessions can highlight an active role of the users with positive impact on their self-efficacy and engagement. In this context, biofeedback strategies (self-regulation of perceptually represented physiological changes) in co-adaptive systems could be especially useful for easing the integration of both implicit and explicit aspects within the user-prosthesis co-training (even during daily re-calibrations) (Kalampratsidou and Torres 2017).

6.4. Amputation matters: the key role of the surgeon

A further reflection that emerges in order to maximize patient acceptance is related to the clinical situation immediately preceding amputation. In fact, although it is clear that in case of injury/accident the doctor/surgeon has to manage a situation of immediate danger in which the priority is to secure the patient trying to save 'as much as possible of the injured limb', on the other hand, on the engineer side, the ideal would be to be able to standardize the level of amputation. In this case, in fact, it would be possible to avoid 'extreme' situations in terms of amputation levels: a very distal amputation does not allow the orthopedic technician, and therefore also the engineer, to have enough space to integrate the circuitry and the power system inside the cavity between the internal and external lamination of the reservoir itself, on the other hand, a very proximal amputation does not allow the orthopedic technician to create a socket capable of allowing the patient's stump to support the weight (thus not even having enough residual muscles from

which to extract the EMG signal) of the dedicated prosthetic system. Therefore, having the possibility, even in case of extreme danger, to be able to define a 'standard' level of amputation in which the surgeon is able to have a slightly longer-term vision, would allow a greater number of patients to take advantage of these fantastic biomedical technologies. In fact, at present, many patients find themselves having to request an additional surgical operation, thus further compromising the patient's willingness to use these solutions.

7. Conclusions

In this manuscript, we have detailed and discussed several strategies to substitute and restore the functionality of human upper limb when missing. We specifically focused on input and feedback signals for bidirectional device control and on aspects regarding the user needs that should be addressed with a user-centered prosthetic design approach. We also stressed that in order to have a fully embodied prosthetic system it is essential to implement a synergistic collaboration of three components: mechatronics, control algorithm and the user perception. These have to be combined by a real rehabilitation process that is necessarily user-centered.

Acknowledgments

Conflict of interest

The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

Author contributions

A M, N B, F T, D D D, G B, and M S conceived the study. N B, A M and M S designed the figures. N B and A M prepared the figures. All the authors contributed to the writing, read and approved the final version of the manuscript.

ORCID iDs

Andrea Marinelli ⬤ https://orcid.org/0000-0001-8672-5202
Dario Di Domenico ⬤ https://orcid.org/0000-0002-4603-1692
Marianna Semprini ⬤ https://orcid.org/0000-0001-5504-0251

References

Adrian E D and Bronk D W 1929 The discharge of impulses in motor nerve fibres: part II. The frequency of discharge in reflex and voluntary contractions *J. Physiol.* **67** 9–151
Ahmadi R, Packirisamy M, Dargahi J and Cecere R 2011 Discretely loaded beam-type optical fiber tactile sensor for tissue manipulation and palpation in minimally invasive robotic surgery *IEEE Sens. J.* **12** 22–32
Ajoudani A, Godfrey S B, Catalano M, Grioli G, Tsagarakis N G and Bicchi A 2013 Teleimpedance control of a synergy-driven anthropomorphic hand *2013 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems* (IEEE) pp 1985–91
Alameh M, Abbass Y, Ibrahim A and Valle M 2020 Smart tactile sensing systems based on embedded CNN implementations *Micromachines* **11** 103
Alameh M, Saleh M, Ansovini F, Fares H, Ibrahim A, Franceschi M, Seminara L, Valle M, Dosen S and Farina D 2018 Live demonstration: system based on electronic skin and cutaneous electrostimulation for sensory feedback in prosthetics *2018 IEEE Biomedical Circuits and Systems Conf. (Biocas)* (IEEE) p 1
Alfadhel A and Kosel J 2015 Magnetic nanocomposite cilia tactile sensor *Adv. Mater.* **27** 7888–92
Alkhafaf O S, Wali M K and Al-timemy A H 2020 Improved prosthetic hand control with synchronous use of voice recognition and inertial measurements *IOP Conf. Ser.: Mater. Sci. Eng.* **745** 012088
Amado Laezza R 2018 Deep neural networks for myoelectric pattern recognition-an implementation for multifunctional control
Amrani M Z-E-A, Daoudi A, Achour N and Tair M 2017 Artificial neural networks based myoelectric control system for automatic assistance in hand rehabilitation *2017 26th IEEE Int. Symp. on Robot and Human Interactive Communication (RO-MAN)* (IEEE) pp 968–73
Anderson F and Bischof W F 2014 Augmented reality improves myoelectric prosthesis training *Int. J. Disabil. Hum. Dev.* **13** 349–54
Antfolk C, Cipriani C, Carrozza M C, Balkenius C, Björkman A, Lundborg G, Rosén B and Sebelius F 2013a Transfer of tactile input from an artificial hand to the forearm: experiments in amputees and able-bodied volunteers *Disabil. Rehabil. Assist. Technol.* **8** 249–54
Antfolk C, D'alonzo M, Rosén B, Lundborg G, Sebelius F and Cipriani C 2013b Sensory feedback in upper limb prosthetics *Expert Rev. Med. Devices* **10** 45–54

Aopa A O P A 2016 Where science meets art (American Orthotic & Prosthetic Association (AOPA))

Bark K, Hyman E, TAN F, CHA E, Jax S A, Buxbaum L J and Kuchenbecker K J 2014 Effects of vibrotactile feedback on human learning of arm motions *IEEE Trans. Neural Syst. Rehabil. Eng.* **23** 51–63

Barresi G *et al* 2021 Exploring the embodiment of a virtual hand in a spatially augmented respiratory biofeedback setting *Front. Neurorobot.* **15** 683653

Barsakcioglu D Y and Farina D 2018 A real-time surface EMG decomposition system for non-invasive human-machine interfaces *2018 IEEE Biomedical Circuits and Systems Conf. (Biocas)* (IEEE) pp 1–4

Basu T 2021 Facebook is making a bracelet that lets you control computers with your brain (available at: www.technologyreview.com/2021/03/18/1021021/facebook-augmented-reality-wristband/?truid=6d1563793eef118b900759ed00bfef6f&utm_source=the_download&utm_medium=email&utm_campaign=the_download.unpaid.engagement&utm_term=&utm_content=03-19-2021&mc_cid=f1e3644b4e&mc_eid=de470f7d14) (Accessed 18 March 2021)

Battaglia E, Clark J P, Bianchi M, Catalano M G, Bicchi A and O'malley M K 2019 Skin stretch haptic feedback to convey closure information in anthropomorphic, under-actuated upper limb soft prostheses *IEEE Trans. Haptics* **12** 508–20

Beccai L, Roccella S, Arena A, Valvo F, Valdastri P, Menciassi A, Carrozza M C and Dario P 2005 Design and fabrication of a hybrid silicon three-axial force sensor for biomechanical applications *Sens. Actuators* A **120** 370–82

Beckerle P 2021 Virtual hand experience *Human-Robot Body Experience* (Springer)

Beckerle P, Castellini C and Lenggenhager B 2019 Robotic interfaces for cognitive psychology and embodiment research: a research roadmap *Wiley Interdiscip. Rev. Cogn. Sci.* **10** e1486

Beckerle P, Kõiva R, Kirchner E A, Bekrater-bodmann R, Dosen S, Christ O, Abbink D A, Castellini C and Lenggenhager B 2018 Feel-good robotics: requirements on touch for embodiment in assistive robotics *Front. Neurorobot.* **12** 84

Belter J T and Dollar A M 2011 Performance characteristics of anthropomorphic prosthetic hands *2011 IEEE Int. Conf. on Rehabilitation Robotics* (IEEE) pp 1–7

Belter J T, Segil J L, Dollar A M and Weir R F 2013 Mechanical design and performance specifications of anthropomorphic prosthetic hands: a review *J. Rehabil. Res. Dev.* **50** 599

Bennett D A, Dalley S A, Truex D and Goldfarb M 2014 A multigrasp hand prosthesis for providing precision and conformal grasps *IEEE/ASME Trans. Mechatronics* **20** 1697–704

Benvenuto A *et al* 2010 Intrafascicular thin-film multichannel electrodes for sensory feedback: evidences on a human amputee *2010 Annual Int. Conf. IEEE Engineering in Medicine and Biology* (IEEE) pp 1800–03

Bettoni M C and Castellini C 2021 Interaction in assistive robotics: a radical constructivist design framework *Front. Neurorobot.* **15** 675657

Biddiss E A and Chau T T 2007 Upper limb prosthesis use and abandonment: a survey of the last 25 years *Prosthet. Orthot. Int.* **31** 236–57

Biddiss E, Beaton D and Chau T 2007 Consumer design priorities for upper limb prosthetics *Disabil. Rehabil. Assist. Technol.* **2** 346–57

Bionics M 2022 *The Luke arm* (available at: www.mobiusbionics.com/) (https://doi.org/10.1080/03091902.2022.2095049)

Bleichner M G, Freudenburg Z V, Jansma J M, Aarnoutse E J, Vansteensel M J and Ramsey N F 2016 Give me a sign: decoding four complex hand gestures based on high-density ECoG *Brain Struct. Funct.* **221** 203–16

Boostani R and Moradi M H 2003 Evaluation of the forearm EMG signal features for the control of a prosthetic hand *Physiol. Meas.* **24** 309–19

Boschmann A, Neuhaus D, Vogt S, Kaltschmidt C, Platzner M and Dosen S 2021 Immersive augmented reality system for the training of pattern classification control with a myoelectric prosthesis *J. Neuroeng. Rehabil.* **18** 1–15

Botvinick M and Cohen J 1998 Rubber hands 'feel' touch that eyes see *Nature* **391** 756

Bouwsema H, Van Der Sluis C K and Bongers R M 2014 Effect of feedback during virtual training of grip force control with a myoelectric prosthesis *PLoS One* **9** e98301

Brainrobotics 2022 *BrainRobotics hand* (available at: https://brainrobotics.com/brainrobotics-hand/)

Burrello A, Morghet F B, Scherer M, Benatti S, Benini L, Macii E, Poncino M and Pagliari D J 2022 Bioformers: embedding transformers for ultra-low power sEMG-based gesture recognition *2022 Design, Automation & Test in Europe Conf. & Exhibition (DATE)* (IEEE) pp 1443–8

Carey S L, Lura D J and Highsmith M J 2015 Differences in myoelectric and body-powered upper-limb prostheses: systematic literature review *J. Rehabil. Res. Dev.* **52** 247–62

Carrozza M C, Persichetti A, Laschi C, Vecchi F, Lazzarini R, Vacalebri P and Dario P 2007 A wearable biomechatronic interface for controlling robots with voluntary foot movements *IEEE/ASME Trans. Mechatronics* **12** 1–11

Castellini C 2020 Upper limb active prosthetic systems—overview *Wearable Robotics* (Amsterdam: Elsevier)

Castellini C, Bongers R M, Nowak M and Van Der Sluis C K 2016 Upper-limb prosthetic myocontrol: two recommendations *Front. Neurosci.* **9** 496

Castillo C S M, Wilson S, Vaidyanathan R and Atashzar S F 2020 Wearable MMG-plus-one armband: evaluation of normal force on mechanomyography (MMG) to enhance human-machine interfacing *IEEE Trans. Neural Syst. Rehabil. Eng.* **29** 196–205

Catalano M G, Grioli G, Farnioli E, Serio A, Piazza C and Bicchi A 2014 Adaptive synergies for the design and control of the Pisa/IIT SoftHand *Int. J. Robot. Res.* **33** 768–82

Chapin J K, Moxon K A, Markowitz R S and Nicolelis M A 1999 Real-time control of a robot arm using simultaneously recorded neurons in the motor cortex *Nat. Neurosci.* **2** 664–70

Chavarriaga R, Sobolewski A and Millán J D R 2014 Errare machinale est: the use of error-related potentials in brain-machine interfaces *Front. Neurosci.* **8** 208

Cheesborough J E, Smith L H, Kuiken T A and Dumanian G A 2015 Targeted muscle reinnervation and advanced prosthetic arms *Semin. Plast. Surg.* **29** 062–72

Chen J, BI S, Zhang G and Cao G 2020 High-density surface EMG-based gesture recognition using a 3D convolutional neural network *Sensors* **20** 1201

Chen X, Zhang X, Zhao Z-Y, Yang J-H, Lantz V and Wang K-Q 2007 Hand gesture recognition research based on surface EMG sensors and 2D-accelerometers *2007 11th IEEE Int. Symp. on Wearable Computers* (IEEE) pp 11–14

Cheng J, Amft O, Bahle G and Lukowicz P 2013 Designing sensitive wearable capacitive sensors for activity recognition *IEEE Sens. J.* **13** 3935–47

Cho E, Chen R, Merhi L-K, Xiao Z, Pousett B and Menon C 2016 Force myography to control robotic upper extremity prostheses: a feasibility study *Front. Bioeng. Biotechnol.* **4** 18

Ciancio A L *et al* 2016 Control of prosthetic hands via the peripheral nervous system *Front. Neurosci.* **10** 116

Cipriani C, Sassu R, Controzzi M and Carrozza M C 2011 Influence of the weight actions of the hand prosthesis on the performance of pattern recognition based myoelectric control: preliminary study *2011 Annual Int. Conf. IEEE Engineering in Medicine and Biology Society* (IEEE) pp 1620–3

Cipriani C, Segil J L, Birdwell J A and Ff WEIR R F 2014 Dexterous control of a prosthetic hand using fine-wire intramuscular electrodes in targeted extrinsic muscles *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 828–36

Cipriani C, Zaccone F, Micera S and Carrozza M C 2008 On the shared control of an EMG-controlled prosthetic hand: analysis of user–prosthesis interaction *IEEE Trans. Robot.* **24** 170–84

Clemente F, D'alonzo M, Controzzi M, Edin B B and Cipriani C 2015 Non-invasive, temporally discrete feedback of object contact and release improves grasp control of closed-loop myoelectric transradial prostheses *IEEE Trans. Neural Syst. Rehabil. Eng.* **24** 1314–22

Clemente F, Dosen S, Lonini L, Markovic M, Farina D and Cipriani C 2016 Humans can integrate augmented reality feedback in their sensorimotor control of a robotic hand *IEEE Trans. Hum.-Mach. Syst.* **47** 583–9

Clemente F, Ianniciello V, Gherardini M and Cipriani C 2019 Development of an embedded myokinetic prosthetic hand controller *Sensors* **19** 3137

Cloutier A and Yang J 2013a Control of hand prostheses: a literature review *Int. Design Engineering Technical Conf. and Computers and Information in Engineering Conf.* (American Society of Mechanical Engineers) p V06AT07A016

Cloutier A and Yang J 2013b Design, control, and sensory feedback of externally powered hand prostheses: a literature review *Critical Rev. Biomed. Eng.* **41** 161–81

Coapt 2017 *COAPT—Complete Control Handbook* (available at: https://coaptengineering.com/wp-content/uploads/2020/03/coapt_gen2_handbook_v3.0.pdf)

Collinger J L, Wodlinger B, Downey J E, Wang W, Tyler-kabara E C, Weber D J, Mcmorland A J, Velliste M, Boninger M L and Schwartz A B 2013 High-performance neuroprosthetic control by an individual with tetraplegia *Lancet* **381** 557–64

Company S R 2020 *Shadow Robot* (available at: www.shadowrobot.com/)

Controzzi M, Clemente F, Barone D, Ghionzoli A and Cipriani C 2016 The SSSA-MyHand: a dexterous lightweight myoelectric hand prosthesis *IEEE Trans. Neural Syst. Rehabil. Eng.* **25** 459–68

Cordella F, Ciancio A L, Sacchetti R, Davalli A, Cutti A G, Guglielmelli E and Zollo L 2016 Literature review on needs of upper limb prosthesis users *Front. Neurosci.* **10** 209

Corke P I and Khatib O 2011 *Robotics, Vision and Control: Fundamental Algorithms in MATLAB* (Berlin: Springer)

Cuberovic I, Gill A, Resnik L J, Tyler D J and Graczyk E L 2019 Learning of artificial sensation through long-term home use of a sensory-enabled prosthesis *Front. Neurosci.* **13** 853

Cutrone A and Micera S 2019 Implantable neural interfaces and wearable tactile systems for bidirectional neuroprosthetics systems *Adv. Healthcare Mater.* **8** 1801345

D'avella A and Bizzi E 2005 Shared and specific muscle synergies in natural motor behaviors *Proc. Natl Acad. Sci.* **102** 3076–81

Davis C and Onge M S 2017 Myoelectric and body-powered upper-limb prostheses: the users' perspective *J. Prosthet. Orthot.* **29** P30–4

De Luca C J 1997 The use of surface electromyography in biomechanics *J. Appl. Biomech.* **13** 135–63

De Oliveira Barata S 2021 The alternative limb project (available at: https://thealternativelimbproject.com/)

De Santis D 2021 A framework for optimizing co-adaptation in body-machine interfaces *Front. Neurorobot.* **15** 40

Deepmind 2021 DeepMind (available at: www.deepmind.com/)

Del Valle J and Navarro X 2013 Interfaces with the peripheral nerve for the control of neuroprostheses *Int. Rev. Neurobiol.* **109** 63–83

Del Vecchio A, Castellini C and Beckerle P 2021 Peripheral neuroergonomics—an elegant way to improve human-robot interaction? *Front. Neurorobot.* **15**

Dellacasa Bellingegni A, Gruppioni E, Colazzo G, Davalli A, Sacchetti R, Guglielmelli E and Zollo L 2017 NLR, MLP, SVM, and LDA: a comparative analysis on EMG data from people with trans-radial amputation *J. Neuroeng. Rehabil.* **14** 1–16

Deng H, Xu X, Zhuo W and Zhang Y 2020 Current-sensor-based contact stiffness detection for prosthetic hands *IEEE Access* **8** 29456–66

Dhawan A S, Mukherjee B, Patwardhan S, Akhlaghi N, Diao G, Levay G, Holley R, Joiner W M, Harris-love M and Sikdar S 2019 Proprioceptive sonomyographic control: a novel method for intuitive and proportional control of multiple degrees-of-freedom for individuals with upper extremity limb loss *Sci. Rep.* **9** 1–15

Di Domenico D *et al* 2021 Hannes prosthesis control based on regression machine learning algorithms *2021 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems (IROS 2021)*

Di Pino G, Guglielmelli E and Rossini P M 2009 Neuroplasticity in amputees: main implications on bidirectional interfacing of cybernetic hand prostheses *Prog. Neurobiol.* **88** 114–26

Digiovanna J, Mahmoudi B, Fortes J, Principe J C and Sanchez J C 2008 Coadaptive brain–machine interface via reinforcement learning *IEEE Trans. Biomed. Eng.* **56** 54–64

Dimante D, Logina I, Sinisi M and Krūmiņa A 2020 Sensory feedback in upper limb prostheses *Proc. Latvian Acad. Sci.* B 308–17 Natural, Exact, and Applied Sciences

Dornfeld C, Swanston M, Cassella J, Beasley C, Green J, Moshayev Y and Wininger M 2016 Is the prosthetic homologue necessary for embodiment? *Front. Neurorobot.* **10** 21

Dosen S, Markovic M, Somer K, Graimann B and Farina D 2015 EMG biofeedback for online predictive control of grasping force in a myoelectric prosthesis *J. Neuroeng. Rehabil.* **12** 1–13

Drost G, Stegeman D F, Van Engelen B G and Zwarts M J 2006 Clinical applications of high-density surface EMG: a systematic review *J. Electromyogr. Kinesiol.* **16** 586–602

Edelman B J, Meng J, Suma D, Zurn C, Nagarajan E, Baxter B, Cline C C and HE B 2019 Noninvasive neuroimaging enhances continuous neural tracking for robotic device control *Sci. Robot.* **4**

Ehrsson H H, Rosén B, Stockselius A, Ragnö C, Köhler P and Lundborg G 2008 Upper limb amputees can be induced to experience a rubber hand as their own *Brain* **131** 3443–52

Engdahl S M, Meehan S K and Gates D H 2020b Differential experiences of embodiment between body-powered and myoelectric prosthesis users *Sci. Rep.* **10** 1–10

Engdahl S, Dhawan A, Lévay G, Bashatah A, Kaliki R and Sikdar S 2020a Motion prediction using electromyography and sonomyography for an individual with transhumeral limb loss *medRxiv*

Englehart K, Hudgins B and Chan A D 2003 Continuous multifunction myoelectric control using pattern recognition *Technol. Disabil.* **15** 95–103

Esposito D, Andreozzi E, Fratini A, Gargiulo G D, Savino S, Niola V and Bifulco P 2018 A piezoresistive sensor to measure muscle contraction and mechanomyography *Sensors* **18** 2553

Farina D, LI X and Madeleine P 2008 Motor unit acceleration maps and interference mechanomyographic distribution *J. Biomech.* **41** 2843–9

Farina D, Merletti R and Enoka R M 2004 The extraction of neural strategies from the surface EMG *J. Appl. Physiol.* **96** 1486–95

Farina D, Vujaklija I, Brånemark R, Bull A M, Dietl H, Graimann B, Hargrove L J, Hoffmann K-P, Huang H H and Ingvarsson T 2021 Toward higher-performance bionic limbs for wider clinical use *Nat. Biomed. Eng.* 1–13

Ferigo D, Merhi L-K, Pousett B, Xiao Z G and Menon C 2017 A case study of a force-myography controlled bionic hand mitigating limb position effect *J. Bionic Eng.* **14** 692–705

Fetz E E 1969 Operant conditioning of cortical unit activity *Science* **163** 955–8

Fifer M S, Hotson G, Wester B A, Mcmullen D P, Wang Y, Johannes M S, Katyal K D, Helder J B, Para M P and Vogelstein R J 2013 Simultaneous neural control of simple reaching and grasping with the modular prosthetic limb using intracranial EEG *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 695–705

Figliolia A, Medola F, Sandnes F, Rodrigues A C T and Paschoarelli L C 2019 Avoiding product abandonment through user centered design: a case study involving the development of a 3D printed customized upper limb prosthesis *Int. Conf. on Applied Human Factors and Ergonomics* (Springer) pp 289–97

Fillauer 2021 *Motion Control Hand* (available at: https://fillauer.com/products/motion-control-hand/)

Fillauer 2022a *Motion E2 Elbow* (available at: https://fillauer.com/products/motion-e2-elbow/)

Fillauer 2022b *Utah Arm 3* (available at: https://fillauer.com/products/utah-arm-3/)

Fishel J A and Loeb G E 2012 Sensing tactile microvibrations with the BioTac—comparison with human sensitivity *2012 4th IEEE RAS & EMBS Int. Conf. on Biomedical Robotics and Biomechatronics (Biorob)* (IEEE) pp 1122–7

Fougner A, Scheme E, Chan A D, Englehart K and Stavdahl Ø 2011 A multi-modal approach for hand motion classification using surface EMG and accelerometers *2011 Annual Int. Conf. IEEE Engineering in Medicine and Biology Society* (IEEE) pp 4247–50

Fuentes-Gonzalez J, Infante-alarcón A, Asanza V and Loayza F R A 2021 3D-Printed EEG based prosthetic arm *2020 IEEE Int. Conf. on E-health Networking, Application & Services (HEALTHCOM)* (IEEE) pp 1–5

Fujiwara E, WU Y T, Suzuki C K, De Andrade D T G, Neto A R and Rohmer E 2018 Optical fiber force myography sensor for applications in prosthetic hand control *2018 IEEE 15th Int. Workshop on Advanced Motion Control (AMC)* (IEEE) pp 342–7

Fukuma R, Yanagisawa T, Saitoh Y, Hosomi K, Kishima H, Shimizu T, Sugata H, Yokoi H, Hirata M and Kamitani Y 2016 Real-time control of a neuroprosthetic hand by magnetoencephalographic signals from paralysed patients *Sci. Rep.* **6** 1–14

Furui A, Eto S, Nakagaki K, Shimada K, Nakamura G, Masuda A, Chin T and Tsuji T 2019 A myoelectric prosthetic hand with muscle synergy–based motion determination and impedance model–based biomimetic control *Sci. Robot.* **4** eaaw6339

Gallagher M, Colzi C and Sedda A 2021 Dissociation of proprioceptive drift and feelings of ownership in the somatic rubber hand illusion *Acta Psychol.* **212** 103192

Garenfeld M A, Mortensen C K, Strbac M, Dideriksen J L and Dosen S 2020 Amplitude versus spatially modulated electrotactile feedback for myoelectric control of two degrees of freedom *J. Neural Eng.* **17** 046034

Garske C A, Dyson M, Dupan S, Morgan G and Nazarpour K 2021a Serious games are not serious enough for myoelectric prosthetics *JMIR Serious Games* **9** e28079

Garske C A, Dyson M, Dupan S and Nazarpour K 2021b Perception of game-based rehabilitation in upper limb prosthetic training: survey of users and researchers *JMIR Serious Games* **9** e23710

Geng W, DU Y, Jin W, WEI W, HU Y and LI J 2016 Gesture recognition by instantaneous surface EMG images *Sci. Rep.* **6** 1–8

Geng Y, Samuel O W, WEI Y and LI G 2017 Improving the robustness of real-time myoelectric pattern recognition against arm position changes in transradial amputees *Biomed. Res. Int.* **2017** 1–10

George J A, Kluger D T, Davis T S, Wendelken S M, Okorokova E, HE Q, Duncan C C, Hutchinson D T, Thumser Z C and Beckler D 2019 Biomimetic sensory feedback through peripheral nerve stimulation improves dexterous use of a bionic hand *Sci. Robot.* **4**

Georgi M, Amma C and Schultz T 2015 Recognizing hand and finger gestures with IMU based motion and EMG based muscle activity sensing *Biosignals* (Citeseer) pp 99–108

Gigli A, Brusamento D, Meattini R, Melchiorri C and Castellini C 2020 Feedback-aided data acquisition improves myoelectric control of a prosthetic hand *J. Neural Eng.* **17** 056047

Gijsberts A, Bohra R, Sierra González D, Werner A, Nowak M, Caputo B, ROA M A and Castellini C 2014 Stable myoelectric control of a hand prosthesis using non-linear incremental learning *Front. Neurorobot.* **8** 8

Godfrey S B, Zhao K D, Theuer A, Catalano M G, Bianchi M, Breighner R, Bhaskaran D, Lennon R, Grioli G and Santello M 2018 The SoftHand Pro: functional evaluation of a novel, flexible, and robust myoelectric prosthesis *PLoS One* **13** e0205653

Gonzalez J, Soma H, Sekine M and YU W 2012 Psycho-physiological assessment of a prosthetic hand sensory feedback system based on an auditory display: a preliminary study *J. Neuroeng. Rehabil.* **9** 1–14

Gonzalez M A, LEE C, Kang J, Gillespie R B and Gates D H 2021 Getting a grip on the impact of incidental feedback from body-powered and myoelectric prostheses *IEEE Trans. Neural Syst. Rehabil. Eng.* **29** 1905–1912

Gow D, Douglas W, Geggie C, Monteith E and Stewart D 2001 The development of the Edinburgh modular arm system *Proc. Inst. Mech. Eng. H* **215** 291–298

Grebenstein M, Albu-schäffer A, Bahls T, Chalon M, Eiberger O, Friedl W, Gruber R, Haddadin S, Hagn U and Haslinger R 2011 The DLR hand arm system *2011 IEEE Int. Conf. on Robotics and Automation* (IEEE) pp 3175–82

Grechuta K, Guga J, Maffei G, Ballester B R and Verschure P F 2017 Visuotactile integration modulates motor performance in a perceptual decision-making task *Sci. Rep.* **7** 1–13

Grushko S, Spurný T and Černý M 2020 Control methods for transradial prostheses based on remnant muscle activity and its relationship with proprioceptive feedback *Sensors* **20** 4883

Gunasekaran H, Spigler G, Mazzoni A, Cataldo E and Oddo C M 2019 Convergence of regular spiking and intrinsically bursting Izhikevich neuron models as a function of discretization time with Euler method *Neurocomputing* **350** 237–47

Guo S, Pang M, Gao B, Hirata H and Ishihara H 2015 Comparison of sEMG-based feature extraction and motion classification methods for upper-limb movement *Sensors* **15** 9022–38

Guo W, Sheng X, Liu H and Zhu X 2017a Mechanomyography assisted myoeletric sensing for upper-extremity prostheses: a hybrid approach *IEEE Sens. J.* **17** 3100–8

Guo W, Sheng X, Liu H and Zhu X 2017b Toward an enhanced human–machine interface for upper-limb prosthesis control with combined EMG and NIRS signals *IEEE Trans. Hum.-Mach. Syst.* **47** 564–75

Hagan M T, Demuth H B and Beale M 1997 *Neural Network Design* (PWS Publishing Co)

Hahne J M, Biessmann F, Jiang N, Rehbaum H, Farina D, Meinecke F C, Müller K-R and Parra L C 2014 Linear and nonlinear regression techniques for simultaneous and proportional myoelectric control *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 269–79

Hahne J M, Schweisfurth M A, Koppe M and Farina D 2018 Simultaneous control of multiple functions of bionic hand prostheses: performance and robustness in end users *Sci. Robot.* **3**

Hammock M L, Chortos A, Tee B C K, Tok J B H and Bao Z 2013 25th anniversary article: the evolution of electronic skin (e-skin): a brief history, design considerations, and recent progress *Adv. Mater.* **25** 5997–6038

Hao Y, Controzzi M, Cipriani C, Popovic D B, Yang X, Chen W, Zheng X and Carrozza M C 2013 Controlling hand-assistive devices: utilizing electrooculography as a substitute for vision *IEEE Robot. Autom. Mag.* **20** 40–52

Hargrove L J, Englehart K and Hudgins B 2007 A comparison of surface and intramuscular myoelectric signal classification *IEEE Trans. Biomed. Eng.* **54** 847–53

Harte R, Glynn L, Rodríguez-molinero A, Baker P M, Scharf T, Quinlan L R and Ólaighin G 2017 A human-centered design methodology to enhance the usability, human factors, and user experience of connected health systems: a three-phase methodology *JMIR Hum. Factors* **4** e5443

Hartwell A, Kadirkamanathan V and Anderson S R 2018 Compact deep neural networks for computationally efficient gesture classification from electromyography signals *2018 7th IEEE Int. Conf. on Biomedical Robotics and Biomechatronics (Biorob)* (IEEE) pp 891–6

Hazubski S, Hoppe H and Otte A 2020 Hand prosthetic controlled via augmented reality

Heckathorne C W and Childress D S 2001 Cineplasty as a control input for externally powered prosthetic components *J. Rehabil. Res. Dev.* **38**

Heerschop A, Van Der Sluis C K and Bongers R M 2021 Transfer of mode switching performance: from training to upper-limb prosthesis use *J. Neuroeng. Rehabil.* **18** 1–16

Hochberg L R, Bacher D, Jarosiewicz B, Masse N Y, Simeral J D, Vogel J, Haddadin S, LIU J, Cash S S and Van Der Smagt P 2012 Reach and grasp by people with tetraplegia using a neurally controlled robotic arm *Nature* **485** 372–5

Hochberg L R, Serruya M D, Friehs G M, Mukand J A, Saleh M, Caplan A H, Branner A, Chen D, Penn R D and Donoghue J P 2006 Neuronal ensemble control of prosthetic devices by a human with tetraplegia *Nature* **442** 164–71

Holobar A and Zazula D 2007 Multichannel blind source separation using convolution kernel compensation *IEEE Trans. Signal Process.* **55** 4487–96

Hotson G, Mcmullen D P, Fifer M S, Johannes M S, Katyal K D, Para M P, Armiger R, Anderson W S, Thakor N V and Wester B A 2016 Individual finger control of a modular prosthetic limb using high-density electrocorticography in a human subject *J. Neural Eng.* **13** 026017

Huang H H, SI J, Brandt A and Li M 2021 Taking both sides: seeking symbiosis between intelligent prostheses and human motor control during locomotion *Curr. Opin. Biomed. Eng.* **20** 100314

Hudgins B, Parker P and Scott R N 1993 A new strategy for multifunction myoelectric control *IEEE Trans. Biomed. Eng.* **40** 82–94

Humphrey D R, Schmidt E and Thompson W 1970 Predicting measures of motor performance from multiple cortical spike trains *Science* **170** 758–62

Huynh T V, Bekrater-bodmann R, Fröhner J, Vogt J and Beckerle P 2019 Robotic hand illusion with tactile feedback: unravelling the relative contribution of visuotactile and visuomotor input to the representation of body parts in space *PLoS One* **14** e0210058

I-Biomed 2021 Sense pattern recognition system (available at: www.i-biomed.com/sense.html)

Igual Bañó C 2021 Co-adaptive myoelectric control for upper limb prostheses (Universitat Politècnica de València)

Iskarous M M and Thakor N V 2019 E-skins: biomimetic sensing and encoding for upper limb prostheses *Proc. IEEE* **107** 2052–64

Jamal M Z 2012 Signal acquisition using surface EMG and circuit design considerations for robotic prosthesis *Computational Intelligence in Electromyography Analysis-A Perspective on Current Applications and Future Challenges* vol 18 pp 427–48

Jamali N, Maggiali M, Giovannini F, Metta G and Natale L 2015 A new design of a fingertip for the iCub hand *2015 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems (IROS)* (IEEE) pp 2705–10

Jiang N, Dosen S, Muller K-R and Farina D 2012 Myoelectric control of artificial limbs—is there a need to change focus? [In the spotlight] *IEEE Signal Proces. Mag.* **29** 152–0

Jiang N, Rehbaum H, Vujaklija I, Graimann B and Farina D 2013 Intuitive, online, simultaneous, and proportional myoelectric control over two degrees-of-freedom in upper limb amputees *IEEE Trans. Neural Syst. Rehabil. Eng.* **22** 501–10

Jiang S, GAO Q, LIU H and Shull P B 2020 A novel, co-located EMG-FMG-sensing wearable armband for hand gesture recognition *Sens. Actuators* A **301** 111738

Jiang S, Li L, Xu H, Xu J, Gu G and Shull P B 2019 Stretchable e-skin patch for gesture recognition on the back of the hand *IEEE Trans. Ind. Electron.* **67** 647–57

Johannes M S, Bigelow J D, Burck J M, Harshbarger S D, Kozlowski M V and Van Doren T 2011 An overview of the developmental process for the modular prosthetic limb *Johns Hopkins APL Tech. Dig.* **30** 207–16

Johansen D, Cipriani C, Popović D B and Struijk L N 2016 Control of a robotic hand using a tongue control system—a prosthesis application *IEEE Trans. Biomed. Eng.* **63** 1368–76

Johansen D, Popović D B, Dosen S and Struijk L N A 2021 Hybrid tongue-myoelectric control improves functional use of a robotic hand prosthesis *IEEE Trans. Biomed. Eng.* **68** 2011–20

Jones H, Dupan S, Dyson M, Krasoulis A, Kenney L P J, Donovan-Hall M, Memarzadeh K, Day S, Coutinho M and Nazarpour K 2021 Co-creation and user perspectives for upper limb prosthetics *Front. Neurorobot.* **15**

Jorgovanovic N, Dosen S, Djozic D J, Krajoski G and Farina D 2014 Virtual grasping: closed-loop force control using electrotactile feedback *Comput. Math. Methods Med.* **2014** 1–13

Kalampratsidou V and Torres E B 2017 Body-brain-avatar interface: a tool to study sensory-motor integration and neuroplasticity *4th Int. Symp. on Movement and Computing, MOCO*

Kayhan O, Nennioglu A K and Samur E 2018 A skin stretch tactor for sensory substitution of wrist proprioception *2018 IEEE Haptics Symp. (HAPTICS)* (IEEE) pp 26–31

Kenney L P, Lisitsa I, Bowker P, Heath G H and Howard D 1999 Dimensional change in muscle as a control signal for powered upper limb prostheses: a pilot study *Med. Eng. Phys.* **21** 589–97

Kerver N, Van Twillert S, Maas B and Van Der Sluis C K 2020 User-relevant factors determining prosthesis choice in persons with major unilateral upper limb defects: a meta-synthesis of qualitative literature and focus group results *PLoS One* **15** e0234342

Kilteni K, Groten R and Slater M 2012 The sense of embodiment in virtual reality *Presence: Teleoperators Virtual Environ.* **21** 373–87

Krasoulis A, Kyranou I, Erden M S, Nazarpour K and Vijayakumar S 2017 Improved prosthetic hand control with concurrent use of myoelectric and inertial measurements *J. Neuroeng. Rehabil.* **14** 1–14

Krasoulis A, Vijayakumar S and Nazarpour K 2019a Effect of user practice on prosthetic finger control with an intuitive myoelectric decoder *Front. Neurosci.* **13** 891

Krasoulis A, Vijayakumar S and Nazarpour K 2019b Multi-grip classification-based prosthesis control with two EMG-IMU sensors *IEEE Trans. Neural Syst. Rehabil. Eng.* **28** 508–18

Krausz N E, Lamotte D, Batzianoulis I, Hargrove L J, Micera S and Billard A 2020 Intent prediction based on biomechanical coordination of EMG and vision-filtered gaze for end-point control of an arm prosthesis *IEEE Trans. Neural Syst. Rehabil. Eng.* **28** 1471–80

Kuiken T A, Barlow A K, Hargrove L and Dumanian G A 2017 Targeted muscle reinnervation for the upper and lower extremity *Tech. Orthop.* **32** 109

Kuiken T A, LI G, Lock B A, Lipschutz R D, Miller L A, Stubblefield K A and Englehart K B 2009 Targeted muscle reinnervation for real-time myoelectric control of multifunction artificial arms *Jama* **301** 619–28

Kyranou I, Vijayakumar S and Erden M S 2018 Causes of performance degradation in non-invasive electromyographic pattern recognition in upper limb prostheses *Front. Neurorobot.* **12**

Laffranchi M *et al* 2020 The Hannes hand prosthesis replicates the key biological properties of the human hand *Sci. Robot.* **5**

Lah U, Lewis J R and Šumak B 2020 Perceived usability and the modified technology acceptance model *Int. J. Hum.–Comput. Interact.* **36** 1216–30

Lamounier E, Lopes K, Cardoso A, Andrade A and Soares A 2010 On the use of virtual and augmented reality for upper limb prostheses training and simulation *2010 Annual Int. Conf. IEEE Engineering in Medicine and Biology* (IEEE) pp 2451–4

Lenzi T, Lipsey J and Sensinger J W 2016 The RIC arm—a small anthropomorphic transhumeral prosthesis *IEEE/ASME Trans. Mechatronics* **21** 2660–71

LI W, Shi P and Yu H 2021 Gesture recognition using surface electromyography and deep learning for prostheses hand: state-of-the-art, challenges, and future *Front. Neurosci.* **15** 621885

LI Y, Chen J and Yang Y 2019 A method for suppressing electrical stimulation artifacts from electromyography *Int. J. Neural Syst.* **29** 1850054

Lin C H, Erickson T W, Fishel J A, Wettels N and Loeb G E 2009 Signal processing and fabrication of a biomimetic tactile sensor array with thermal, force and microvibration modalities *2009 IEEE Int. Conf. on Robotics and Biomimetics (ROBIO)* (IEEE) pp 129–34

Liu G, Wang L and Wang J 2021a A novel energy-motion model for continuous sEMG decoding: from muscle energy to motor pattern *J. Neural Eng.* **18** 016019

LIU M, Batista A, Bensmaia S and Weber D J 2021b Information about contact force and surface texture is mixed in the firing rates of cutaneous afferent neurons *J. Neurophysiol.* **125** 496–508

Longfellow H W 2014 Upper-limb prostheses, acceptance, and acceptance factors *Improving the Design of Upper-Limb Robotic Prostheses Using Biomechanical Properties and Sensorimotor Control Principles* 5

Longo L 2018 Experienced mental workload, perception of usability, their interaction and impact on task performance *PLoS One* **13** e0199661

Lucarotti C, Oddo C M, Vitiello N and Carrozza M C 2013 Synthetic and bio-artificial tactile sensing: a review *Sensors* **13** 1435–66

Luchetti M, Cutti A G, Verni G, Sacchetti R and Rossi N 2015 Impact of Michelangelo prosthetic hand: findings from a crossover longitudinal study *J. Rehabil. Res. Dev.* **52** 605–18

Lundborg G, Rosén B and Lindberg S 1999 Hearing as substitution for sensation: a new principle for artificial sensibility *J. Hand Surg. Am.* **24** 219–24

Maimon Mor R O and Makin T R 2020 Is an artificial limb embodied as a hand? Brain decoding in prosthetic limb users *PLoS Biol.* **18** e3000729

Mainardi E and Davalli A 2007 Controlling a prosthetic arm with a throat microphone *2007 29th Annual Int. Conf. IEEE Engineering in Medicine and Biology Society* (IEEE) pp 3035–9

Maiolino P, Maggiali M, Cannata G, Metta G and Natale L 2013 A flexible and robust large scale capacitive tactile system for robots *IEEE Sens. J.* **13** 3910–17

Makin T R, De Vignemont F and Faisal A A 2017 Neurocognitive barriers to the embodiment of technology *Nat. Biomed. Eng.* **1** 1–3

Mamidanna P, Dideriksen J L and Dosen S 2021 The impact of objective functions on control policies in closed-loop control of grasping force with a myoelectric prosthesis *J. Neural Eng.* **18** 056036

Marasco P D, Hebert J S, Sensinger J W, Beckler D T, Thumser Z C, Shehata A W, Williams H E and Wilson K R 2021 Neurorobotic fusion of prosthetic touch, kinesthesia, and movement in bionic upper limbs promotes intrinsic brain behaviors *Sci. Robot.* **6** eabf3368

Marasco P D, Kim K, Colgate J E, Peshkin M A and Kuiken T A 2011 Robotic touch shifts perception of embodiment to a prosthesis in targeted reinnervation amputees *Brain* **134** 747–58

Marinelli A *et al* 2020 Performance evaluation of pattern recognition algorithms for upper limb prosthetic applications *8th IEEE RAS/EMBS Int. Conf. for Biomedical Robotics and Biomechatronics (Biorob)* (IEEE)

Marinelli A, Boccardo N, Semprini M, Succi A, Canepa M, Stedman S, Lombardi L, Bellingegni A D, Chiappalone M and Gruppioni E 2021 Miniature EMG sensors for prosthetic applications *2021 10th Int. IEEE/EMBS Conf. on Neural Engineering (NER)* (IEEE) pp 1022–5

Markovic M, Dosen S, Cipriani C, Popovic D and Farina D 2014 Stereovision and augmented reality for closed-loop control of grasping in hand prostheses *J. Neural Eng.* **11** 046001

Markovic M, Karnal H, Graimann B, Farina D and Dosen S 2017 GLIMPSE: Google glass interface for sensory feedback in myoelectric hand prostheses *J. Neural Eng.* **14** 036007

Markovic M, Schweisfurth M A, Engels L F, Farina D and Dosen S 2018 Myocontrol is closed-loop control: incidental feedback is sufficient for scaling the prosthesis force in routine grasping *J. Neuroeng. Rehabil.* **15** 1–11

Markovic M, Varel M, Schweisfurth M A, Schilling A F and Dosen S 2019 Closed-loop multi-amplitude control for robust and dexterous performance of myoelectric prosthesis *IEEE Trans. Neural Syst. Rehabil. Eng.* **28** 498–507

Massari L, Oddo C M, Sinibaldi E, Detry R, Bowkett J and Carpenter K C 2019 Tactile sensing and control of robotic manipulator integrating fiber Bragg grating strain-sensor *Front. Neurorobot.* **13** 8

Mazzoni A, Oddo C M, Valle G, Camboni D, Strauss I, Barbaro M, Barabino G, Puddu R, Carboni C and Bisoni L 2020 Morphological neural computation restores discrimination of naturalistic textures in trans-radial amputees *Sci. Rep.* **10** 1–14

Mcdermott E J, Metsomaa J, Belardinelli P, Grosse-wentrup M, Ziemann U and Zrenner C 2021 Predicting motor behavior: an EEG signal processing pipeline to detect relevant brain states with potential therapeutic relevance for VR-based neurorehabilitation *Virtual Real* (https://doi.org/10.1007/s10055-021-00538-x)

Mcdonnell P M, Scott R N, Dickison J, Theriault R A and Wood B 1989 Do artificial limbs become part of the user? New evidence *J. Rehabil. Res. Dev.* **26** 17–24

Mcfarland D J, Sarnacki W A and Wolpaw J R 2010 Electroencephalographic (EEG) control of three-dimensional movement *J. Neural Eng.* **7** 036007

Meagher C, Franco E, Turk R, Wilson S, Steadman N, Mcnicholas L, Vaidyanathan R, Burridge J and Stokes M 2020 New advances in mechanomyography sensor technology and signal processing: validity and intrarater reliability of recordings from muscle *J. Rehabil. Assist. Technol. Eng.* **7** 2055668320916116

Meattini R, Biagiotti L, Palli G, De Gregorio D and Melchiorri C 2019 A control architecture for grasp strength regulation in myocontrolled robotic hands using vibrotactile feedback: preliminary results *2019 IEEE 16th Int. Conf. on Rehabilitation Robotics (ICORR)* (IEEE) pp 1272–7

Melcer E F, Astolfi M T, Remaley M, Berenzweig A and Giurgica-tiron T 2018 CTRL-labs: hand activity estimation and real-time control from neuromuscular signals *Extended Abstracts of the 2018 CHI Conf. on Human Factors in Computing Systems* pp 1–4

Merad M, De Montalivet É, Touillet A, Martinet N, Roby-brami A and Jarrassé N 2018 Can we achieve intuitive prosthetic elbow control based on healthy upper limb motor strategies? *Front. Neurorobot.* **12** 1

Mereu F, Leone F, Gentile C, Cordella F, Gruppioni E and Zollo L 2021 Control strategies and performance assessment of upper-limb TMR prostheses: a review *Sensors* **21** 1953

Merletti R, Aventaggiato M, Botter A, Holobar A, Marateb H and Vieira T M 2010 Advances in surface EMG: recent progress in detection and processing techniques *Crit. Rev. Biomed. Eng.* **38**

Merletti R and Farina D 2009 Analysis of intramuscular electromyogram signals *Phil. Trans. R. Soc.* A **367** 357–68

Merletti R and Farina D 2016 Surface electromyography for man-machine interfacing in rehabilitation technologies *Surface Electromyography: Physiology, Engineering, and Applications* (IEEE) pp 540–60

Merletti R, Holobar A and Farina D 2008 Analysis of motor units with high-density surface electromyography *J. Electromyogr. Kinesiol.* **18** 879–90

Merletti R and Muceli S 2019 Tutorial. Surface EMG detection in space and time: best practices *J. Electromyogr. Kinesiol.* **49** 102363

Milgram P and Kishino F 1994 A taxonomy of mixed reality visual displays *IEICE Trans. Inf. Syst.* **77** 1321–9

Millstein S, Heger H and Hunter G 1986 Prosthetic use in adult upper limb amputees: a comparison of the body powered and electrically powered prostheses *Prosthet. Orthot. Int.* **10** 27–34

Montazerin M, Zabihi S, Rahimian E, Mohammadi A and Naderkhani F 2022 ViT-HGR: vision transformer-based hand gesture recognition from high density surface EMG signals (arXiv:2201.10060)

Moon I, Lee M, Chu J and Mun M 2005 Wearable EMG-based HCI for electric-powered wheelchair users with motor disabilities *Proc. 2005 IEEE Int. Conf. on Robotics and Automation* (IEEE) pp 2649–54

Moore C H, Corbin S F, Mayr R, Shockley K, Silva P L and Lorenz T 2021 Grasping embodiment: haptic feedback for artificial limbs *Front. Neurorobot.* **15** 66

Mouchoux J, Carisi S, Dosen S, Farina D, Schilling A F and Markovic M 2021 Artificial perception and semiautonomous control in myoelectric hand prostheses increases performance and decreases effort *IEEE Trans. Robot.* **37** 1298–312

Musk E 2021 *Neuralink* (available at: https://neuralink.com/)

Nakagawa-silva A, Sunkesula S P R, Prach A, Cabibihan J-J, Thakor N V and Soares A B Slip suppression in prosthetic hands using a reflective optical sensor and MPI controller 2018 *2018 IEEE Biomedical Circuits and Systems Conf. (Biocas)* (IEEE) pp 1–4

Navaraj W T, Nassar H and Dahiya R 2019 Prosthetic hand with biomimetic tactile sensing and force feedback *2019 IEEE Int. Symp. on Circuits and Systems (ISCAS)* (IEEE) pp 1–4

Nelson E C, Sools A M, Vollenbroek-hutten M M, Verhagen T and Noordzij M L 2020 Embodiment of wearable technology: qualitative longitudinal study *JMIR mHealth uHealth* **8** e16973

Nguyen A T, Drealan M W, Luu D K, Jiang M, Xu J, Cheng J, Zhao Q, Keefer E W and Yang Z 2021. A portable, self-contained neuroprosthetic hand with deep learning-based finger control *J. Neural Eng.* **18** 056051

Nguyen A T, Xu J, Jiang M, Luu D K, Wu T, Tam W-K, Zhao W, Drealan M W, Overstreet C K and Zhao Q 2020 A bioelectric neural interface towards intuitive prosthetic control for amputees *J. Neural Eng.* **17** 066001

Nicolelis M A and Chapin J K 2002 Controlling robots with the mind *Sci. Am.* **287** 46–53

Nissler C, Mouriki N and Castellini C 2016 Optical myography: detecting finger movements by looking at the forearm *Front. Neurorobot.* **10** 3

Nissler C, Nowak M, Connan M, Büttner S, Vogel J, Kossyk I, Márton Z-C and Castellini C 2019 VITA—an everyday virtual reality setup for prosthetics and upper-limb rehabilitation *J. Neural Eng.* **16** 026039

Nowak M, Castellini C and Massironi C 2018 Applying radical constructivism to machine learning: a pilot study in assistive robotics *Constr. Found.* **13** 250–62

Nowak M, Eiband T, Ramírez E R and Castellini C 2020 Action interference in simultaneous and proportional myocontrol: comparing force-and electromyography *J. Neural Eng.* **17** 026011

O'sullivan L, Power V, De Eyto A and Ortiz J 2017 User centered design and usability of bionic devices *Converging Clinical and Engineering Research on Neurorehabilitation II* (Berlin: Springer)

Oddo C M, Raspopovic S, Artoni F, Mazzoni A, Spigler G, Petrini F, Giambattistelli F, Vecchio F, Miraglia F and Zollo L 2016 Intraneural stimulation elicits discrimination of textural features by artificial fingertip in intact and amputee humans *elife* **5** e09148

Olsson A E, Sager P, Andersson E, Björkman A, Malešević N and Antfolk C 2019 Extraction of multi-labelled movement information from the raw HD-sEMG image with time-domain depth *Sci. Rep.* **9** 1–10

Orizio C, Solomonow M, Diemont B and Gobbo M 2008 Muscle-joint unit transfer function derived from torque and surface mechanomyogram in humans using different stimulation protocols *J. Neurosci. Methods* **173** 59–66

Ortiz-catalan M, Mastinu E, Sassu P, Aszmann O and Brånemark R 2020 Self-contained neuromusculoskeletal arm prostheses *New Engl. J. Med.* **382** 1732–8

Osborn L E, Dragomir A, Betthauser J L, Hunt C L, Nguyen H H, Kaliki R R and Thakor N V 2018 Prosthesis with neuromorphic multilayered e-dermis perceives touch and pain *Sci. Robot.* **3**

Ossur 2020a i-limb apps (available at: www.ossur.com/it-it/ossur-per-i-professionisti/apps-ossur/i-limb-mobile-apps)

Ossur 2020b *i-limb® ultra titanium* (available at: www.ossur.com/it-it/protesi/arto-superiore/i-limb-ultra-titanium)

Ottobock 2019 Myo plus pattern recognition (available at: www.ottobock.it/protesizzazioni/arto-superiore/panoramica-delle-soluzioni/myo-plus-pattern-recognition/)

Ottobock 2020a Bebionic hand (available at: www.ottobock.it/protesizzazioni/prodotti-dalla-a-alla-z/mano-bebionic/)

Ottobock 2020b Michelangelo hand (available at: www.ottobock.it/soluzioni-protesiche/arto-superiore/panoramica-delle-soluzioni/sistema-axon-bus-con-mano-michelangelo/)

Ottobock 2020c MyoHand VariPlus speed (available at: https://shop.ottobock.us/Prosthetics/Upper-Limb-Prosthetics/Myo-Hands-and-Components/Myo-Terminal-Devices/MyoHand-VariPlus-Speed/p/8E38∼59)

Ottobock 2021 SensorHand speed (available at: https://shop.ottobock.us/Prosthetics/Upper-Limb-Prosthetics/Myo-Hands-and-Components/Myo-Terminal-Devices/SensorHand-Speed/p/8E39∼58)

Ottobock 2022a DynamicArm (available at: https://shop.ottobock.us/Prosthetics/Upper-Limb-Prosthetics/Myoelectric-Elbows/DynamicArm-Elbow/c/2125)

Ottobock 2022b DynamicArm plus (available at: https://shop.ottobock.us/Prosthetics/Upper-Limb-Prosthetics/Myoelectric-Elbows/DynamicArm-Plus-Elbow/c/2126)

Ottobock 2022c ErgoArm (available at: https://shop.ottobock.us/Prosthetics/Upper-Limb-Prosthetics/Myoelectric-Elbows/ErgoArm-Elbow/c/2127)

Padilha Leitzke J and Zangl H 2020 A review on electrical impedance tomography spectroscopy *Sensors* **20** 5160

Page D M, George J A, Kluger D T, Duncan C, Wendelken S, Davis T, Hutchinson D T and Clark G A 2018 Motor control and sensory feedback enhance prosthesis embodiment and reduce phantom pain after long-term hand amputation *Front. Hum. Neurosci.* **12** 352

Paleari M, Luciani R and Ariano P 2017 Towards NIRS-based hand movement recognition *2017 Int. Conf. on Rehabilitation Robotics (ICORR)* (IEEE) pp 1506–11

Paljic A 2017 Ecological validity of virtual reality: three use cases *Int. Conf. on Image Analysis and Processing* (Springer) pp 301–10

Pan L, Crouch D L and Huang H 2018 Myoelectric control based on a generic musculoskeletal model: toward a multi-user neural-machine interface *IEEE Trans. Neural Syst. Rehabil. Eng.* **26** 1435–42

PAN L, Zhang D, Jiang N, Sheng X and ZHU X 2015 Improving robustness against electrode shift of high density EMG for myoelectric control through common spatial patterns *J. Neuroeng. Rehabil.* **12** 1–16

Parajuli N, Sreenivasan N, Bifulco P, Cesarelli M, Savino S, Niola V, Esposito D, Hamilton T J, Naik G R and Gunawardana U 2019 Real-time EMG based pattern recognition control for hand prostheses: a review on existing methods, challenges and future implementation *Sensors* **19** 4596

Pelayo S, Marcilly R and Bellandi T 2021 Human factors engineering for medical devices: European regulation and current issues *Int. J. Quality Health Care* **33** 31–36

Phelan I, Arden M, Matsangidou M, Carrion-plaza A and Lindley S 2021 Designing a virtual reality myoelectric prosthesis training system for amputees *Extended Abstracts of the 2021 CHI Conf. on Human Factors in Computing Systems* pp 1–7

Phillips S L and Craelius W 2005 Residual kinetic imaging: a versatile interface for prosthetic control *Robotica* **23** 277–82

Phinyomark A and Scheme E 2018 EMG pattern recognition in the era of big data and deep learning *Big Data Cogn. Comput.* **2** 21

Piazza C, Rossi M, Catalano M G, Bicchi A and Hargrove L J 2020 Evaluation of a simultaneous myoelectric control strategy for a multi-DoF transradial prosthesis *IEEE Trans. Neural Syst. Rehabil. Eng.* **28** 2286–95

Prakash A, Sahi A K, Sharma N and Sharma S 2020 Force myography controlled multifunctional hand prosthesis for upper-limb amputees *Biomed. Signal Process. Control* **62** 102122

PSYONIC 2022 Ability hand (available at: www.psyonic.io/ability-hand)

Pyasik M, Tieri G and Pia L 2020 Visual appearance of the virtual hand affects embodiment in the virtual hand illusion *Sci. Rep.* **10** 1–11

Radmand A, Scheme E and Englehart K 2016 High-density force myography: a possible alternative for upper-limb prosthetic control *J. Rehabil. Res. Dev.* **53** 443–56

Rapetti L, Tirupachuri Y, Darvish K, Dafarra S, Nava G, Latella C and Pucci D 2020 Model-based real-time motion tracking using dynamical inverse kinematics *Algorithms* **13** 266

Raspopovic S *et al* 2014 Restoring natural sensory feedback in real-time bidirectional hand prostheses *Sci. Transl. Med.* **6** 222ra19

Raspopovic S, Valle G and Petrini F M 2021a Sensory feedback for limb prostheses in amputees *Nat. Mater.* 1–15

Raspopovic S, Valle G and Petrini F M 2021b Sensory feedback for limb prostheses in amputees *Nat. Mater.* **20** 925–39

Resnik L 2011 Development and testing of new upper-limb prosthetic devices: research designs for usability testing *J. Rehabil. Res. Dev.* 48

Resnik L, Etter K, Klinger S L and Kambe C 2011 Using virtual reality environment to facilitate training with advanced upper-limb prosthesis *J. Rehabil. Res. Dev.* **48**

Resnik L, Huang H H, Winslow A, Crouch D L, Zhang F and Wolk N 2018 Evaluation of EMG pattern recognition for upper limb prosthesis control: a case study in comparison with direct myoelectric control *J. Neuroeng. Rehabil.* **15** 1–13

Resnik L, Klinger S L, Etter K and Fantini C 2014 Controlling a multi-degree of freedom upper limb prosthesis using foot controls: user experience *Disabil. Rehabil. Assist. Technol.* **9** 318–29

Ribeiro J, Mota F, Cavalcante T, Nogueira I, Gondim V, Albuquerque V and Alexandria A 2019 Analysis of man-machine interfaces in upper-limb prosthesis: a review *Robotics* **8** 16

Riener R 2016 The cybathlon promotes the development of assistive technology for people with physical disabilities *J. Neuroeng. Rehabil.* **13** 1–4

Robot S 2022 Shadow dexterous hand series (available at: www.shadowrobot.com/dexterous-hand-series/)

Roche A D, Lakey B, Mendez I, Vujaklija I, Farina D and Aszmann O C 2019 Clinical perspectives in upper limb prostheses: an update *Curr. Surg. Rep.* **7**

Rodgers M M, Alon G, PAI V M and Conroy R S 2019 Wearable technologies for active living and rehabilitation: current research challenges and future opportunities *J. Rehabil. Assist. Technol. Eng.* **6** 2055668319839607

Romano D, Caffa E, Hernandez-arieta A, Brugger P and Maravita A 2015 The robot hand illusion: inducing proprioceptive drift through visuo-motor congruency *Neuropsychologia* **70** 414–20

Romano D, Maravita A and Perugini M 2021 Psychometric properties of the embodiment scale for the rubber hand illusion and its relation with individual differences *Sci. Rep.* **11** 1–16

Rubin D I 2019 Needle electromyography: basic concepts *Handbook of Clinical Neurology* vol 160 pp 243–56

Salminger S, Stino H, Pichler L H, Gstoettner C, Sturma A, Mayer J A, Szivak M and Aszmann O C 2020 Current rates of prosthetic usage in upper-limb amputees–have innovations had an impact on device acceptance? *Disabil. Rehabil.* 1–12

Sani H N and Meek S G 2011 Characterizing the performance of an optical slip sensor for grip control in a prosthesis *2011 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems* (IEEE) pp 1927–32

Sartori M, Durandau G, Došen S and Farina D 2018 Robust simultaneous myoelectric control of multiple degrees of freedom in wrist-hand prostheses by real-time neuromusculoskeletal modeling *J. Neural Eng.* **15** 066026

Scalise L, Casaccia S, Marchionni P, Ercoli I and EP T 2013 Laser Doppler myography (LDMi): a novel non-contact measurement method for the muscle activity *Laser Ther.* **22** 261–8

Scheme E and Englehart K 2011 Electromyogram pattern recognition for control of powered upper-limb prostheses: state of the art and challenges for clinical use *J. Rehabil. Res. Dev.* **48**

Schmidl H 1965 The INAIL myoelectric B/E prosthesis *Orthot. Prosthet.* **19** 298–303

Schmitz A, Maggiali M, Randazzo M, Natale L and Metta G 2008 A prototype fingertip with high spatial resolution pressure sensing for the robot iCub *Humanoids 2008—8th IEEE-RAS Int. Conf. on Humanoid Robots* (IEEE) pp 423–8

Schneider D M, Ripplinger C M, Gilbertson K, Katti R and Schroeder M J 2003 Optically-based control of a prosthetic device *Summer Bioengineering Conf.* (*Sonesta Beach Resort in Key Biscayne, Florida, 25–29 June*)

Schweisfurth M A, Markovic M, Dosen S, Teich F, Graimann B and Farina D 2016 Electrotactile EMG feedback improves the control of prosthesis grasping force *J. Neural Eng.* **13** 056010

Scott R and Parker P 1988 Myoelectric prostheses: state of the art *J. Med. Eng. Technol.* **12** 143–51

Sensinger J W and Dosen S 2020 A review of sensory feedback in upper-limb prostheses from the perspective of human motor control *Front. Neurosci.* **14**

Shahzaib M and Shakil S 2018 Hand electromyography circuit and signals classification using artificial neural network *2018 14th Int. Conf. on Emerging Technologies (ICET)* (IEEE) pp 1–6

Shannon G 1979 A myoelectrically-controlled prosthesis with sensory feedback *Med. Biol. Eng. Comput.* **17** 73–80

Sharma A, Hunt C L, Maheshwari A, Osborn L, Lévay G, Kaliki R R, Soares A B and Thakor N 2018 A mixed-reality training environment for upper limb prosthesis control *2018 IEEE Biomedical Circuits and Systems Conf. (Biocas)* (IEEE) pp 1–4

Shaw H, Ellis D A and Ziegler F V 2018 The technology integration model (TIM). Predicting the continued use of technology *Comput. Hum. Behav.* **83** 204–14

Shull P B, Jiang S, ZHU Y and ZHU X 2019 Hand gesture recognition and finger angle estimation via wrist-worn modified barometric pressure sensing *IEEE Trans. Neural Syst. Rehabil. Eng.* **27** 724–32

Siciliano B, Khatib O and Kröger T 2008 *Springer Handbook of Robotics* (Berlin: Springer)

Siciliano B, Sciavicco L, Villani L and Oriolo G 2010 *Robotics: Modelling, Planning and Control* (Springer Science & Business Media)

Simon A M, Hargrove L J, Lock B A and Kuiken T A 2011 The target achievement control test: evaluating real-time myoelectric pattern recognition control of a multifunctional upper-limb prosthesis *J. Rehabil. Res. Dev.* **48** 619

Smail L C, Neal C, Wilkins C and Packham T L 2020 Comfort and function remain key factors in upper limb prosthetic abandonment: findings of a scoping review *Disabil. Rehabil. Assist. Technol.* 1–10

Smith L H, Kuiken T A and Hargrove L J 2014 Real-time simultaneous and proportional myoelectric control using intramuscular EMG *J. Neural Eng.* **11** 066013

Speicher M, Hall B D and Nebeling M 2019 What is mixed reality? *Proc. 2019 CHI Conf. on Human Factors in Computing Systems* pp 1–15

Steeper 2022 Espire elbow (available at: www.steepergroup.com/prosthetics/upper-limb-prosthetics/espire%E2%84%A2-elbow/)

Štrbac M, Belić M, Isaković M, Kojić V, Bijelić G, Popović I, Radotić M, Došen S, Marković M and Farina D 2016 Integrated and flexible multichannel interface for electrotactile stimulation *J. Neural Eng.* **13** 046014

Sun T, Hu Q, Gulati P and Atashzar S F 2021a Temporal dilation of deep LSTM for agile decoding of sEMG: application in prediction of upper-limb motor intention in NeuroRobotics *IEEE Robot. Autom. Lett.* **6** 6212–9

Sun Y *et al* 2021b A comparison between virtual reality and augmented reality on upper-limb prosthesis control *2021 Int. Symp. on Electrical, Electronics and Information Engineering* pp 521–8

Svensson P, Wijk U, Björkman A and Antfolk C 2017 A review of invasive and non-invasive sensory feedback in upper limb prostheses *Expert Rev. Med. Devices* **14** 439–47

Syed U, Kausar Z and Yousaf N 2020 Control of a prosthetic arm using fNIRS, a neural-machine interface *Data Acquisition-Recent Advances and Applications in Biomedical Engineering*

Systems V 2020 Vincent hand (available at: www.vincentsystems.de/)

Tabor A, Hill W, Bateman S and Scheme E 2017 Quantifying muscle control in myoelectric training games *Proc. Myoelectr Control Symp. (MEC)*

TAN D W, Schiefer M A, Keith M W, Anderson J R, Tyler J and Tyler D J 2014 A neural interface provides long-term stable natural touch perception *Sci. Transl. Med.* **6** 257ra138

Tanaka M, Okuyama T and Saito K 2011 Study on evaluation of muscle conditions using a mechanomyogram sensor *IEEE Int. Conf. Syst. Man Cybern.* 741–5

Tarantino S, Clemente F, Barone D, Controzzi M and Cipriani C 2017 The myokinetic control interface: tracking implanted magnets as a means for prosthetic control *Sci. Rep.* **7** 1–11

Taska 2022. Taska Hand Gen 2 (available at: www.taskaprosthetics.com/it/prodotti/taska-hand/)

Taunyazov T, Song L S, Lim E, See H H, Lee D, Tee B C and Soh H 2021 Extended tactile perception: vibration sensing through tools and grasped objects. *2021 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems (IROS)* (IEEE) pp 1755–62

Tchimino J, Marković M, Dideriksen J and Dosen S 2021 The effect of calibration parameters on the control of a myoelectric hand prosthesis using EMG feedback *J. Neural Eng.* **18** 046091

Tee B C, Wang C, Allen R and Bao Z 2012 An electrically and mechanically self-healing composite with pressure-and flexion-sensitive properties for electronic skin applications *Nat. Nanotechnol.* **7** 825–32

Thakor N V, Fifer M S, Hotson G, Benz H L, Newman G I, Milsap G W and Crone N E 2014 Neuroprosthetic limb control with electrocorticography: approaches and challenges *2014 36th Annual Int. Conf. IEEE Engineering in Medicine and Biology Society* (IEEE) pp 5212–5

Toet A, Kuling I A, Krom B N and Van Erp J B 2020 Toward enhanced teleoperation through embodiment *Front. Robot. AI* **7** 14

Tomo T P *et al* 2018 A new silicone structure for uSkin—A soft, distributed, digital 3-axis skin sensor and its integration on the humanoid robot iCub *IEEE Robot. Autom. Lett.* **3** 2584–91

Trent L, Intintoli M, Prigge P, Bollinger C, Walters L S, Conyers D, Miguelez J and Ryan T 2019 A narrative review: current upper limb prosthetic options and design *Disabil. Rehabil. Assist. Technol.* **15** 604–13

Tropea P, Mazzoni A, Micera S and Corbo M 2017 Giuliano Vanghetti and the innovation of "cineplastic operations" *Neurology* **89** 1627–32

Truong H, Zhang S, Muncuk U, Nguyen P, Bui N, Nguyen A, Lv Q, Chowdhury K, Dinh T and Vu T 2018 Capband: battery-free successive capacitance sensing wristband for hand gesture recognition *Proc. 16th ACM Conf. on Embedded Networked Sensor Systems* pp 54–67

Tsakiris M and Haggard P 2005 The rubber hand illusion revisited: visuotactile integration and self-attribution *J. Exp. Psychol. Hum. Percept. Perform.* **31** 80

Uellendahl J 2017 Myoelectric versus body-powered upper-limb prostheses: a clinical perspective *J. Prosthet. Orthot.* **29** P25–9

Urbanchek M G, Baghmanli Z, Moon J D, Sugg K B, Langhals N B and Cederna P S 2012 Quantification of regenerative peripheral nerve interface signal transmission *Plast. Reconstr. Surg.* **130** 55–6

Valle G, Mazzoni A, Iberite F, D'anna E, Strauss I, Granata G, Controzzi M, Clemente F, Rognini G and Cipriani C 2018 Biomimetic intraneural sensory feedback enhances sensation naturalness, tactile sensitivity, and manual dexterity in a bidirectional prosthesis *Neuron* **100** e7

Varol H A, Dalley S A, Wiste T E and Goldfarb M 2014 Biomimicry and the design of multigrasp transradial prostheses *The Human Hand as an Inspiration for Robot Hand Development* (Berlin: Springer)

Vasan G and Pilarski P M 2017 Learning from demonstration: teaching a myoelectric prosthesis with an intact limb via reinforcement learning *2017 Int. Conf. on Rehabilitation Robotics (ICORR)* (IEEE) pp 1457–64

Vazhapilli Sureshbabu A, Metta G and Parmiggiani A 2019 A systematic approach to evaluating and benchmarking robotic hands—the FFP index *Robotics* **8** 7

VU P P, Chestek C A, Nason S R, Kung T A, Kemp S W and Cederna P S 2020a The future of upper extremity rehabilitation robotics: research and practice *Muscle Nerve* **61** 708–18

Vu P P, Vaskov A K, Irwin Z T, Henning P T, Lueders D R, Laidlaw A T, Davis A J, NU C S, Gates D H and Gillespie R B 2020b A regenerative peripheral nerve interface allows real-time control of an artificial hand in upper limb amputees *Sci. Transl. Med.* **12**

Wang W, Collinger J L, Degenhart A D, Tyler-kabara E C, Schwartz A B, Moran D W, Weber D J, Wodlinger B, Vinjamuri R K and Ashmore R C 2013 An electrocorticographic brain interface in an individual with tetraplegia *PLoS One* **8** e55344

Weiner P, Starke J, Hundhausen F, Beil J and Asfour T 2018 The KIT prosthetic hand: design and control *2018 IEEE/RSJ Int. Conf. on Intelligent Robots and Systems (IROS)* (IEEE) pp 3328–34

Weir R F, Troyk P R, Demichele G A, Kerns D A, Schorsch J F and Maas H 2008 Implantable myoelectric sensors (IMESs) for intramuscular electromyogram recording *IEEE Trans. Biomed. Eng.* **56** 159–171

Wilke M A, Niethammer C, Meyer B, Farina D and Dosen S 2019 Psychometric characterization of incidental feedback sources during grasping with a hand prosthesis *J. Neuroeng. Rehabil.* **16** 1–13

Wilson S and Vaidyanathan R 2017 Upper-limb prosthetic control using wearable multichannel mechanomyography *2017 Int. Conf. on Rehabilitation Robotics (ICORR)* (IEEE) pp 1293–8

Winters J M 1990 Hill-based muscle models: a systems engineering perspective *Multiple Muscle Systems*

Wodlinger B, Downey J, Tyler-kabara E, Schwartz A, Boninger M and Collinger J 2014 Ten-dimensional anthropomorphic arm control in a human−machine interface: difficulties, solutions, and limitations *J. Neural Eng.* **12** 016011

Woodward R B, Shefelbine S J and Vaidyanathan R 2017 Pervasive monitoring of motion and muscle activation: inertial and mechanomyography fusion *IEEE/ASME Trans. Mechatronics* **22** 2022–33

Wu R, Li M, Yao Z, Liu W, Si J and Huang H 2022 Reinforcement learning impedance control of a robotic prosthesis to coordinate with human intact knee motion *IEEE Robot. Autom. Lett.* **7** 7014–20

Wu Y T, Fujiwara E and Suzuki C K 2019 Evaluation of optical myography sensor as predictor of hand postures *IEEE Sens. J.* **19** 5299–306

Wu Y, Jiang D, Liu X, Bayford R and Demosthenous A 2018 A human–machine interface using electrical impedance tomography for hand prosthesis control *IEEE Trans. Biomed. Circuits Syst.* **12** 1322–33

Xia W, Zhou Y, Yang X, He K and Liu H 2019 Toward portable hybrid surface electromyography/A-mode ultrasound sensing for human–machine interface *IEEE Sens. J.* **19** 5219–28

Xiao Z G and Menon C 2019 A review of force myography research and development *Sensors* **19** 4557

Xu H, Zhang D, Huegel J C, Xu W and Zhu X 2015 Effects of different tactile feedback on myoelectric closed-loop control for grasping based on electrotactile stimulation *IEEE Trans. Neural Syst. Rehabil. Eng.* **24** 827–36

Yang D and Liu H 2021 Human-machine shared control: new avenue to dexterous prosthetic hand manipulation *Sci. China Technol. Sci.* **64** 767–73

Yang J, Singh H, Hines E L, Schlaghecken F, Iliescu D D, Leeson M S and Stocks N G 2012 Channel selection and classification of electroencephalogram signals: an artificial neural network and genetic algorithm-based approach *Artif. Intell. Med.* **55** 117–26

Yeung D, Farina D and Vujaklija I 2019 Directional forgetting for stable co-adaptation in myoelectric control *Sensors* **19** 2203

Youn W and Kim J 2010 Estimation of elbow flexion force during isometric muscle contraction from mechanomyography and electromyography *Med. Biol. Eng. Comput.* **48** 1149–57

Zahabi M, White M M, Zhang W, Winslow A T, Zhang F, Huang H and Kaber D B 2019 Application of cognitive task performance modeling for assessing usability of transradial prostheses *IEEE Trans. Hum.-Mach. Syst.* **49** 381–7

Zainuddin N A, Anuar N F, Mansur A L, Fauzi N I M, Hanim W F and Herman S H 2015 Resistive-based sensor system for prosthetic fingers application *Proc. Comput. Sci.* **76** 323–9

Zbinden J, Lendaro E and Ortiz-catalan M 2021 Prosthetic embodiment: review and perspective on definitions, measures, and experimental paradigms *TechRxiv Preprint* (https://doi.org/10.36227/techrxiv.14450058.v1)

Zbyszynski M, Di Donato B and Tanaka A 2019 The effect of co-adaptive learning & feedback in interactive machine learning *ACM CHI: Human-Centered Machine Learning Perspectives Workshop (Glasgow, United Kingdom, 4 May 2019)*

Zecca M, Micera S, Carrozza M C and Dario P 2002 Control of multifunctional prosthetic hands by processing the electromyographic signal *Crit. Rev. Biomed. Eng.* **30** 459–85

Zhai X, Jelfs B, Chan R H and TIN C 2017 Self-recalibrating surface EMG pattern recognition for neuroprosthesis control based on convolutional neural network *Front. Neurosci.* **11** 379

Zhang X, Chen X, LI Y, Lantz V, Wang K and Yang J 2011 A framework for hand gesture recognition based on accelerometer and EMG sensors *IEEE Trans. Syst. Man Cybern.* A **41** 1064–76

Zhang Y, Xiao R and Harrison C 2016 Advancing hand gesture recognition with high resolution electrical impedance tomography *Proc. 29th Annual Symp. on User Interface Software and Technology* pp 843–50

Zhao J, YU Y, Wang X, MA S, Sheng X and ZHU X 2022 A musculoskeletal model driven by muscle synergy-derived excitations for hand and wrist movements *J. Neural Eng.*

Zhao T, LIU J, Wang Y, LIU H and Chen Y 2019 Towards low-cost sign language gesture recognition leveraging wearables *IEEE Trans. Mob. Comput.* **20** 1685–701

Zheng M, Crouch M S and Eggleston M S 2021 Surface electromyography as a natural human-machine interface: a review (arXiv:2101.04658)

Zuckerberg M 2021 Facebook reality labs (available at: https://tech.fb.com/ar-vr/)

Zuo S, Heidari H, Farina D and Nazarpour K 2020 Miniaturized magnetic sensors for implantable magnetomyography *Adv. Mater. Technol.* **5** 2000185