

Spectral QC Summary Report

This QC report is for guidance only. Spectra are expected to vary somewhat by instrument and protocol; minor deviations from library references are normal.

Page 1 of 2

Fluorophore	Similarity	Status
BV421	0.9694	WARNING
cFluor V450	0.9962	PASS
BV510	0.995	PASS
V547	0.9985	PASS
BV570	0.9959	PASS
BV650	0.9946	PASS
BV711	0.9974	PASS
BV750	0.9988	PASS
BV785	0.999	PASS
cFluor B515	N/A	No Reference Available
cFluor B532	N/A	No Reference Available
cFluor B548	N/A	No Reference Available
cFluor B690	N/A	No Reference Available
cFluor BYG575	N/A	No Reference Available
cFluor YG584	N/A	No Reference Available
cFluor BYG610	N/A	No Reference Available
cFluor BYG667	N/A	No Reference Available
cFluor BYG710	N/A	No Reference Available
cFluor BYG781	N/A	No Reference Available
cFluor R659	0.9963	PASS

Spectral QC Summary Report


This QC report is for guidance only. Spectra are expected to vary somewhat by instrument and protocol; minor deviations from library references are normal.

Page 2 of 2


Fluorophore	Similarity	Status
cFluor R668	N/A	No Reference Available
cFluor R685	N/A	No Reference Available
cFluor R720	0.9977	PASS
cFluor R780	N/A	No Reference Available
cFluor R840	0.9962	PASS
ViaDye Red	N/A	No Reference Available

BV421
Cosine Similarity: 0.9694

BV421 — Experiment ■ Library Reference




cFluor V450
Cosine Similarity: 0.9962



BV510
Cosine Similarity: 0.995


BV510 Experiment Library Reference



V547


Cosine Similarity: 0.9985

V547 Experiment Library Reference




BV570
Cosine Similarity: 0.9959

BV570 — Experiment ■ Library Reference




BV650
Cosine Similarity: 0.9946

BV650 — Experiment ■ Library Reference




BV711
Cosine Similarity: 0.9974

BV711 — Experiment ■ Library Reference




BV750
Cosine Similarity: 0.9988

BV750 — Experiment ■ Library Reference




BV785
Cosine Similarity: 0.999

BV785 — Experiment ■ Library Reference




cFluor B515
No Reference Available

cFluor B515 — Experiment ■ Library Reference




cFluor B532
No Reference Available

cFluor B532 — Experiment ■ Library Reference




cFluor B548
No Reference Available

cFluor B548 — Experiment ■ Library Reference




cFluor B690
No Reference Available

cFluor B690 — Experiment ■ Library Reference




cFluor BYG575
No Reference Available

cFluor BYG575 — Experiment ■ Library Reference




cFluor YG584
No Reference Available

cFluor YG584 Experiment Library Reference




cFluor BYG610
No Reference Available

cFluor BYG610 — Experiment ■ Library Reference




cFluor BYG667
No Reference Available

cFluor BYG667 — Experiment ■ Library Reference




cFluor BYG710
No Reference Available

cFluor BYG710 — Experiment ■ Library Reference




cFluor BYG781
No Reference Available

cFluor BYG781 — Experiment ■ Library Reference




cFluor R659
Cosine Similarity: 0.9963

cFluor R659 — Experiment ■ Library Reference




cFluor R668
No Reference Available

cFluor R668 — Experiment ■ Library Reference




cFluor R685
No Reference Available

cFluor R685 — Experiment ■ Library Reference




cFluor R720
Cosine Similarity: 0.9977

cFluor R720 Experiment Library Reference




cFluor R780
No Reference Available

cFluor R780 — Experiment ■ Library Reference



cFluor R840
Cosine Similarity: 0.9962

cFluor R840 — Experiment ■ Library Reference



ViaDye Red
No Reference Available

ViaDye Red — Experiment ■ Library Reference

