

## Master data

Customer no.	Report no.	20240226135614
Customer name	Company	Gossen Metrawatt Südwestpark 15 90449 Nürnberg
Representative	Inspector	Wilfred van den berg

## Testdevice

Testdevice ident	426739392	Designation
Manufacturer		Serial number
Type		Notice
Cost center		Department











## Reporting

Next reporting	1/5/2025
Utilized measuring	SecuTest S4 BL5710890004

## Classifications

Name	Value
Protection class	Class II
Connection	test socket
Meas. Mode	Passive

## Measured Values

		1/5/2024 12:02 PM		
Measurement step name	Measurement name	Measured value	Limit value	Passed
Visuele insp.	Sight check			
RISO SK II	RINS	>300 MΩ	≥2.00 MΩ	
IT Alt.	IT≈	2 μA	≤500 μA	
Functietest	P	0 W		
Functietest	S	0 VA		
Functietest	PF	1.00		
Functietest	I	0.00 A		
Functietest	U	221.2 V		
Functietest	f	50.0 Hz		

## Signature

Representative			Inspector		
2/26/2024			Nürnberg 2/26/2024		
Location	Date	Signature	Location	Date	Signature