ACSM Submission Tool v.1.0

Installation

To load the ipf procedure within Igor, the user can:

- Place the file in the Igor Procedure folder, to automatically loaded at the startup of Igor
- Double click on the file

Setup

The panel can be loaded using PopUp Menu “ACMCC Export”.

- Set station name : The drop down menu contains a list of different stations. If not present, select Other, and type the name. This variable is only indicative, it is not used for metadata, only for filenaming. It is however required to be consistent.
- Procedure automatically detects type of ACSM, based on folders present in the data browser. In the case of Q-ACSM, serial number is automatically detected from DAQ_Matrix. In the case of ToF-ACSM, serial number shall be entered manually.
- Lens & vaporizer type can be set using corresponding dropdown menus.
- Data folder where text files will be saved needs to be set. The adress by default points towards a “default” NextCloud folder. It can be changed if necessary.

Requirements

- ACSM measurements need to have already started
- On acsm_local (for Q-ACSM), TimeSeries and Ion Transmission correction needs to be applied (and corresponding checkboxes checked)

Once set, the user can press start. It will check every 5min if there is a new data point. If yes, then export is triggered.