



These inputs are backwards.
On the PCB, the traces should be cut
and wires used to switch the inputs.

* 1% tolerance or better. 5–10% otherwise.

** Ceramic X5R or X7R.

*** The combination of the divider on V– and the voltage increase into V+ due to input bias current should provide adequate offset over the full range of Vin so that the output of the comparator is stable, but does not false-trigger due to Vout regulation tolerance.

Sheet: /

File: psu.sch

Title: Simple Lab Power Supply Alex Striff <striffa@reed.edu> <<https://github.com/jfjhh/lab-psu>>

Size: A4 Date: 2019–06–01

Rev: B

KiCad E.D.A. kicad 5.1.0

Id: 1/1