

PromQ Prime

Ultrapure Water Purification System Specifications

SI. No	Name of the part	Specification		Qty	Material
		Parameter/unit	Value		
1	Cabinet	Height/mm Width/mm Depth/mm	490 250 562	01	Filter- Polypropylene Housing PP
2	Solenoid Valve	Power Requirement	220V AC	04	Brass alloy
3	Booster pump	Nominal Flow Rate/ LPM Ordinary Pressure/ psi Operating Pressure/ Kg/cm ² Voltage Amps	1.8 80 1.5 48V 2.0A	01	Anodized Aluminium Casing Neoprene Diaphragm
4	All water Tubing's connectors/fittings with gaskets	Dia/mm	6	-	Polypropylene with Silicone gaskets
5	UPC	Dia/Inch Length/Inch	2.5 10	05	Mixed Bed Resin Polypropylene
6	UV lamp	Ultraviolet sterilizer Model: UV Mini LITE Wavelength; 254nm, Power consumption/Watt	11	01	UV tube with SS housing
7	End point filter	Nylon capsule Filter Pore Size/μm	0.22μ	01	Hydrophilic Nylon 66 membrane
8	Conductivity	K=0.01		01	Cell electrodes made up of SS 316 L Housing Nylon