Figure 1. Distributed Control and extraction system at NMGRL. Each hardware component, such as a mass spectrometer or laser, is independently controlled by a dedicated computer and application. Applications communicate and interact with one another via a Local Area Newtork (LAN) using the TCP or UDP protocols. For visual clarity the connections between Pychron Valve and all of the valves it controls are not display.