

 <b>COROMANDEL INTERNATIONAL LIMITED</b>	Format No.	INST/F/06
	Issue No.	01
	Issue Date	03/10/2017
	Revision No.	01
	Revision Date:	03/10/2017

PREVENTIVE MAINTENANCE CHECKPOINTS-ANALYTICAL INSTRUMENTS (pH/DO/TOC/O2/TSS)				
SECTION:			PM DATE:	
TAG NO:				
DESCRIPTION/LOCATION				
SR.	JOB	POINTS TO BE CHECKED	CHECK (v)	REMARK
1	ANALYTICAL INSTRUMENTS (pH/DO/TOC/O2/TSS)	Take permit from the plant		
2		Isolate the power and clean the instrument		
3		Check all connections and tighten it		
4		Switch on power		
5		Take standard solution according to analytical instrument		
6		Put the sensor in standard solution and check the reading on display, adjust if required		
7		Fix back to original location, update instrument calibration tag and cover with polythene		
8		Note down reading in calibration record and mention all tags in shift Log book		
9		Close the permit and handover equipment		

Special Remarks:		
Checked by:	Verified by:	Reviewed by: