

 COROMANDEL INTERNATIONAL LIMITED	Format No.	INST/F/19
	Issue No.	00
	Issue Date	06/10/2018
	Revision No.	00
	Revision Date:	06/10/2018

PREVENTIVE MAINTENANCE CHECKPOINTS-CEMS (Online Stack Monitoring System)				
SECTION:			PM DATE:	
TAG NO:				
DESCRIPTION/LOCATION				
SR.	JOB	POINTS TO BE CHECKED	CHECK (✓)	REMARK
1	CEMS PANNEL	Take permit from the plant		
2		Is Sample Probe Filter Clean & Dry?		
3		Is SS Filter Clean & Dry?		
4		Is Drain Pot chocked?		
5		Do Leakage Test. Is there any leakage?		
6		Clean the panel properly		
7		Tighten the terminations. Provide proper glanding to avoid dust and moisture ingress.		
8		IS the SPM lens clean?		
9		Close the permit and handover equipment		

Special Remarks:		
Checked by:	Verified by:	Reviewed by: