

 COROMANDEL INTERNATIONAL LIMITED	Format No.	INST/F/05
	Issue No.	01
	Issue Date	03/10/2017
	Revision No.	01
	Revision Date:	03/10/2017

PREVENTIVE MAINTENANCE CHECKPOINTS-TEMPERATURE GAUGES				
SECTION:			PM DATE:	
TAG NO:				
DESCRIPTION/LOCATION				
SR.	JOB	POINTS TO BE CHECKED	CHECK (v)	REMARK
1	TEMPERATURE GAUGES	Take permit from the plant		
2		Isolate the system		
3		Check the glass & clean		
4		Clean the gauge with cotton waste		
5		Check the pointer, if not ok then set it with ambient temp.		
7		If not ok, then remove it from the location		
8		Take to test with Temp. bath and calibrate		
9		Fix back to original location, cover with polythene and close the permit		
10		Note down reading in calibration record and mention all tags in shift Log book		

Special Remarks:		
Checked by:	Verified by:	Reviewed by: