

Scioto Usage Report

11 Nov 2024

FW: 1.4.0

BTID: iQ-50C8

SN: 10444

Setup

Target Pressure (inHg)		Step Rate	
Donning	3	25	MED
Low	4		Steps/min
Med	7	50	HIGH
High	10		Steps/min

Performance

Low

Med

High

ASSESSING

Seal Quality

Graph is not available

Outcome

PEQ Patient Survey

N/A

/100

Satisfaction

N/A

/100

Utility

Cadence Variation

N/A

The ambulation cadence with this prosthesis has been recorded and analyzed. The user demonstrates a sustainable variability of cadence over the following range in steps/minute.

Steps/min

Step Count (session)

Low	0	Total
Med	0	0
High	0	