



Nanoscience and nanotechnologies: hopes and concerns

RFID technology for human implant devices

Technologie RFID pour implants dans le corps humain

Hervé Aubert^{a,b,*}

^a Laboratoire d'Analyse et d'Architecture des Systèmes (LAAS) – CNRS, France

^b University of Toulouse, INP, UPS, INSA, ISAE, 7, avenue du colonel Roche, 31000 Toulouse, France

ARTICLE INFO

Article history:

Available online 3 September 2011

Keywords:

RFID technology
Human implant devices
Nanotechnology

Mots-clés:

Technologie RFID
Implant humains
Nanotechnologies

ABSTRACT

This article presents an overview on Radio Frequency Identification (RFID) technology for human implants and investigates the technological feasibility of such implants for locating and tracking persons or for remotely controlling human biological functions. Published results on the miniaturization of implantable passive RFID devices are reported as well as a discussion on the choice of the transmission frequency in wireless communication between a passive RFID device implanted inside human body and an off-body interrogator. The two techniques (i.e., inductive coupling and electromagnetic coupling) currently used for wirelessly supplying power to and read data from a passive implantable RFID device are described and some documented biomedical and therapeutic applications of human RFID-implant devices are finally reported.

© 2011 Académie des sciences. Published by Elsevier Masson SAS. All rights reserved.

RÉSUMÉ

Cet article dresse un état de l'art sur la technologie des puces RFID implantées dans le corps humain. Il explore la faisabilité d'implants permettant de localiser en temps réel des personnes ou encore de contrôler à distance les fonctions biologiques du corps humain. Des résultats sur la miniaturisation des puces RFID passives et implantables dans le corps sont présentés. Des considérations sur le choix de la fréquence de la communication sans fil entre l'implant RFID et son lecteur sont formulées. Deux techniques usuelles (à savoir le couplage inductif et le couplage électromagnétique) pour alimenter sans fil et à distance l'implant RFID passif et pour lire les données de l'implant sont décrites. Des applications biomédicales et thérapeutiques des implants RFID sont finalement données.

© 2011 Académie des sciences. Published by Elsevier Masson SAS. All rights reserved.

1. Introduction

Implantable Radio Frequency Identification (RFID) devices are known as a technology for the remote identification (tagging) of animals or persons. A classical example of human identification based on RFID implant is given by the *VeriChip tag* (11 mm long and about 1 mm in diameter) which contains numerical code (identifiers) readable at 10 centimeters or less using a handheld reader [1]. When illuminating by the low-frequency magnetic field (134 kHz) generated by the reader, the RFID tag powers itself up and transmits a 16-digit code that is unique to the tag. Recent developments in RFID technology

* Correspondence to: University of Toulouse, INP, UPS, INSA, ISAE, LAAS – CNRS, 7 Avenue du colonel Roche, 31000 Toulouse, France.

E-mail address: aubert@enseiht.fr.

have opened up new potential applications, where RFID devices are implanted in the human body for locating and tracking persons or for remotely controlling human biological functions. Such applications may raise ethical and juridical problems, but these are outside the scope of the present survey. In this article only the advances and challenges in RFID technology for human implant devices are reported.

Sub-micron sized “spychip” implant connected to the Global Positioning System (GPS) is not achievable using the now-days technology. A GPS device that would be required for the real-time tracking on Earth presents dimensions that are too large for implanting inside the human body. Miniaturization of such a device for building a “rice-grain-sized GPS” implant is a very challenging technological issue that has to be solved before addressing the feasibility of “Human implant Personal GPS Tracker” [2].

RFID technology for human implants is generally based on battery-less (passive) devices and allows achieving very short read range, typically 10 cm or much less (see, e.g., Freudenthal et al. [3]). Consequently RFID implants cannot communicate with GPS satellites (or other equivalent satellite-based positioning system) that would be required for the real-time tracking of a person. RFID technology is then not suitable for the real-time tracking of a person anywhere on the planet through a RFID chip deeply implanted in his body.

The remote control of human biological functions by using RFID technology seems to be more realistic. Following the estimations of Burke et al. [4], a single chip (including antenna) radio system with on-board sensors of size $100 \times 100 \times 1 \mu\text{m}^3$ seems to be feasible with available technology. Consequently, the wireless communication with small RFID implants in humans for reading information about chemical or physical quantities in biological systems or for providing the remote activation/deactivation of biochemical activity inside a single living cell seems to be “not completely out of reach” [4]. Moreover, using nanotechnology, Jensen et al. [5] have reported the fabrication of an FM-radio using a single nanotube of 1 μm length and 10 nm in diameter. The nanotube radio could be inserted inside the living human cell to give a subcellular remote and real-time controlled interface [6]. This recent result in miniaturization of radio-chip could also find application in fabrication of smartdusts [7], i.e., tiny objects with sensing and communication abilities that may be massively distributed over some area for the remote tracking of human presence/displacement [8] (a recent review presented in [9] discusses various techniques to take smartdust in sensor network beyond millimeter dimensions to the micrometer level). Another application could be on the fabrication of the very small physiological sensors implanted in human and communicating with a single central unit in Wireless Body Area Networks (WBAN) [10].

Using current technology, a RFID device implanted inside the human body and wirelessly communicating with an off-body (or body-worn) reader can be *very small* and *passive* (see Section 2): the smallest reported passive RFID device operating at 2.45 GHz occupies an area of $400 \times 400 \mu\text{m}^2$ [11] (see Section 3) while, at 13.56 MHz a RFID device with an area of $2.5 \times 2.5 \text{ mm}^2$ is available [12]. However the wireless communication with passive RFID device implanted inside the human body is generally performed at *very short range*. Based on inductive coupling (in free space) this range was measured to be 1 cm at 13.56 MHz [13,14] and shorter than 1 mm at 2.45 GHz [15,16] (see Section 4.1). Based on electromagnetic coupling (in free space) this range is between 1 m to 15.7 m at 900 MHz and may reach 7.5 cm at 5.8 GHz [17] (see Section 4.2). For small passive transponders ($2 \text{ mm} \times 10 \text{ mm}$) deeply implanted in the human body (at 10 cm to 30 cm in depth) and power supplied by inductive coupling at 13.56 MHz, the read range was found to be shorter than 10 cm from the body [3].

This article presents an overview on RFID technology for human implant devices. Section 2 reports the technical requirements that human implants having sensing and communication abilities have to satisfy. Justifications for the use of passive RFID devices implanted inside human body for achieving the remote electronic control at bio-molecular scale level conclude Section 2. Section 3 is devoted to documented results on the miniaturization of RFID devices, while Sections 4.1 and 4.2 report wireless techniques for powering such devices. A discussion on the choice of the transmission frequency in wireless communication between RFID device implanted inside human body and an off-body interrogator is given in Section 4.3. Nowadays RFID devices implanted in the human body are expected: (1) to transmit measurements of chemical or biological data inside the body; and (2) to monitor biological activity or physiological functions of the human. Some published biomedical and therapeutic applications of human RFID-implant devices are finally reported in Section 5. The conclusion identifies some technological challenges to face for demonstrating the feasibility of *in vivo* remote control over human bio-molecular machinery.

2. RFID technology for human implant devices

2.1. Technical requirements for human implant devices

Many technical requirements have to be fulfilled by a human implantable device having sensing, computing and communication abilities. *In vivo* implants have to be as small as possible. Moreover, it may be impractical to use a battery for supplying power to an electronic device injected into the body: the large size of batteries compromises their utilization and a passive (also called *battery-less* or *fully autonomous*) device may be preferred in practice. Another requirement concerns the *selectivity* of the remote control of human biological functions. For example, the remote control of a specific bio-molecular event inside a living cell has to be spatially localized to avoid affecting other biological activities in the cell. Additionally *bio-compatible material* has to be used for fabricating the implant device and its packaging in order to avoid tissue reactions. Eventually, undesirable migration of the device inside the body has to be controlled or avoided. Moreover

secure and reliable wireless transmission of the personal/confidential data recorded by the device inside the body has to be established. Guaranteed confidentiality and high reliability are certainly among the main criteria that must be given priority to implement such wireless technology. Finally, for *safety issues* related to antennas implanted in human tissue, electromagnetic energy radiated (or backscattered/re-radiated) by an implant device for wirelessly communicating with the outside world must respect the IEEE human tissue exposure standard [18,19]. To satisfy safety recommendations, an implantable antenna operating at 868 MHz when placed inside tissue layers has been designed in [20] to ensure a Specific Absorption Rate (SAR) lower than 2 W/kg. In Kim and Rahmat-Samii [21] implantable antennas have been designed for operating at the *Medical Implant Communications Service* frequency band (402–405 MHz), which is regulated by the *Federal Communications Commission* and the *European Radiocommunications Committee* for ultra-low-power active medical implants. To satisfy the recommendations of these institutions, antennas allowing a peak SAR lower than 1.6 mW/g have been designed (estimation of the SAR in practical applications may be found in Kim and Rahmat-Samii [21]).

2.2. Remote control of molecular scale biochemistry

The first attempt to demonstrate the feasibility of the remote electronic control at bio-molecular scale level has been reported in 2002 by Hamad-Schifferli et al. [22]. The control of the hybridization behavior of DNA molecules has been performed by monitoring the heating of gold nanoparticles interfaced (linked) to DNA. Such local controlled heating performed by electromagnetic coupling at 1 GHz induces the reversible hybridization/dehybridization of the DNA while leaving surrounding molecules relatively unaffected. Although the physics of heating nanometer particles has been little explored [23] and will certainly be the subject of future investigations, this experiment demonstrates for the first time the direct radio frequency (RF) control over bio-molecular machinery in a specific and fully reversible manner. Remote control at such a very small scale level seems not to alter bio-molecular events occurring in the surrounding medium.

A discussion of the technological feasibility of tiny devices implanted inside a living cell for the remote control of a wide range of biological functions or the detection of presence or concentration of a protein or antibody can be found in Burke and Rutherglen [4]. Based on known technological achievements and patents in nanoelectronics a “nanorobot” consisting in embedded smart sensors implanted in human body for tele-operated surgical procedures or early disease diagnosis has been designed in Cavalcanti et al. [24]. Up to now, such nano-devices have not been fabricated.

2.3. RFID technology for human implant devices

Basically, RFID is a technology that uses communication via electromagnetic waves to exchange data between an *interrogator* (also known as the *reader*) and an object called the *transponder* for identification and tracking purposes. The transponder is generally composed of an integrated circuit for storing and processing the data signal and an antenna for receiving and transmitting the signal [25]. Such technology may fulfill the technical requirements listed in Section 2.1. Fabrication of very small RFID device with communicating ability is achievable (see Section 3) and, following the recent estimations of Burke and Rutherglen [4], it appears technically possible to fabricate RFID device at the bio-molecular scale that interrogates its local chemical environment (sensing) or/and controls the biological machinery inside a single living cell. The remote monitoring at such a very small scale level is assumed to provide the selectivity requirement, that is, the local control of bio-molecular machinery without damaging cells or even affecting the biological events in the surrounding medium.

Moreover RFID technology allows real-time remote monitoring of *fully autonomous implants* in human body (see Sections 4.1 and 4.2). In the power harvesting technique based on inductive or electromagnetic couplings, the RFID implant receives power from the interrogator and emits back to it a signal that contains data. For such passive solution RF power deposition inside tissue can be easily made compliant with safety regulations mentioned in Section 2.1. In addition some technical requirements, relative to the fabrication of bio-compatible implant devices and to the establishment of secure and reliable wireless transmission of data recorded by implants, have already been addressed in the framework of RFID technology and preliminary solutions have been found (see, e.g., [26] and [27]). The read range of communication for RFID implanted in human body is found to be shorter than 10 cm (see Section 4). Such short read range could be advantageous when RFID implants are subject to security/confidentiality requirement.

3. Technological results on the miniaturization of RFID devices

As mentioned in [4], the size of single chip RFID transponders with available technology is dominated by the antenna size and not by one of the integrated circuit (IC) used by the RFID transponder for storing and processing the data signal. The smallest size of the IC is $50 \times 50 \times 5 \mu\text{m}^3$ without antenna, i.e., for an off-chip antenna [28–30]. However, using either electromagnetic waves or inductive coupling for powering passive RFID device (see Sections 4.1 and 4.2) an off-chip antenna is typically centimeters in size. Smaller RF antennas (typically few mm by a few mm in size) can be fabricated when incorporating antenna on the same chip as the signal processing components (on-chip antenna or OCA). Using $0.25 \mu\text{m}$ CMOS technology Abrial et al. [31] have reported in 2001 a System-On-Chip (operating at 13.56 MHz) including coil antenna (1.5 mm^2), RF communication and digital processing within a surface of 16 mm^2 (including pads) while *Maxell Corp.* [12] has reported an OCA designed on a chip area of $2.5 \text{ mm} \times 2.5 \text{ mm}$. At the same operating frequency, a mini-RFID transponder




Fig. 1. RFID system based on inductive coupling: (a) Constitutive elements and (b) equivalent circuit.

occupying an area of $8 \times 3 \text{ mm}^2$ (and a volume of 25 mm^3) has been recently published in [13]. At 2.45 GHz, a RFID-chip area of $1 \times 0.5 \text{ mm}^2$ has been reported [15], while using $0.18 \mu\text{m}$ CMOS technology, Usami et al. have presented in 2004 a smaller RFID chip occupying an area of $400 \times 400 \mu\text{m}^2$ only [11,32,33]. Using existing CMOS technology, Burke et al. estimate that the fabrication of a RFID-chip volume of $100 \times 100 \times 5 \mu\text{m}^3$ “is not completely out of reach” [4].

New manufacturing techniques at nano-scale level may help to miniaturize RFID devices. They may allow, for example, exploring the feasibility of a very small antenna by using a carbon nanotube [34] (overview of the high-frequency properties nanotube and nanowire technology for microwave applications can be found in Burke et al. [35]). Using a single carbon nanotube Jensen et al. have reported in 2007 the first nano-sized radio-wave receivers [5]. A single carbon nanotube in vacuum ($0.1 \mu\text{Torr}$) of $1 \mu\text{m}$ length and 10 nm in diameter served at the same time as antenna, tuner, amplifier and frequency demodulator. The antenna receives signals via high-frequency mechanical vibrations of the nanotube and the magnitude of these vibrations is high only when the frequency of the incident wave coincides with the (tunable) resonance frequency of the nanotube. A battery has been connected as a power source for the operation of the radio and consequently, a transmission and reception of signals on a distance of several meters may be achieved with this active devices. Another tiny radio using nanotechnology is also demonstrated in Rutherglen and Burke [36] where a carbon nanotube is used to demodulate amplitude modulated signal. These advances in fabrication of radio-chip at nano-scale level may help the advent of nano-scale RFID device. Due to the small size, such devices may then potentially be inserted inside the living human cell to provide a sub-cellular remote controlled interface.

4. Wireless techniques for powering passive RFID-implant devices

Passive RFID technology allows one to overcome the limitations in terms of lifetime, reliability and size of RFID devices requiring a battery. The electromagnetic field received from the interrogator is the only source for supplying power to the passive transponder. The two techniques currently used for wirelessly supplying power to (and read data from) passive RFID chip are reported in this section. The analysis given in Sections 4.1 and 4.2 are performed in free-space environment and in absence of electromagnetic noise, while Section 4.3 considers the performances of actual wireless communication between an implanted RFID device and an off-body interrogator.

4.1. Inductive coupling

The power required for operating a passive transponder may be supplied by an interrogator via *inductive (or near field) coupling*. Fig. 1(a) shows the constitutive elements of the corresponding RFID system. Time-harmonic magnetic field is generated by the low-frequency current in the interrogator coil. In the near field region inductive coupling may occur between this coil and the transponder coil.

The equivalent circuit of the RFID system based on inductive coupling is shown in Fig. 1(b). The resistor R_T models the total losses in the integrated circuit (IC) used by the RFID transponder for storing and processing the data signal.




Fig. 2. Constitutive elements of a RFID system based on electromagnetic coupling.

L_1 and L_2 represent the transponder and the interrogator coil inductances, respectively, while M denotes the mutual inductance. r_1 and r_2 are the series resistances of the interrogator and transponder coils, respectively. The capacitor C connected in parallel with the transponder antenna coil L_2 form a resonant circuit at the chosen transmission frequency f . Then the magnitude V of the time-harmonic voltage at the terminal of L_2 is given by (see, e.g., [25]):

$$V = \frac{\omega k \sqrt{L_1 L_2}}{\sqrt{\left(\frac{L_2 \omega}{R_T} + r_2 C \omega\right)^2 + (1 - L_2 C \omega^2 + \frac{r_2}{R_T})^2}} I \quad (1)$$

It reaches its maximum value at resonance frequency. In Eq. (1), I denotes the magnitude of the time-harmonic current in the interrogator coil, $\omega = 2\pi f$ is the radian frequency of the transmission. The coupling coefficient $k = M/\sqrt{L_1 L_2}$ between the two coils depends on the distance d between the transponder coil and the interrogator coil [37]. In practice this coefficient is close to 10^{-3} (e.g., $k = 4 \times 10^{-3}$ at a distance $d = 5$ mm and $k = 1.5 \times 10^{-3}$ at $d = 1$ cm in [13]). The voltage V given by Eq. (1), once rectified and smoothed with a regulator, is used for supplying power to the IC. It must be large enough to operate this circuit. For a given distance d between the two coils, this condition allows to derive the minimum magnetic field that must be generated by the interrogator coil. This field decreases rapidly with distance inside the near field region (as $1/d^3$ where d designates the distance from the center of the interrogator coil to the transponder coil [25]) and consequently, when the two coils are separated, it may rapidly become too weak for providing the transponder with the energy it needs to operate. Following [13], at a frequency $f = 13.56$ MHz, voltages $V \approx 1$ Volt (required for operating the RFID IC) may be induced in a small implantable transponder coil $L_2 \approx 100$ nH from an interrogator coil $L_1 \approx 1$ μ H located at a distance d smaller than 1 cm.

Lu et al. [13] have recently reported wireless communication with implantable RFID device (occupying approximately 25 mm³ in volume) in the frequency range of 13.56–27 MHz at a distance of 1 cm. A successful wireless communication at 13.56 MHz has also been established by Sasaki et al. [14] at the same distance. Guo et al. [15] have reported a successful communication with a small 2.45 GHz RFID device (with a chip-area of 1 mm \times 0.5 mm) at a distance of 1 mm. The design of a 2.45 GHz RFID device with its on-chip antenna smaller than 0.5 mm² with a thickness of 0.1 mm is reported in Chen et al. [16]: with the reader generating an output power of 500 mW, the RFID system was able to perform with RF read/write functions at a distance smaller than 0.5 mm.

4.2. Electromagnetic coupling

The technique based on *electromagnetic (or far field) coupling* involves the use of electromagnetic waves that propagate from the antenna in the far field region to power the passive RFID chip. To transmit data, the IC changes the impedance of the transponder antenna which backscatters some of the incident RF energy to the interrogator. Fig. 2 shows the constitutive elements of the RFID system based on such a coupling.

In free space, the electromagnetic power P_{rect} received by the transponder antenna (gain G_T) is given by Balanis [38]:

$$P_{\text{rect}} = G_I \left(\frac{\lambda}{4\pi d} \right)^2 G_T P_I \quad (2)$$

where P_I designates the power transmitted by the interrogator antenna (gain G_I), λ is the free-space wavelength and d denotes the distance between the interrogator and transponder antennas (polarization mismatch between antennas is assumed negligible; this can be obtained by choosing appropriate antenna orientation). The attenuation factor $(\frac{\lambda}{4\pi d})^2$ is called the *free-space path loss*. The power P_R back-scattered by the transponder is given as follows [38]:

$$P_R = \sigma \frac{(\lambda_0 G_I)^2}{(4\pi)^3 d^4} P_I \quad (3)$$

where σ designates the radar cross-section of the transponder, that is, the efficiency with which the transponder reflects incoming electromagnetic waves. As it can be observed from Eq. (3), power is strongly attenuated by a factor $1/d^4$ (in free space) when it returns back to the interrogator. The range of the wireless reading of the transponder data can be derived from the interrogator sensitivity $P_{R\text{min}}$.

Systems based on electromagnetic coupling operate usually at UHF frequencies of 868 MHz (Europe) and 915 MHz (USA) and at microwave frequencies of 2.5 GHz and 5.8 GHz [25]. Jia et al. [39] have reported a RFID humidity sensor capable of

passive wireless sensing at 868 MHz. The reader antenna was positioned at a fixed distance of 1.5 m in front of the RFID chip. Very recently, a great improvement with respect to performances of existing passive UHF/microwave RFID chip has been reported in [17]: a fully integrated and implantable 900 MHz RFID chip has been fabricated and the measured range is found to be 15.7 m, assuming an off-chip $G_T = 0$ dB antenna gain, while the device with its on-chip antenna operating at 5.8 GHz can be read and powered-up up to 7.5 cm distance.

A passive RFID device incorporating sensor based on electromagnetic transduction has been recently proposed in Jatlaoui et al. [40–42]. By using this device, wireless data transmission and measurement of the physical quantity (e.g., pressure, temperature, gas concentration) can share the same electromagnetic wave. No electronics, no rectifier and no IC are needed and consequently, the overall RFID device losses are minimized. This allows transmission ranges of some tens of meters. Based on this technique, passive RFID devices incorporating pressure sensor [43–47], temperature sensor [48,49] or gas concentration sensor [50–54] have been successfully tested.

4.3. Frequency and read range in wireless communication between RFID device implanted inside human body and an off-body interrogator

Low-Frequency (LF) electromagnetic waves are not strongly attenuated by human tissues. However the read range associated with LF (135 kHz) RFID implant is generally shorter than 1 m due to rapid attenuation of the magnetic field with distance (see Section 4.1) and the necessarily small cross-sectional area of implanted transponder coil antenna. Moreover the bandwidth of an LF RFID chip is very small. This technical characteristic implies: (1) very slow data rates (low data transfer); (2) inability of the interrogator to read simultaneously multiple RFID-implant devices; (3) vulnerability to transmission collapse in presence of strong environmental electrical noise (lack of robustness and reliability of the wireless link); and (4) difficulty to incorporate any cryptography in the transmitted signal (lack of security/confidentiality of the data transfer, although the very short range associated with such close coupling systems could be interesting for achieving secure transmission).

In Freudenthal et al. [3] the authors investigate the effectiveness of *High-Frequency (HF)* waves for communicating with implanted devices. Considerations that may justify the use of HF RFID for implanted medical applications can also be found in [55]. HF waves ranging from 1 MHz to 20 MHz are not significantly attenuated by human tissues and thus are suitable for communication with RFID devices implanted inside the human body [56]. Experiments performed at 4 MHz emulating the human tissue electrical properties by introducing water bearing colloids between coils of the inductive coupling RFID system have revealed a small decrease in the amount of power transmitted by the interrogator to the transponder [27]. This attenuation was not large enough to cause data or power flow disruption or to lower the power supply voltage below the IC operating range. The read range of such HF inductive coupling system was then found to be quite the same than one observed in free-space environment (that is 1 cm at 13.56 MHz and shorter than 1 mm at 2.45 GHz). Moreover successful wireless communication with ten HF RFID devices (each one occupying an area of $2 \times 10 \text{ mm}^2$) implanted at three locations within a human body (at 10 cm to 30 cm in depth) has been established in [3]. It has been observed that deeply implanted transponders may be read at distance of 10 cm from the body, i.e., at distances-to-transponders similar to those observed for non-implanted transponders.

From the review of state of the art it can be observed that inductive coupling operating in HF (13.56 MHz and below) frequency range is now one of the most common methods to wirelessly send power and data from off-body interrogator to RFID device implanted inside human body. In 1988, the design of millimeter- and sub-millimeter-sized power sources operating at 2 MHz and 20 MHz for neural implant based on inductive coupling has been proven feasible [57]. Power harvesting of implanted devices through inductive coupling has been successfully performed for various applications, e.g., for recording the neural signals from axons by using implanted telemetry system occupying an area of $4 \times 6 \text{ mm}^2$ and operating at 2 MHz [58], for wireless implantable neural micro-stimulation system [59], or for replacing defective photoreceptors in patients by implanting a retinal chip that achieves the transmission link between the extra-ocular and intra-ocular coil antennas at a frequency ranging from 1 to 10 MHz [60].

Ultra-High Frequency (UHF) RFID device offers significantly higher bandwidth than LF or HF RFID. This allows, e.g., high data rates and the incorporation of privacy/integrity preserving cryptographic protocols for wireless transmission in order to protect information [61]. Moreover, using UHF electromagnetic waves, high level of miniaturization of RFID chip (including its antenna) can be achieved (see Section 3). However UHF waves may be problematic for human RFID implants, due to the following reasons: (1) UHF and microwave fields are strongly attenuated by water (which is the primary constituent of human tissue) and consequently a passive RFID solution allows a very limited communication range (typically shorter than 1 m); (2) the UHF field transmitted by an interrogator located outside the human body is subject to multiple reflections by environmental objects creating interferences and undesirable *shadow zones* (in which RFID chip is not detectable by the interrogator) and peaks or *hot spot* in the field distribution; (3) HF frequencies (in particular 2.45 GHz) may be unsafe for human for long-term exposure and/or when high electromagnetic densities are involved.

In the UHF band the human body, essentially composed of water, bones and tissues, can be viewed as a highly scattering and dissipative electromagnetic propagation channel in which multiple reflections (multi-paths) and strong attenuation occur. Specifically input impedance and radiation pattern of the transponder antenna may depend on the location of the antenna implanted inside the body [21]. In order to perform an analysis of the UHF wireless communication efficiency and to analysis electromagnetic characteristics of antennas implanted inside the human body, a realistic electromagnetic model

of the human body (with its various tissues and geometries) has to be elaborated. It will allow designing a reliable wireless communication system between a RFID device implanted inside human body and an off-body interrogator. The propagation analysis of UHF electromagnetic field in muscle tissue has been presented in [62] while in [21,20] the authors analyze the antenna performances (efficiency, radiation pattern, and input impedance) when implanted in human body or in an equivalent simplified tissue model. In [21] it is observed that the return-loss characteristics of 400 MHz antennas change with the location of the antennas inside the tissue-simulating fluid. The resulting range of the wireless communication with an UHF RFID implant may then be greatly dependent on the implant location in human body.

At an operating frequency between 915 MHz and 1 GHz wireless communication between a transponder antenna and an interrogator antenna separated by 15 mm of bovine muscle tissue has been established by O'Driscoll [63]. In Warty et al. [64] communication between antennas positioned below the scalp covering the frontal region of the skull was presented and the effect of the antenna packaging on the impedance matching of an implanted antenna was also analyzed. Using an inductively powered and implantable wireless neural recording device working at 3.2 GHz, the *in vivo* continuous recording continuous of electro-encephalogram from the hippocampus of an un-tethered rat is possible [65].

5. Biomedical/therapeutic applications of human RFID-implant devices

For *biomedical applications* RFID-implant devices are expected to transmit various chemical or biological parameters measured inside the body. Such devices can be used by diabetes sufferers [66]: a wireless alarm unit implanted in the abdominal region of the patient, continually reads the data of a sensor chip sensitive to glucose concentration of surrounding fluids. Recently, Sauer et al. [27] have used RFID-implant technology for *in vivo* data collecting and wirelessly transmission of electro-encephalograms taken during animal recording.

For *therapeutic applications* RFID-implant devices are expected to monitor biological activity or physiological functions of human body and to modify them. Example is provided by the monitoring of brain function via implanted probes implanted that communicate via a transponder embedded within the skull [67,68]. An array of 16 activated iridium microelectrodes (5–6 mm in length within a cluster approximately 1.8 mm in diameter) suitable for long-term implantation into the brain, records signal from single neurons, and delivers as well localized microstimulation. These microelectrodes may also be used therapeutically [3]. However such therapeutic applications of RFID implants are still at the proof of concept stage. The benefits of such technology in terms of, e.g., improvement of quality of living or of patient care, have to be demonstrated.

6. Conclusion

RFID technology has opened up new expectations in the fabrication of tiny devices implanted in the human body for locating and tracking persons or for remotely controlling human biological functions. A RFID chip implanted inside the human body can transmit its data to an off-body interrogator at very short distance (typically 10 cm or less). It cannot communicate, e.g., with GPS satellites (or other equivalent satellite-based positioning system) that would be required for the real-time tracking of a person. RFID technology is then not suitable for the real-time tracking of a person anywhere on the planet through RFID chip deeply implanted in its body. However, a RFID-based remote and reversible control of human bio-molecular machinery seems to be not completely out of reach using existing technology [4]. The recent advances in fabrication of radio-chip at nano-scale level [5,36] may help the advent of nano-scale RFID chip. Due to their small size, such device could be inserted inside the human living cell to provide a sub-cellular remote controlled interface.

Many technological problems have to be solved for demonstrating the feasibility of the reliable remote control of human bio-molecular machinery using RFID implants: *heat removal* and elaboration of *bio-compatible miniaturized packaging* are two of them [69]. Other questions have to be addressed: Could *new power harvesting technique* based on the properties of surrounding human tissue be exploited for supplying power to RFID devices implanted inside the body? Would electromagnetic fields transmitted by nano-radio implant *damage in surrounding tissue*? Despite very promising recent technological advances, many challenges must be addressed for clearly validating the expected benefits of RFID-implant technology.

References

- [1] K.R. Foster, J. Jaeger, IEEE spectrum: RFID inside, <http://spectrum.ieee.org/print/4939>, July 2007.
- [2] A. Masters, K. Michael, Human-centric applications of RFID implants: The usability contexts of control, convenience and care, in: Proceedings of the Second IEEE International Workshop on Mobile Commerce and Services (WMCS'05), 2005.
- [3] E. Freudenthal, D. Herrera, F. Kautz, C. Natividad, A. Ogrey, J. Sipla, A. Sosa, C. Betancourt, L. Estevez, Evaluation of HF RFID for implanted medical applications, Departmental Technical Reports (CS), 2007, Paper 162, http://digitalcommons.utep.edu/cs_techrep/162.
- [4] Peter Burke, Christopher Rutherglen, Towards a single-chip, implantable RFID system: is a single-cell radio possible?, Biomed. Microdevices 24 (January 2009), doi:10.1007/s10544-008-9266-4, <http://www.ncbi.nlm.nih.gov/pmc/articles/PMC2896640/>.
- [5] K. Jensen, J. Weldon, H. Garcia, A. Zettl, Nanotube radio, Nano Lett. 7 (11) (2007) 3508–3511.
- [6] Ujjal Kumar Sur, World's first single carbon nanotube radio, Current Science 94 (2) (January 2008) 166–167.
- [7] W. Cook, S. Lanzisera, K.S.J. Pister, SoC issues for RF smart dust, Proceedings of the IEEE 94 (6) (June 2006) 1177–1196.
- [8] B. Warneke, M. Last, B. Liebowitz, K.S.J. Pister, Smart dust: Communicating with a cubic-millimeter, Computer 34 (2001) 44–51.
- [9] Michael J. Sailor, Jamie R. Link, Smart dust: nanostructured devices in a grain of sand, Chemical Communications 11 (2005) 1375.
- [10] M. Chen, Body Area Networks: A Survey, ACM/Springer Mobile Networks and Applications (MONET), 2010.
- [11] M. Usami, An ultra small RFID chip: μ -chip, in: Radio Frequency Integrated Circuits (RFIC) Symposium, Digest of Papers, IEEE, 2004, pp. 241–244.
- [12] Coil-on-chip RFID tag, Maxell, Fairlawn, NJ, 2010 [online]. Available: <http://www.maxell-usa.com/index.aspx?id=4;41;432;0>.

- [13] H.M. Lu, C. Goldsmith, L. Cauller, Jeong-Bong Lee, MEMS-based inductively coupled RFID transponder for implantable wireless sensor applications, *IEEE Transactions on Magnetics* 43 (6) (June 2007) 2412–2414.
- [14] S. Sasaki, T. Seki, K. Imanaka, M. Kimata, T. Toriyama, T. Miyano, S. Sugiyama, Batteryless-wireless MEMS sensor system with a 3D loop antenna, *IEEE Sensors 2007* (October 2007) 252–255.
- [15] L.H. Guo, A.P. Popov, H.Y. Li, Y.H. Wang, V. Bliznetsov, G.Q. Lo, N. Balasubramanian, D.-L. Kwong, A small OCA on a $1 \times 0.5\text{-mm}^2$ 2.45-GHz RFID tag—design and integration based on a CMOS-compatible manufacturing technology, *IEEE Electron Device Lett.* 27 (2) (2006) 96–98.
- [16] X. Chen, W.G. Yeoh, Y.B. Choi, H. Li, R. Singh, A 2.45-GHz near-field RFID system with passive on-chip antenna tags, *IEEE Trans. Microw. Theory Tech.* 56 (6) (June 2008) 1397–1404.
- [17] S. Radiom, M. Baghaei-Nejad, G. Vandenbosch, L.-R. Zheng, G. Gielen, Far-field RF powering system for RFID and implantable devices with monolithically integrated on-chip antenna, in: *IEEE Radio Frequency Integrated Circuits Symposium*, 2010.
- [18] IEEE standard for safety levels with respect to human exposure to radio frequency electromagnetic fields, 3 kHz to 300 GHz, *IEEE Standard C95.1-1999*, 1999.
- [19] IEEE recommended practice for measurements and computations of radio frequency electromagnetic fields with respect to human exposure to such fields, 100 kHz to 300 GHz, *IEEE Standard C95.3-2002*, 2002.
- [20] A. Sani, M. Rajab, R. Foster, Y. Hao, Antennas and propagation of implanted RFIDs for pervasive healthcare applications, *Proceedings of the IEEE* 98 (9) (September 2010) 1648–1655.
- [21] J. Kim, Y. Rahmat-Samii, Planar inverted-F antennas on implantable medical devices: Meandered type versus spiral type, *Microw. Opt. Technol. Lett.* 48 (3) (2006) 567–572.
- [22] K. Hamad-Schifferli, J.J. Schwartz, et al., Remote electronic control of DNA hybridization through inductive coupling to an attached metal nanocrystal antenna, *Nature* 415 (6868) (2002) 152–155.
- [23] G.W. Hanson, S.K. Patch, Optimum electromagnetic heating of nanoparticle thermal contrast agents at RF frequencies, *Journal of Applied Physics* 106 (2009) 054309.
- [24] Ariano Cavalcanti, Bijan Shirinzadeh, Robert A. Freitas Jr., Luiz C. Kretly, Medical Nanorobot Architecture Based on Nanobioelectronics, *Recent Patents on Nanotechnology* 2007, vol. 1 (No. 1), Bentham Science Publishers Ltd., February 2007, available at: <http://www.bentham.org/nanotec/samples/nanotec1-1/Cavalcanti.pdf>.
- [25] K. Finkenzeller, *RFID Handbook: Fundamentals and Applications in Contactless Smart Cards and Identification*, Wiley, Hoboken, NJ, 2003.
- [26] L. Ricciardi, I. Pitz, S.F.A. Sarawi, V. Varadan, D. Abbott, Investigation into the future of RFID in biomedical applications, *Proc. of SPIE – Int. Soc. Optical Eng.* 5119 (2003) 199–209.
- [27] C. Sauer, M. Stanacevic, G. Cauwenberghs, N. Thakor, Power harvesting and telemetry in CMOS for implanted devices, *IEEE Trans. Circ. Syst.* 52 (12) (2003) 2605–2613.
- [28] M. Usami, A. Sato, K. Sameshima, K. Watanabe, H. Yoshigi, R. Imura, Powder LSI: an ultra small RF identification chip for individual recognition applications, in: *Solid-State Circuits Conference, Digest of Technical Papers, ISSCC, IEEE International*, 2003.
- [29] <http://www.engadget.com/2007/02/14/hitachis-rfid-powder-freaks-us-the-heck-out>.
- [30] M. Usami, H. Tanabe, A. Sato, I. Sakama, Y. Maki, T. Iwamatsu, T. Ipposhi, Y. Inoue, A $0.05 \times 0.05\text{ mm}^2$ RFID chip with easily scaled-down ID-memory, in: *Solid-State Circuits Conference, Digest of Technical Papers, ISSCC 2007, IEEE International*, 2007.
- [31] A. Abrial, J. Bouvier, M. Renaudin, P. Senn, P. Vivet, A new Contactless Smart Card IC using an on-chip antenna and an asynchronous microcontroller, *IEEE J. Solid-State Circuits* 36 (7) (2001) 1101–1107.
- [32] M. Usami, An ultra small RFID chip: μ -chip, in: *IEEE Asia-Pacific Conference on Advanced System Integrated Circuits (AP-ASIC2004)*, August 4–5, 2004.
- [33] M. Usami, Ultra-small RLID chip technology, in: *13th IEEE International Conference on Circuits and Systems (ICECS '06)*, 2006, pp. 708–711.
- [34] P.J. Burke, S. Li, Z. Yu, Quantitative theory of nanowire and nanotube antenna performance, *IEEE Trans. Nanotechnol.* 5 (4) (2006) 314–334.
- [35] P.J. Burke, Z. Yu, S. Li, C. Rutherglen, Nanotube technology for microwave applications, in: *Proc. of IEEE MTT International Microwave Symposium*, 2005.
- [36] C. Rutherglen, P. Burke, Carbon nanotube radio, *Nano Lett.* 7 (11) (2007) 3296–3299.
- [37] C. Reinhold, P. Scholz, W. John, U. Hillerlingmann, Efficient antenna design of inductive coupled RFID-systems with high power demand, *Journal of Communications* 2 (6) (November 2007) 14–23.
- [38] C.A. Balanis, *Antenna Theory – Analysis and Design*, third edition, John Wiley & Sons, 2005.
- [39] Y. Jia, M. Heiss, Q. Fu, N.A. Gay, A prototype RFID humidity sensor for built environment monitoring, in: *International Workshop on Education Technology and Training & International Workshop on Geoscience and Remote Sensing*, 2008, pp. 496–499.
- [40] M.M. Jatlaoui, P. Pons, H. Aubert, Radio-frequency pressure transducer, in: *37th European Microwave Conference (EuMC'07)*, München, Germany, 8–12 October 2007, pp. 983–986.
- [41] M.M. Jatlaoui, P. Pons, H. Aubert, Pressure micro-sensor based on radio-frequency transducer, in: *IEEE International Microwave Symposium (IMS)*, Atlanta, Georgia, USA, 15–20 June 2008.
- [42] M.M. Jatlaoui, F. Chebila, I. Gmati, P. Pons, H. Aubert, New electromagnetic transduction micro-sensor concept for passive wireless pressure monitoring application, in: *Transducers 2009, 15th International Conference on Solid-State Sensors, Actuators and Microsystems*, Denver, Colorado, USA, June 21–25, 2009.
- [43] M.M. Jatlaoui, F. Chebila, P. Pons, H. Aubert, Pressure sensing approach based on electromagnetic transduction principle, in: *Asia Pacific Microwave Conference (APMC 2008)*, Hong Kong and Macao, China, 16–19 December 2008.
- [44] M.M. Jatlaoui, F. Chebila, P. Pons, H. Aubert, Wireless interrogation techniques for a passive pressure micro-sensor using an EM transducer, in: *European Microwave Week, Nuova Fiera di Roma, Rome, Italy*, 28 September–2 October 2009.
- [45] M.M. Jatlaoui, F. Chebila, P. Pons, H. Aubert, New micro-sensors identification techniques based on reconfigurable multi-band scatterers, in: *Asia-Pacific Microwave Conference (APMC 2009)*, Singapore, December 7–10, 2009.
- [46] M.M. Jatlaoui, F. Chebila, S. Bouaziz, P. Pons, H. Aubert, Original identification technique of passive EM sensors using loaded transmission delay lines, in: *European Microwave Week (EuMW 2010)*, Paris, France, 26 September–1 October 2010.
- [47] M.M. Jatlaoui, F. Chebila, T. Idda, P. Pons, H. Aubert, Phenomenological theory and experimental characterizations of passive wireless EM pressure micro-sensor prototype, in: *The 9th Annual IEEE Conference on Sensors (IEEE Sensors 2010)*, Waikaloa, Hawaii, USA, November 1–4, 2010.
- [48] T.T. Thai, M. Jatlaoui, H. Aubert, P. Pons, G.R. Dejean, M.M. Tentzeris, R. Plana, A novel passive wireless ultrasensitive temperature RF transducer for remote sensing, in: *IEEE International Microwave Symposium (IMS)*, Anaheim, California, USA, May 23–28, 2010.
- [49] T.T. Thai, F. Chebila, M.M. Jatlaoui, M. Mehdi, P. Pons, H. Aubert, G.R. Dejean, M.M. Tentzeris, R. Plana, Design and development of a millimetre-wave novel passive ultrasensitive temperature transducer for remote sensing and identification, in: *European Microwave Week (EuMW 2010)*, Paris, France, 26 September–1 October 2010.
- [50] H. Hallil, F. Chebila, P. Menini, H. Aubert, Feasibility of passive gas sensor based on whispering gallery modes and its RADAR interrogation: Theoretical and experimental investigations, *Sensors & Transducers Journal* 116 (5) (May 2010) 38–48.
- [51] H. Hallil, F. Chebila, P. Menini, P. Pons, H. Aubert, Feasibility of wireless gas detection with an FMCW RADAR interrogation of passive RF gas sensor, in: *The 9th Annual IEEE Conference on Sensors (IEEE Sensors 2010)*, Waikaloa, Hawaii, USA, November 1–4, 2010.
- [52] H. Hallil, P. Menini, H. Aubert, Novel microwave gas sensor using dielectric resonator with SnO_2 sensitive layer, in: *EuroSensors 2009*, Lausanne, Switzerland, 6–9 September 2009.

- [53] H. Hallil, P. Menini, H. Aubert, New microwave gas detector using dielectric resonator based on a Whispering-Gallery-Mode, in: European Microwave Week, Nuova Fiera di Roma, Rome, Italy, 28 September–2 October 2009.
- [54] H. Hallil, P. Menini, H. Aubert, Novel millimeter-wave gas sensor using dielectric resonator with sensitive layer on TiO₂, in: 8th IEEE Conference on Sensors, Christchurch, New Zealand, 25–28 October 2009.
- [55] Magellan Technology, White paper: A comparison of RFID frequencies and protocols, March 2006 [online]. Available: <http://www.magtech.com.au/res/PDF/Brochures%20and%20Documents/Whitepapers/RFID%20Frequencies%20and%20Protocols.pdf>.
- [56] P. Vaillancourt, A. Djemouai, J.F. Harvey, M. Sawan, EM radiation behavior upon biological tissues in a radio-frequency power transfer link for a cortical visual implant, in: Proc. IEEE EMBS Conf., vol. 6, 1997, pp. 2499–2502.
- [57] W.J. Heetderks, RF powering of millimeter- and submillimeter-sized neural prosthetic implants, IEEE Trans. Biomed. Eng. 33 (3) (March 1988) 323–327.
- [58] T. Akin, K. Najafi, R.M. Bradley, A wireless implantable multichannel digital neural recording system for a micromachined sieve electrode, IEEE J. Solid-State Circuits 33 (1) (January 1998) 109–118.
- [59] M. Ghovanloo, K. Najafi, A wide-band frequency-shift keying wireless link for inductively powered biomedical implants, IEEE Trans. Circuits Syst. I, Reg. Papers 51 (12) (December 2004) 2374–2383.
- [60] W. Liu, K. Vichienchom, M. Clements, S.C. DeMarco, C. Hughes, E. McGucken, M.S. Humayun, E. De Juan, J.D. Weiland, R. Greenberg, A neuro-stimulus chip with telemetry unit for retinal prosthetic device, IEEE J. Solid-State Circuits 35 (7) (July 2000) 1487–1497.
- [61] A. Juels, RFID security and privacy: A research survey, Journal of Selected Areas in Communication (J-SAC) 24 (2) (February 2006) 381–395.
- [62] D.D. Arumugam, D.W. Engels, Characterization of RF propagation in muscle tissue for passive UHF RFID tags, in: XXIX General Assembly of the International Union of Radio Science URSI, Chicago, August 7–17, 2008.
- [63] O'Driscoll, 2009.
- [64] R. Warty, M. Tofighi, U. Kawoos, A. Rosen, Characterization of implantable antennas for intracranial pressure monitoring: Reflection by and transmission through a scalp phantom, IEEE Trans. Microw. Theory Tech. 56 (10) (October 2008) 2366–2376.
- [65] P. Irazoqui-Pastor, I. Mody, J.W. Judy, In-vivo EEG recording using a wireless implantable neural transceiver, in: Proc. 1st Int. IEEE EMBS Conf. Neural Engineering, vol. 1, 2003, pp. 622–625.
- [66] M-Biotech, Biosensor Technology, M-Biotech, Salt Lake City, 2003, <http://www.m-biotech.com/technology1.html>.
- [67] D. McCreery, A. Lossinsky, V. Pikov, X. Liu, Microelectrode array for chronic deep-brain microstimulation and recording, IEEE Trans. Biomedical Engineering 53 (4) (April 2006) 726–737.
- [68] D. McCreery, Microelectrode array for chronic deep-brain microstimulation for recording, US 3006/0276866 A1, December 2006.
- [69] K.K. O, K. Kim, B.A. Floyd, J.L. Metha, H. Yoon, C.-M. Hung, D. Bravo, T.O. Dickson, X. Guo, R. Li, N. Trichy, J. Caserta, W.R. Bomstad II, J. Branch, D.-J. Yang, J. Bohorquez, E. Seok, L. Gao, A. Sugavanam, J.-J. Lin, J. Chen, J.E. Brewer, On-chip antennas in silicon ICs and their application, IEEE Trans. Electron Devices 52 (7) (July 2005) 1312–1323.