

A Products and closest off-target primer site

AA1 assay	90 bp	3	>=4	>=4	3	3	>=4
AA2 assay	>=4	145 bp	>=4	>=4	3	3	>=4
AA3 assay	>=4	3	145 bp	3	3	3	3
AA4 assay	>=4	>=4	>=4	125 bp	3	>=4	3
AA5 assay	>=4	>=4	>=4	>=4	107 bp	>=4	>=4
AA6 assay	>=4	3	>=4	3	3	112 bp	>=4
AA7 assay	>=4	3	>=4	>=4	>=4	>=4	95 bp
	AA1	AA2	AA3	AA4	AA5	AA6	AA7

genome (all replicons)

B Knock-out test (all 14 primers pooled)

AA1 target masked	lost						
AA2 target masked		lost					
AA3 target masked			lost				
AA4 target masked				lost			
AA5 target masked					lost		
AA6 target masked						lost	
AA7 target masked							lost
	AA1	AA2	AA3	AA4	AA5	AA6	AA7

assay

A: blue = in-silico PCR product of the designed length; grey cells give the fewest mismatches of any primer site in that genome (bowtie1, exhaustive; >=4 = no site with <=3). No off-target product with <=3 mismatches per primer.

B: each row masks one target amplicon; blue = product still formed; only the masked target is lost.