

Certificate of Calibration

Megger[®]

Issued By: Megger Instruments Limited
 Archcliffe Road
 Dover
 Kent
 CT17 9EN
www.megger.com

Instrument Details

Model: DLRO100XB
Description: Digital Low Resistance Ohmmeter
Part Number: 1004914
Serial Number: 102664261

Environment Details

Temperature	Relative Humidity	Mains Frequency	Mains Voltage	Date Tested	Date Commissioned
20°C ±3°C	50% ±20%	50Hz ±0.5Hz	230V -2% +10%	2026-Mar-23	

Traceability Information

Serial N°	Test Equipment Description	Procedure Version
T3068	DLRO100A AutoCal Test Equipment	V1.17

Test Results

Test Title	Applied Value	Reading Value	Tolerance	Uncertainties	Pass/Fail
Resistance Ranges					
150μΩ @ 100A	151.79μΩ	151.8μΩ	0.5μΩ	±0.42μΩ	Pass
1mΩ @ 10A	1.01834mΩ	1.0179mΩ	3μΩ	±0.08μΩ	Pass
1mΩ @ 50A	1.01834mΩ	1.0186mΩ	2.2μΩ	±0.08μΩ	Pass
1mΩ @ 100A	1.01834mΩ	1.0182mΩ	2.2μΩ	±0.08μΩ	Pass