

Certificate of Calibration

Megger

Issued By: Megger Instruments Limited
Archcliffe Road
Dover
Kent
CT17 9EN
www.megger.com

Instrument Details

Model: MIT1025/2-EU
Description: MIT1025 Insulation Tester
Part Number: 1016095
Serial Number: 102692031

Environment Details

Temperature	Relative Humidity	Mains Frequency	Mains Voltage	Date Tested	Date Commissioned
20°C ±3°C	50% ±20%	50Hz ±0.5Hz	230V -2% +10%	2026-Mar-31	

Traceability Information

Serial N°	Test Equipment Description	Procedure Version
T2842	High Voltage Insulation Test Equipment	V4.17

Test Results

Test Title	Applied Value	Reading Value	Tolerance	Uncertainties	Pass/Fail
Voltage Ranges					
500V Voltage Range	500.0V	507.0V	+30V / -0V	±0.5% ±1d	Pass
2500V Voltage Range	2500.0V	2546.0V	+110V / -0V	±0.5% ±1d	Pass
10000V Voltage Range	10000.0V	10206.0V	+410V / -0V	±0.5% ±1d	Pass
Insulation Ranges					
10GΩ @ 5kV Insulation Range	9.975GΩ	9.95GΩ	0.25GΩ	±0.4% ±1d	Pass
100MΩ @ 500V Insulation Range	99.86MΩ	99.9MΩ	2.5MΩ	±0.4% ±1d	Pass
100MΩ @ 10kV Insulation Range	99.86MΩ	99.6MΩ	2.5MΩ	±0.5% ±1d	Pass
47nF @ 1kV	45.9nF	48.8nF	12.35nF	±0.1% ±1d	Pass
3.3μF @ 1kV	3.3μF	3.33μF	0.175μF	±0.1% ±1d	Pass
250V AC DVM Range	250.0V	250.0V	8V	±0.1% ±1d	Pass