CTL DECISION SHEET

Standard(s): **IEC 60065/2001** **A1/2005** **IEC 60950-1/2005** **IEC 60601-1/2005** **IEC 61010-1/2001** **IEC 60335-1/2001** **A1/2004** **A2/2006**	**Sub clause(s): Various**	**Sheet N°: 624/07 CTL-WG4** **Page: 1(1)**
Subject(s): Humidity test requirements	**Key words:** **- 93% RH** **- +/-3% RH tolerance** **- 20 – 30°C**	**Meeting:** **Decision confirmed at 45th CTL meeting**

Question: Various IEC standards have requirements for humidity/temperature conditioning, but have different specifications for the test conditions.

IEC 60065 – 93 +2%/-3% (t 30°C +0/-2°C)
IEC 60950-1 – 91% - 95% (t 20-30°C +/-1°C)
IEC 60601-1 – 93 + 3%/-3% (t 20-30°C)
IEC 61010-1 – 92,5 + 2,5%/-2,5% (t 40°C +/- 2°C)
IEC 60335-1 – 93 + 3%/- 3% (t 20-30°C +/-1°C)

There does not seem to be a technical justification for the differences in the specifications.
Accordingly, it is suggested and recommended that the humidity/temperature specifications in the above standards be consistent.

Decision: The chamber is to be set at 93% RH, with +/-3% RH tolerance, at a convenient temperature between 20°C and 30°C, but the chosen temperature to be kept within +/- 2°C during the test.
The standard specifications for the humidity are to be met at the reference/measuring point, where the controlling humidity sensor is located.
The tolerances permitted in CTL DSH 251A (+/-6% RH) are allowed in other parts of the humidity chamber, away from the reference point.

Explanatory notes: --