Revised CTL DECISION SHEET

Standard(s) (incl. year)	Subclause(s)	Tracking No.	Year
IEC 60065/2014 IEC 60950-1/2013 IEC 60601-1/2012 IEC 61010-1/2010 IEC 60335-1/2010	Various	DSH 0624A	2016
Category			
GENERAL			
Subject	**Keywords**	**Developed by**	**Approved at**
Humidity test requirements	- 93% RH - +/- 3% RH tolerance - 20-30°C	WG4	2017 CTL Plenary Meeting

Question/Task

Various IEC standards have requirements for humidity / temperature conditioning, but have different specifications for the test conditions.

- IEC 60065 – 93% +/- 3% (t 20-30°C +/- 2°C or 40°C +/- 2°C);
- IEC 60950-1 – 93% +/- 3% (t 20-30°C +/- 2°C or 40°C +/- 2°C);
- IEC 60601-1 – 93% +/- 3% (t 20-30°C +/- 1°C);
- IEC 61010-1 – 93% +/- 3% (t 40°C +/- 2°C);
- IEC 60335-1 – 93% +/- 3% (t 20-30°C +/- 1°C).

There does not seem to be a technical justification for the differencies in the specifications. Accordingly, it is suggested and recommended that the humidity / temperature specifications in the above standards be consistent.

Decision

The chamber is to be set at 93% RH, with +/- 3% RH tolerance, at a convenient temperature between 20°C and 30°C or an appropriate higher temperature when required by the specific product standard, but the chosen temperature to be kept within +/- 2°C during the test.

The standard specifications for the humidity are to be met at the reference / measuring point, where the controlling humidity sensor is located.

Explanatory notes