CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.	Year
IEC 60601-1-11:2010 (1st Ed.)	8.9.1.12	2028	2015
Category			
MED			
Subject	**Keywords**	**Developed by**	**Approved at**
Environmental operating conditions a) to j)	Monitoring, basic safety, essential performance, specified atmospheric pressure	ETF-3	2016 CTL Plenary Meeting

Question

1. How to monitor/evaluate basic safety and essential performance of ME equipment for which test steps c) and d) are required (lowest and highest specified atmospheric pressure)?
2. How should pressure test be performed if it is done at an external test house?

Decision

Taking into consideration IEC 60601-1-11:2015 (which constitutes a collateral standard to IEC 60601-1, Ed. 3.1):

- The evaluation of basic safety need not include leakage and dielectric strength testing.
- The manufacturer shall determine and agree with the test house prior start of the test sequence, on which functionality of the ME equipment can appropriately be used for evaluation of its basic safety and essential performance when operated at the least favorable working conditions.
- The full test sequence a) to j) shall be followed unless the manufacturer provides sufficient evidence that the equipment is not pressure sensitive, in which case steps 4.2.2 c) and d) may be waived.
- The pressure testing as per 4.2.2 c) and d) shall be conducted at the full range of pressure. Pressure testing may be conducted independently from the temperature and humidity ranges.

Explanatory notes

The standard does not indicate how basic safety and essential performance shall be monitored, in particular while exposed to lowest and highest specified atmospheric pressure as per c) and d).