X CTL Provisional DECISION SHEET (PDSH)
☐ CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.
IEC 60601-1-9:2007+A1:2013	1.1, 4	PDSH 2138
Category		
MED		
Subject	**Keywords**	**Developed by**
ME-equipment	IEC 60601-1-9	ETF 17

Question

1. Do you as ETF-17 member have experience of the use of this standard? Some countries require this standard for their approval. It is not very clear when we have to use this standard (except for the ME-system)

2. Do NCB have to state this standard on the CB-certificate?

Decision

1. There are at this point in time, no particular or collateral standards that require evaluation according to IEC 60601-1-9. According to OD-2055 60601-1-9 is not required for issuing a certificate. It will be up to the manufcaturer to request evaluation
2. If evaluation is requested by the manufacturer and evaluated in full and compliance is shown, the NCB can state the standard on the certificate

Explanatory notes

Enquiry: The scope of the standard IEC 60601-1-9 applies to all MEDICAL ELECTRICAL EQUIPMENT.

1.1 * Scope

This International Standard applies to the reduction of adverse ENVIRONMENTAL IMPACTS of MEDICAL ELECTRICAL EQUIPMENT, hereafter referred to as ME EQUIPMENT.

MEDICAL ELECTRICAL SYSTEMS are excluded from the scope of this collateral standard.

Date: 2018-11-19

PDSH:
Year of publication: 2025
To be approved at CTL plenary: 2026

DSH
Publication date:

Click here to enter text.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.