CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.	Year
IEC 60601-1, ed. 3, 2005 + A1: 2012	8.7.4	2024	2015
Category			
MED			
Subject	**Keywords**	**Developed by**	**Approved at**
Internally powered ME Leakage current and Touch current	Internally powered ME equipment; without connection to mains.	ETF-3	2016 CTL Plenary Meeting

Question

In the second edition of the standard, specific figures were dedicated for testing of internally powered ME Equipment. This is not the case in the third edition.

How shall the Leakage and Touch current measurement be performed in the third edition?

Decision

Subclause 8.7.4.6 shall be followed.

Explanatory notes

To WG14: To reinstate Figure 23 of IEC 60601 Ed. 2 in IEC 60601 Ed.3.