☐ CTL Provisional DECISION SHEET (PDSH)
x CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.
IEC 60601-1:2005 + A1:2012 IEC 60601-1:2005 + A1:2012; IEC 60601-1:2005 + A1:2012 IEC 60601-1:2005 + A1:2012 + A2:2020	Clause 1.3 (Scope)	2208A
Category		
MED		
Subject	**Keywords**	**Developed by**
Office / IT equipment under IEC 60601-1	Definition MEE/MES,	ETF 3

Question

Can Office / IT equipment be tested and certified to IEC 60601-1 in IECEE CB scheme as a stand alone product?

Decision

The scope of IEC 60601-1 is limited to MEE and MES. Typical OFF/ITE/AV_ICT Equipment does not meet the 60601-1 definition of MEE/MES and cannot be evaluated to IEC 60601-1 under the CB scheme as a stand alone device.

See IECEE OD-2055 (Operational Document on Medical Electrical Equipment in the CB Scheme according to the IEC 60601 and ISO/IEC 80601 Series of Standards) for specific rules associated with power supply and component/subassemblies certifications under the CB Scheme.

Explanatory notes

Enquiry:
Date: 2022-09-13

PDSH:
Year of publication: 2023
To be approved at CTL plenary: 2024

DSH
Publication date: 2024-09-13

Medical electrical equipment and medical electrical systems are defined terms in IEC 60601-1. Currently, different non-medical equipment is tested and certified to IEC 60601-1. This includes off-the-shelf PCs, monitors, routers etc. OFF/ITE/AV_ICT equipment not meeting the definition of MEE/MES as per clause 3.63 and 3.64 is not covered by the scope of the standard.

Clause 16 covers medical electrical system (MES) requirements and allows IEC 60950-1 or IEC 62368-1 equipment to be used inside or in combination with MEE / MES under certain conditions.

Additionally, Annex I provide examples on the use of such equipment together or inside MEE. We recognize that manufactures may wish to provide OFF/ITE/AV_ICT equipment that meet means of patient protection in order to more easily combine in medical electrical systems (MES). However, OD-2055 edition 2.3 does not have a framework for consistent application of requirements for these situations.

Note while it is recognized that there may be a need to consider requirements for OFF/ITE/AV_ICT Equipment to be evaluated to aspects of 60601-1, there are no agreed methodology for these evaluations in OD-2055 and

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.

updates to the document would be required in order to allow for CB certifications for OFF/ITE/AV_ICT Equipment to 60601-1.