DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed.2 Am1+Am2	**Sub clause(s):** 16	**Sheet n°:** **DSH-404**
Subject: ENCLOSURES and PROTECTIVE COVERS	**Key words:** ENCLOSURES, PROTECTIVE COVERS, SIP/SOP	**Confirmed by CTL at its 39th meeting, in Cologne**

Question:

In practice, many standard SIP/SOPs are constructed so that live parts of the SIP/SOPs are accessible with the standard test finger or, alternatively, the air clearances and/or creepage distances between these live parts and the standard test finger are too short. This means that these live parts of SIP/SOPs are to be considered as part of the enclosure and therefore the corresponding enclosure leakage currents shall be measured, and hence, many existing standard SIP/SOPs fail to comply with the standard.

Should, for instance, concession be granted for SIP/SOPs with accessible live parts with a voltage (to earth) not exceeding 25V AC or 60V DC?

Decision:

Include SIP/SOP in the list of sub-clause 16e)1.
Instructions for use must contain the warning given in 16a)5.

Explanatory notes:

For the operator, access to parts at potentials not exceeding 25V AC or 60V DC is consider not to present a hazard, provided that simultaneous contact between operator and patient is avoided.