DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed.2 Am1+Am2	**Sub clause(s):** 20.1	**Sheet n°:** **DSH-411**
Subject: General requirements	**Key words:** Leakage current, SIP/SOP, working voltage, A-k	**Confirmed by CTL at its 39th meeting, in Cologne**

Question:

I) Does exemption a) refer to voltages arising within the EQUIPMENT, to voltages appearing from external sources, or both?
ii) Does exemption b) refer to the SFC of mains voltage on the SIP or SOP? Is insulation which ensures absence of excessive LEAKAGE CURRENTS not to be tested?
iii) For exemption d) to apply, does the manufacturer have to restrict connection of the SIP and SOP to other MEDICAL ELECTRICAL EQUIPMENT or can connection to non-medical equipment be permitted, subject to some restrictions?

Decision:

i) a) applies only if the voltages within the SIP and SOP are less than SELV when connected to EQUIPMENT consistent with the instructions for use.
ii) b) applies to a SFC resulting from a single component failure within the SIP or SOP in NORMAL USE i.e. connected as specified in the instructions for use. If such a SFC does not produce excessive LEAKAGE CURRENTS then no additional insulation test is needed.
iii) d) applies only if the manufacturer restricts connection of the SIP or SOP to MEDICAL ELECTRICAL EQUIPMENT

Explanatory notes:

a) is intended to cover the working voltages on the SIP or SOP in NORMAL USE regardless of their origin.
b) is intended to cover the SFC caused by a component failure within the SIP or SOP in NORMAL USE.
d) exemption does not allow connection to equipment, only Medical Electrical EQUIPMENT, as defined is covered.