DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed.2 Am1+Am2	**Sub clause(s):** 4.10	**Sheet n°:** **DSH-415**
Subject: Preconditioning of equipment intended for conditioned rooms	**Key words:** Preconditioning	**Confirmed by CTL at its 39th meeting, in Cologne**

Question:

Should this test be applied to equipment which is permanently installed and operated only in a controlled temperature and humidity conditions as specified in the ACCOMPANYING DOCUMENTS

Decision:

No exception can be made from the requirements of Sub-clause 4.10. However, attention is drawn to the 3rd paragraph:

The test shall be applied only to those EQUIPMENT parts likely to create a SAFETY HAZARD when influenced by the climatic conditions that are simulated by the test.

Explanatory notes:

Certain equipment is permanently installed and operated in controlled temperature and humidity conditions. Nevertheless IEC 60601-1 requires equipment to comply with all the relevant parts of the standard when exposed to the environmental conditions specified in Sub-clause 10.2.1, even though the operating conditions may be restricted, as allowed in Sub-clause 6.8.3.