DECISION SHEET

Standard(s)- (year and edition):	**Sub clause(s):**	**Sheet n°:**
IEC 60601-1:1988 Ed.2 Am1+Am2	57.4	**DSH-427**
Subject: Appliance coupler on mobile equipment	**Key words:** Mobile, appliance coupler	**Confirmed by CTL at its 39th meeting, in Cologne**

Question:

In a mobile unit (with power supply cord attached), as a result of excessive amount of movement of the equipment, it is possible that damage to internal parts of the appliance coupler can occur which eventually could result in a fire hazard. What requirements should be applied?

Decision:

If the appliance coupler is not according to IEC 60320 apply the tests according to 57.4.

Explanatory notes:

If the appliance coupler complies with IEC 60320, there is no possibility of applying additional requirements.