COLLECTION OF THE CTL DECISIONS
DECISION SHEET

Standard(s)- (year and edition): IEC 60601-1:1988 Ed.2 Am1+Am2	**Sub clause(s):** General	**Sheet n°: 537**
Subject: PCB layer transformers,	**Key words:** PCB layer transformers	Decision taken at the 40th meeting 2003

Question:

What requirements are needed for PCB layer transformers using etched windings?

Decision:

The requirements of the IEC 60601:1988 Ed.2 should be used with regard to creepage distances and isolation distances through thickness.

Explanatory notes:

This issue will be put forward to TC 66 and TC62 to deal with this specific problem. In the meantime a temperature cycling test according to IEC 60950 can be used for reference. Base for this test are the temperatures retrieved during normal condition tests calculated/adjusted to an ambient of 40 °C. It is important that the influence of the temperature of other components shall be taken into consideration. The minimum thickness through insulation can be taken as from IEC 60950, being 0,4 mm.