X CTL Provisional DECISION SHEET (PDSH)
☐ CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	**Subclause(s)**	**Tracking No.**
IEC 60601-1:2005+A1:2012	4.8	PDSH 2170
Category		
MED		
Subject	**Keywords**	**Developed by**
Identification of Critical Components – Minimum description criteria	Components	ETF 17

Question

Can the interchangeable method suggested by IECEE OD-G-2060 be used for wiring and printed wiring boards also in MED category?

Decision

This decision is not necessary, OD-G-2060:2017 is applicable for all categories. OD-G-2060:2017, does not explicitly allow or deny wiring and printed wiring boards. Printed wiring boards are part of the examples, eventhough wiring isn't part of the examples, these could still be eligible for interchangeable, for wiring where no testing was conducted in the end-product interchangeable is acceptable according to the OD.

Explanatory notes

Enquiry:
Date:

PDSH:
Year of publication: 2025
To be approved at CTL plenary: 2026

DSH
Publication date:

Click here to enter text.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.