CTL Provisional DECISION SHEET (PDSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.	Publication date
IEC 60601-1:2005/AMD2:2020	All	2241A	2024
Category			
MED			
Subject	**Keywords**	**Developed by**	**To be approved**
Use of part 2 standards with new updated part 1	Part 2 late updates	ETF 17	2026 CTL Plenary Meeting

Question

Can IEC 60601-1:2005/AMD2:2020 (edition 3.2) be used with IEC 60601-2-4:2010/AMD1:2018 (edition 3.1), since there is presently no version of standard IEC 60601-2-4 that is published to be used in conjunction with the IEC 60601-1:2005+AMD1: 2012+AMD2:2020 (edition 3.2)?

Decision

For the MED category, it is possible to issue a CBTR and CBTC when the particular standard (Part 2) has not been updated to the latest edition of the Part 1 standard.

A statement shall be added to the GPI (General products information) section of the CBTR Part 1 and Part 2 standards showing justification to utilize a specific Part 1 Standard TRF based on references found in IEC Part 2 Standard:

Generic Statement:
“Based on a demonstrated RM file evidence usage of IEC 60601-1:2005/Amendment(s) *(1)* standard with Part 2 Standard *(2)* was approved and consequentially TRF IEC60601_X *(3)* was used for completion of this CBTR.”

Note (1): Engineer needs to input correct amendment(s) including Publication Year.

Note (2): Engineer needs to input correct Part 2 Standard including Publication Year.

Note (3): Engineer needs to input correct TRF version.

Example: “Based on a demonstrated RM file evidence usage of IEC 60601-1:2005+AMD1:2012+AMD2:2020 standard with ISO 80601-2-61:2017 was approved and consequentially TRF IEC60601_Q was used for completion of this CBTR.”

As soon as the particular standard aligned with the normative reference is published, the correct collateral or particular standard shall be used. A grace period of two years from the date of publication of the collateral or particular is considered a suitable transition period. After the two year period, it is expected the aligned collateral or particular shall be used.

Explanatory notes

It can be found acceptable to use a Part 2 standard (particular standard : IEC 60601-2-XX or IEC/ISO 80601-2 -XX) that has not been updated to the latest edition of the Part 1 standard (general: IEC 60601-1 and collaterals: IEC 60601-1-XX). A document gap analysis conducted by the manufacturer utilizing the risk management process (ISO 14971) is required to determine the suitability of using a part 2 standard that has not been updated to the latest edition of the part 1 standard.

Example: The IECEE CBTL can use the IEC 60601-1:2005+AMD1:2012+AMD2:2020 TRF combined with an IEC 60601-2-XX and IEC/ISO 80601-2-XX TRF(s) which is/are in conjunction with IEC 60601-1:2005+AMD1:2012 standard as long as the manufacturer’s RM File provides evidence that the manufacturer demonstrates that the hazards (hazardous situation) are addressed and documented in the applicable TRF Part 1 or Part 2 standards.

This permission is stated in all applicable documents of the IEC 60601-1 series to clearly identify where this was considered in the individual TRF clauses.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version

Acknowledging the fact, that due to the complexity of the MEE and MES, testing time can take place over multiple months, a grace period is introduced to ensure a transition within an acceptable timeframe.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version