CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	**Subclause(s)**	**Tracking No.**	**Year**
IEC 61010-1 + Amd 1:2016	1.1.3 and 5.4.4	DSH 2108	2018
Category			
MEAS			
Subject	Keywords	Developed by	Approved at
IT-Equipment	IEC 60950 IEC 62368	ETF 3	2019 CTL Plenary Meeting

Question

In the subjected clauses the IEC 60950 equipment is mentioned.

Can the standard be understood as including IEC 62368 wherever IEC 60950 is mentioned?

Decision

Generally, IEC 62368-1 approved power-supplies can be accepted based on requirements of IEC 61010 clause 14 (see also Figure 15). It is therefore not necessary with a specific exception to allow the use of these IEC 62368-1 approved power-supplies per the opinion of the IEC technical committee (TC 66). Attention shall be given to clauses 1.4, 6.3 and 10 of 61010-1.

Exception is given to external located, enclosed type, power-supplies. This since IEC 62368-1 do allow a greater leakage current than the limits of 6.3.1 / 6.3.2. These externally located, enclosed type power-supplies are still considered as acceptable.

Enclosed type when used internally, shall comply with clauses 1.4, 6.3 and 10 of 61010-1.

Explanatory notes

The IEC 60950-1 will soon be replaced by IEC 62368-1. Eventually, it will be difficult to find IT-equipment and power supplies certified to IEC 60950-1 because the IEC 62368-1 will be the state of the art standard. Equipment regarded as safe for use by laymen and in the home environment, as is the case with IT-equipment, would reasonably also be sufficiently safe for use by operators of equipment for measurement, control and laboratory use.