COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 297** **Page 1(1)**
IEC 61010-1:1990 +A1:1992+A2:1995	5.1.5	
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Measuring Circuit Terminals (Amd. 2)	Terminals Marking	

Question:

Is it acceptable for a measuring equipment to have its sensing terminals marked:

"CAT II (600V)

CAT III (300V)"

Decision:

Marking with both categories with the relevant voltages as shown is acceptable.

Explanation:

The standard allows operation under different categories provided the isolation is suitable for the voltages concerned. In this case, it is possible to mark both categories since the requirements for Cat II at 600V are the same as those for Cat III at 300V.