COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 298** **Page 1(1)**
IEC 61010-1:1990+A1:1992+A2:1995	5.1.6(f)	
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Marking, Terminals and Operating Devices	- Marking - Mains	

Question:

Can symbols 9 and 10 of Table 1 (0,1) be used for any other purpose than mains disconnection?

Decision:

The standard states these symbols are reserved for mains disconnection devices, the symbols should therefore, be used only for this purpose. However, these symbols are allowed for other use in other IEC standards and some Test laboratories therefore, exercise discretion on this matter provided the mains separation device is clearly indicated either on the unit or in the handbook.

Explanation:

It is likely that the use of 0,1 symbols will be reviewed across all IEC documents and, pending this review, it is important that the mains disconnection device is clearly identified either on the equipment or in the instructions for use.