COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992+A2:1995	**Sub clause:** 5.1.8	**Sheet n. 299** **Page 1(1)**
Subject: Marking of Battery Terminals	**Key words:** - Battery - Marking	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

There is no requirement for the marking of battery terminals. Has this caused difficulty in test programmes?

Decision:

No marking is required, no difficulties were reported as a consequence.

Explanation:

Clause 13.2.2. requires batteries intended to be changed by the user to be connected in reverse polarity, if possible, as a type test and this has eliminated hazards.

Note: Care should be taken with certain cells, (eg lithium), when performing this test to ensure there is no risk of explosion.