COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 302**
IEC 61010-1:1990+A1:1992+A2:1995	6.3.1.3	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Capacitance	- Capacitance - Charge	

Question:

How can the stored charge be measured, in particular where the circuit is complex?

Decision and Explanation:

Only the charge accessible to the user is of concern under the standard. Therefore, where the circuit is complex, it is considered that a method of measurement from the accessible part is needed. It is, therefore, recommended that a procedure of discharge through a defined resistor is used, monitoring the voltage – time profile to provide a method of calculation of charge.

When this method is used it is requested that details are recorded in the report for the benefit of CB members.

The method described in IEC 60950 Subclause 2.1.10 is recommended.