COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990+A1:1992+A2:1995	**Sub clause:** 6.12	**Sheet n. 303** **Page 1(1)**
Subject: Disconnection from Supply Source	**Key words:** - Mains - Disconnect - Symbols	**Provisional decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

Where disconnection is by means of a mains plug or appliance inlet, how is this to be indicated?

Decision:

The manual shall provide sufficient information to clearly identify the mains disconnect device.

Explanation: ----