COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 304**
IEC 61010-1:1990+A1:1992+A2:1995	6.7	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Protection against electric shock, Table D18	Creepage	

Question:

What creepages apply when voltages in excess of 1000V are present on printed wiring boards?

Decision:

Use the following columns:

For pollution degree 1 use “Other material”
For pollution degree 2 use CTI>600

Explanation:

IEC 60664 Fig. 1 gives the source of these distances and shows that the values are as given above.