COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 305**
IEC 61010-1:1990+A1:1992+A2:1995	7	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Protection against mechanical hazards	- Fixed - Mounting	

Question:

What requirements are needed on fixing devices for wall mounting equipment?

Decision:

No specific requirements are given in the standard, however, where a hazard could arise if a device becomes detached from its mounting, laboratories are performing a 4x loading test applied vertically and/or horizontally as appropriate.

Explanation:

There is no requirement to perform any test if analysis shows that no hazard can arise. The use of the 4x loading test is in line with other IEC documents (eg IEC 60601).