COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 306**
IEC 61010-1:1990+A1:1992+A2:1995	7.1	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Protection against mechanical hazards	- Guards - Access	

Question:

The requirement states that there must be no access to hazardous moving parts in Normal Condition and no parts must be expelled in Single Fault Condition. Does this mean it is acceptable to have access to hazardous parts in Single Fault Condition?

Decision:

Mechanical parts are not subject to Single Fault Condition examination.

Explanation:

Specific tests are included in the standard for strength, e.g. impact hammer test, which determine the acceptable strength of parts. It is not permissible to remove or presume destruction of mechanical components in simulated Single Fault Conditions in addition to these pre-determined requirements.

Where safety of a moving part is determined by an electronic control, for example a speed control circuit, then single fault analysis of this circuit must be performed if a hazard could arise.