COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 312**
IEC 61010-1:1990+A1:1992+A2:1995	14.2.1	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Components, motor temperature	- Components - Temperature	

Question:

Where a motor is arranged physically or in a circuit such that it cannot present a shock, temperature or fire hazard, is it necessary to perform the tests specified?

Decision:

Tests are only performed where a hazard is likely to arise. Analysis and examination should be made to determine the hazard and tests performed if needed.

Explanation:

There is no reason to perform tests which do not examine the safety of a device. If consideration shows no hazard will arise, the tests are not needed. A clear explanation should be given in the report for a decision not to test.