COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 315**
IEC 61010-1:1990+A1:1992+A2:1995	14.7	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Mains transformers	Transformer Approval	

Question:

Where a transformer is certified to IEC 60742, should further examination and test be performed ?
Note that IEC 60742 has been withdrawn and replaced by IEC 61558 series.

Decision

The entire equipment needs to be in conformity with IEC 61010-1. To achieve this, recognition should be given to those features of the device which meet these requirements and are recorded in the approval report. Other features not covered should be examined for the application concerned.

Explanation

There is not at this time, a common set of requirements for all transformers, power supplies and so on, laboratories must, therefore, consider the application of each device for the use it is employed in the equipment.