COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 317**
IEC 61010-1:1990+A1:1992+A2:1995	1.1	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Scope	- Battery Charger - Scope	

Question:

A battery charger for lead acid batteries may be used in a room. Does the battery charger fall within the scope of IEC61010-1?

Decision:

Battery Chargers – Generally these are within the scope of IEC60335 unless they are supplied as a part of the apparatus falling within IEC61010-1 when they are considered as an integral part of the equipment.

Chargers supplied with apparatus under IEC60335 but with any additional requirements from IEC61010-1 as pertinent for example due to extended climatic conditions.

Explanation

-