COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990+A1:1992+A2:1995	**Sub clause:** 5	**Sheet n. 318** **Page 1(1)**
Subject: Marking and Documentation	**Key words:** - Document - Installation - Mains Supply	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

Where a device has more than one mains cord e.g. laboratory bench, how is the marking requirement fulfilled?

Decision:

IEC61010-1 does not allow a single device to have more than one means of disconnection from the supply. In general this will preclude the use of more than 1 mains supply cord if the cord is the means of disconnection.

Explanation:

The requirements for mains connection are given in Clause 6.12.2.2. which refers to a single means of disconnection from the supply i.e. a switch or an appliance coupler or a separate plug which will in general preclude multiple connection means to the mains supply.