COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 321**
IEC 61010-1:1990+A1:1992+A2:1995	6.11.1	**Page 1(1)**
Subject:	**Key words:**	**Provisional decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Terminals	- Terminals - Contact - Strand	

Question:

The test in this clause allows a stripped conductor 8 mm in length to be moved around to contact other parts. Is it acceptable for the conductor to be sent around objects or is it to remain straight and only act as a radius about the point of entry to the terminal?

If bending is allowed how is the limit determined?

Decision:

The strand is to be kept straight as far as reasonable and moved around as a radius from the terminal point.

Explanation

The strand is to be kept straight as far as reasonable and moved around as a radius from the terminal point.