COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 322**
IEC 61010-1:1990+A1:1992+A2:1995	6.7	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Clearances and Creepage distances	- Clearance - Creepage - Operational insulation	

Question:

What are the creepage and clearance distances for operational insulation.

Decision:

IEC61010-1 does not define creepage or clearances for operational insulation and therefore the values are left to the discretion of the manufacturer.

Explanation

-