COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990+A1:1992+A2:1995	**Sub clause:** 6.7	**Sheet n. 324** **Page 1(1)**
Subject: Clearances and Creepage distances	**Key words:** - Insulation - Thickness	**Provisional decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

What are the distances required through insulation for
- Basic
- Supplementary
- Double
- Reinforced insulation?

Decision and Explanation

Distances through insulation are not defined in IEC61010-1 and are therefore left to the discretion of the manufacturer. The high voltage tests required by IEC61010-1 will reveal whether the insulation complies with requirements.