COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 325**
IEC 61010-1:1990+A11992+A2:1995	5	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Marking and Documentation	- Document - Installation - Mains Supply	

Question:

Where devices are designed to check the installation of power supply wiring within buildings and there is a risk to occupants of the building from the wiring under test due to the power levels employed, what marking and instructions should be provided?

Decision:

The instructions provided should include warnings concerning the hazard and reference to any national regulations covering supply installation.